


LB137799

7470A

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.008	-0.008	13	0.000	13	0			
0.2	0.200	0.171	-0.029	745	0.0 %	745	0			
2.5	2.500	2.464	-0.036	10112	0.0 %	10112	0			
5.0	5.000	5.069	0.069	20752	0.0 %	20752				
7.5	7.500	7.635	0.135	31237	0.0 %	31237				
10.0	10.000	9.870	-0.130	40364	0.0 %	40364				

LB137799 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 06 Nov 2025 13:06:11

Type	Sample ID	Extended ID	Conc.	μ Abs.	Units	Type	Std Conc	Method	Date
S	0.0 - 1	50	-	13	PPB	Std	0.0000	7470A	06 Nov 2025 13:10:53
S	0.2 - 1	50.2	-	745	PPB	Std	0.2000	7470A	06 Nov 2025 13:13:10
S	2.5 - 1	52.5	-	10112	PPB	Std	2.5000	7470A	06 Nov 2025 13:15:28
S	5.0 - 1	55	-	20752	PPB	Std	5.0000	7470A	06 Nov 2025 13:17:45
S	7.5 - 1	57.5	-	31237	PPB	Std	7.5000	7470A	06 Nov 2025 13:22:56
S	10.0 - 1	510	-	40364	PPB	Std	10.0000	7470A	06 Nov 2025 13:28:12
U	ICV101 - 1	ICV101	3.6977	15152	PPB	SMPL	-7470A		06 Nov 2025 13:37:03
U	ICB101 - 1	ICB101	-0.0511	-162	PPB	SMPL	-7470A		06 Nov 2025 13:39:17
U	CCV34 - 1	CCV34	4.9494	20265	PPB	SMPL	-7470A		06 Nov 2025 13:41:35
U	CCB34 - 1	CCB34	-0.0504	-159	PPB	SMPL	-7470A		06 Nov 2025 13:43:50
U	CRA - 1	CRA	0.1726	752	PPB	SMPL	-7470A		06 Nov 2025 13:46:09
U	HighStd - 1	HighStd	9.6960	39655	PPB	SMPL	-7470A		06 Nov 2025 13:48:24
U	ChkStd - 1	ChkStd	6.7467	27607	PPB	SMPL	-7470A		06 Nov 2025 13:50:39
U	PB170408BL - 1	PBW	-0.0746	-258	PPB	SMPL	-7470A		06 Nov 2025 13:53:05
U	PB170408BS - 1	LCSW	3.8536	15789	PPB	SMPL	-7470A		06 Nov 2025 13:55:28
U	Q3532-04 - 1	TP-1	-0.0465	-143	PPB	SMPL	-7470A		06 Nov 2025 13:57:45
U	Q3532-04DUP - 1	TP-1DUP	0.0039	63	PPB	SMPL	-7470A		06 Nov 2025 14:00:04
U	Q3532-04MS - 1	TP-1MS	3.6468	14944	PPB	SMPL	-7470A		06 Nov 2025 14:02:21
U	Q3532-04MSD - 1	TP-1MSD	3.4950	14324	PPB	SMPL	-7470A		06 Nov 2025 14:04:38
U	Q3532-08 - 1	TP-2	-0.0516	-164	PPB	SMPL	-7470A		06 Nov 2025 14:06:56
U	CCV35 - 1	CCV35	4.9330	20198	PPB	SMPL	-7470A		06 Nov 2025 14:09:14
U	CCB35 - 1	CCB35	-0.0813	-285	PPB	SMPL	-7470A		06 Nov 2025 14:11:30
U	Q3536-02 - 1	FILTER-COMP	-0.0007	44	PPB	SMPL	-7470A		06 Nov 2025 14:13:48
U	Q3537-02 - 1	MW-17-SOIL	-0.0112	1	PPB	SMPL	-7470A		06 Nov 2025 14:16:03
U	Q3544-04 - 1	ONYX-1	-0.0078	15	PPB	SMPL	-7470A		06 Nov 2025 14:18:19
U	Q3546-10 - 1	TP-COMP	0.0047	66	PPB	SMPL	-7470A		06 Nov 2025 14:20:35
U	Q3549-02 - 1	RT4922	0.0071	76	PPB	SMPL	-7470A		06 Nov 2025 14:22:50
U	PB170409BL - 1	PBW	-0.0051	26	PPB	SMPL	-7470A		06 Nov 2025 14:25:07
U	PB170409BS - 1	LCSW	3.7361	15309	PPB	SMPL	-7470A		06 Nov 2025 14:27:24
U	Q3534-10 - 1	EB-1	-0.0756	-262	PPB	SMPL	-7470A		06 Nov 2025 14:29:41
U	Q3552-01 - 1	MW-4	-0.0032	34	PPB	SMPL	-7470A		06 Nov 2025 14:32:00
U	Q3552-01DUP - 1	MW-4DUP	0.0052	68	PPB	SMPL	-7470A		06 Nov 2025 14:34:16
U	CCV36 - 1	CCV36	4.7433	19423	PPB	SMPL	-7470A		06 Nov 2025 14:36:31
U	CCB36 - 1	CCB36	-0.0746	-258	PPB	SMPL	-7470A		06 Nov 2025 14:38:47
U	Q3552-01MS - 1	MW-4MS	3.4754	14244	PPB	SMPL	-7470A		06 Nov 2025 14:41:05
U	Q3552-01MSD - 1	MW-4MSD	3.5232	14439	PPB	SMPL	-7470A		06 Nov 2025 14:46:15
U	Q3552-02 - 1	MW-8	0.1567	687	PPB	SMPL	-7470A		06 Nov 2025 14:48:31
U	Q3552-03 - 1	MW-9	0.3011	1277	PPB	SMPL	-7470A		06 Nov 2025 14:50:51
U	Q3552-04 - 1	MW-10	-0.0149	-14	PPB	SMPL	-7470A		06 Nov 2025 14:53:08
U	Q3552-05 - 1	MW-3	0.0810	378	PPB	SMPL	-7470A		06 Nov 2025 14:55:25
U	CCV37 - 1	CCV37	4.7587	19486	PPB	SMPL	-7470A		06 Nov 2025 14:57:42
U	CCB37 - 1	CCB37	-0.0705	-241	PPB	SMPL	-7470A		06 Nov 2025 14:59:59
U	PB170394TB - 1	PB170394TB	-0.0024	37	PPB	SMPL	-7470A		06 Nov 2025 15:02:19
U	Q3532-04lx5 - 1		0.0020	55	PPB	SMPL	-7470A		06 Nov 2025 15:04:35
U	Q3532-04a - 1		3.6901	15121	PPB	SMPL	-7470A		06 Nov 2025 15:06:50
U	Q3552-01lx5 - 1		-0.0617	-205	PPB	SMPL	-7470A		06 Nov 2025 15:09:06
U	Q3552-01A - 1		3.6774	15069	PPB	SMPL	-7470A		06 Nov 2025 15:11:24
U	ccv38 - 1	ccv38	4.8622	19909	PPB	SMPL	-7470A		06 Nov 2025 15:13:40
U	ccb38 - 1	ccb38	-0.0795	-278	PPB	SMPL	-7470A		06 Nov 2025 15:15:58