



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick Engineers

SDG No.: Q3534

Contract: REMI02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB100	Mercury	0.076	+/-0.2	U	0.20	CV	11/06/2025	08:09	LB137791

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick Engineers

SDG No.: Q3534

Contract: REMI02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB31	Mercury	0.076	+/-0.2	U	0.20	CV	11/06/2025	08:13	LB137791
CCB32	Mercury	0.076	+/-0.2	U	0.20	CV	11/06/2025	08:44	LB137791
CCB33	Mercury	0.076	+/-0.2	U	0.20	CV	11/06/2025	09:10	LB137791
CCB34	Mercury	0.076	+/-0.2	U	0.20	CV	11/06/2025	09:19	LB137791

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick Engineers

SDG No.: Q3534

Contract: REMI02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB101	Mercury	0.076	+/-0.2	U	0.20	CV	11/06/2025	13:39	LB137799

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick Engineers

SDG No.: Q3534

Contract: REMI02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB34	Mercury	0.076	+/-0.2	U	0.20	CV	11/06/2025	13:43	LB137799
CCB35	Mercury	0.076	+/-0.2	U	0.20	CV	11/06/2025	14:11	LB137799
CCB36	Mercury	0.076	+/-0.2	U	0.20	CV	11/06/2025	14:38	LB137799
CCB37	Mercury	0.076	+/-0.2	U	0.20	CV	11/06/2025	14:59	LB137799
CCB38	Mercury	0.076	+/-0.2	U	0.20	CV	11/06/2025	15:15	LB137799

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick Engineers

SDG No.: Q3534

Contract: REMI02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	18.6	+/-50	J	100	P	11/05/2025	11:41	LB137798
	Antimony	6.76	+/-25	U	50.0	P	11/05/2025	11:41	LB137798
	Arsenic	5.12	+/-10	U	20.0	P	11/05/2025	11:41	LB137798
	Barium	14.6	+/-50	U	100	P	11/05/2025	11:41	LB137798
	Beryllium	0.56	+/-3	U	6.00	P	11/05/2025	11:41	LB137798
	Cadmium	0.50	+/-3	U	6.00	P	11/05/2025	11:41	LB137798
	Calcium	234	+/-1000	U	2000	P	11/05/2025	11:41	LB137798
	Chromium	2.12	+/-5	U	10.0	P	11/05/2025	11:41	LB137798
	Cobalt	2.26	+/-15	U	30.0	P	11/05/2025	11:41	LB137798
	Copper	4.60	+/-10	U	20.0	P	11/05/2025	11:41	LB137798
	Iron	29.3	+/-50	J	100	P	11/05/2025	11:41	LB137798
	Lead	2.30	+/-6	U	12.0	P	11/05/2025	11:41	LB137798
	Magnesium	244	+/-1000	U	2000	P	11/05/2025	11:41	LB137798
	Manganese	5.94	+/-10	U	20.0	P	11/05/2025	11:41	LB137798
	Nickel	3.06	+/-20	U	40.0	P	11/05/2025	11:41	LB137798
	Potassium	918	+/-1000	U	2000	P	11/05/2025	11:41	LB137798
	Selenium	9.64	+/-10	U	20.0	P	11/05/2025	11:41	LB137798
	Silver	1.62	+/-5	U	10.0	P	11/05/2025	11:41	LB137798
	Sodium	868	+/-1000	U	2000	P	11/05/2025	11:41	LB137798
	Thallium	4.38	+/-20	U	40.0	P	11/05/2025	11:41	LB137798
	Vanadium	6.26	+/-20	U	40.0	P	11/05/2025	11:41	LB137798
	Zinc	16.7	+/-20	U	40.0	P	11/05/2025	11:41	LB137798

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick Engineers

SDG No.: Q3534

Contract: REMI02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	11.3	+/-50	U	100	P	11/05/2025	12:33	LB137798
	Antimony	6.76	+/-25	U	50.0	P	11/05/2025	12:33	LB137798
	Arsenic	5.12	+/-10	U	20.0	P	11/05/2025	12:33	LB137798
	Barium	14.6	+/-50	U	100	P	11/05/2025	12:33	LB137798
	Beryllium	0.56	+/-3	U	6.00	P	11/05/2025	12:33	LB137798
	Cadmium	0.50	+/-3	U	6.00	P	11/05/2025	12:33	LB137798
	Calcium	234	+/-1000	U	2000	P	11/05/2025	12:33	LB137798
	Chromium	2.12	+/-5	U	10.0	P	11/05/2025	12:33	LB137798
	Cobalt	2.26	+/-15	U	30.0	P	11/05/2025	12:33	LB137798
	Copper	4.60	+/-10	U	20.0	P	11/05/2025	12:33	LB137798
	Iron	31.0	+/-50	J	100	P	11/05/2025	12:33	LB137798
	Lead	2.30	+/-6	U	12.0	P	11/05/2025	12:33	LB137798
	Magnesium	244	+/-1000	U	2000	P	11/05/2025	12:33	LB137798
	Manganese	5.94	+/-10	U	20.0	P	11/05/2025	12:33	LB137798
	Nickel	3.06	+/-20	U	40.0	P	11/05/2025	12:33	LB137798
	Potassium	918	+/-1000	U	2000	P	11/05/2025	12:33	LB137798
	Selenium	9.64	+/-10	U	20.0	P	11/05/2025	12:33	LB137798
	Silver	1.62	+/-5	U	10.0	P	11/05/2025	12:33	LB137798
	Sodium	868	+/-1000	U	2000	P	11/05/2025	12:33	LB137798
	Thallium	4.38	+/-20	U	40.0	P	11/05/2025	12:33	LB137798
	Vanadium	6.26	+/-20	U	40.0	P	11/05/2025	12:33	LB137798
	Zinc	16.7	+/-20	U	40.0	P	11/05/2025	12:33	LB137798
CCB02	Aluminum	25.4	+/-50	J	100	P	11/05/2025	13:25	LB137798
	Antimony	6.76	+/-25	U	50.0	P	11/05/2025	13:25	LB137798
	Arsenic	5.12	+/-10	U	20.0	P	11/05/2025	13:25	LB137798
	Barium	14.6	+/-50	U	100	P	11/05/2025	13:25	LB137798
	Beryllium	0.56	+/-3	U	6.00	P	11/05/2025	13:25	LB137798
	Cadmium	0.50	+/-3	U	6.00	P	11/05/2025	13:25	LB137798
	Calcium	234	+/-1000	U	2000	P	11/05/2025	13:25	LB137798
	Chromium	2.12	+/-5	U	10.0	P	11/05/2025	13:25	LB137798
	Cobalt	2.26	+/-15	U	30.0	P	11/05/2025	13:25	LB137798
	Copper	4.60	+/-10	U	20.0	P	11/05/2025	13:25	LB137798
	Iron	23.4	+/-50	U	100	P	11/05/2025	13:25	LB137798
	Lead	2.30	+/-6	U	12.0	P	11/05/2025	13:25	LB137798
	Magnesium	244	+/-1000	U	2000	P	11/05/2025	13:25	LB137798
	Manganese	5.94	+/-10	U	20.0	P	11/05/2025	13:25	LB137798
	Nickel	3.06	+/-20	U	40.0	P	11/05/2025	13:25	LB137798
	Potassium	918	+/-1000	U	2000	P	11/05/2025	13:25	LB137798
	Selenium	9.64	+/-10	U	20.0	P	11/05/2025	13:25	LB137798

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick Engineers

SDG No.: Q3534

Contract: REMI02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	1.62	+/-5	U	10.0	P	11/05/2025	13:25	LB137798
	Sodium	868	+/-1000	U	2000	P	11/05/2025	13:25	LB137798
	Thallium	4.38	+/-20	U	40.0	P	11/05/2025	13:25	LB137798
	Vanadium	6.26	+/-20	U	40.0	P	11/05/2025	13:25	LB137798
	Zinc	16.7	+/-20	U	40.0	P	11/05/2025	13:25	LB137798
CCB03	Aluminum	11.3	+/-50	U	100	P	11/05/2025	14:18	LB137798
	Antimony	6.76	+/-25	U	50.0	P	11/05/2025	14:18	LB137798
	Arsenic	5.12	+/-10	U	20.0	P	11/05/2025	14:18	LB137798
	Barium	14.6	+/-50	U	100	P	11/05/2025	14:18	LB137798
	Beryllium	0.56	+/-3	U	6.00	P	11/05/2025	14:18	LB137798
	Cadmium	0.50	+/-3	U	6.00	P	11/05/2025	14:18	LB137798
	Calcium	234	+/-1000	U	2000	P	11/05/2025	14:18	LB137798
	Chromium	2.12	+/-5	U	10.0	P	11/05/2025	14:18	LB137798
	Cobalt	2.26	+/-15	U	30.0	P	11/05/2025	14:18	LB137798
	Copper	4.60	+/-10	U	20.0	P	11/05/2025	14:18	LB137798
	Iron	23.4	+/-50	U	100	P	11/05/2025	14:18	LB137798
	Lead	2.30	+/-6	U	12.0	P	11/05/2025	14:18	LB137798
	Magnesium	244	+/-1000	U	2000	P	11/05/2025	14:18	LB137798
	Manganese	5.94	+/-10	U	20.0	P	11/05/2025	14:18	LB137798
	Nickel	3.06	+/-20	U	40.0	P	11/05/2025	14:18	LB137798
	Potassium	918	+/-1000	U	2000	P	11/05/2025	14:18	LB137798
	Selenium	9.64	+/-10	U	20.0	P	11/05/2025	14:18	LB137798
	Silver	1.62	+/-5	U	10.0	P	11/05/2025	14:18	LB137798
	Sodium	868	+/-1000	U	2000	P	11/05/2025	14:18	LB137798
	Thallium	4.38	+/-20	U	40.0	P	11/05/2025	14:18	LB137798
	Vanadium	6.26	+/-20	U	40.0	P	11/05/2025	14:18	LB137798
	Zinc	16.7	+/-20	U	40.0	P	11/05/2025	14:18	LB137798
CCB04	Aluminum	11.3	+/-50	U	100	P	11/05/2025	15:21	LB137798
	Antimony	6.76	+/-25	U	50.0	P	11/05/2025	15:21	LB137798
	Arsenic	5.12	+/-10	U	20.0	P	11/05/2025	15:21	LB137798
	Barium	14.6	+/-50	U	100	P	11/05/2025	15:21	LB137798
	Beryllium	0.56	+/-3	U	6.00	P	11/05/2025	15:21	LB137798
	Cadmium	0.50	+/-3	U	6.00	P	11/05/2025	15:21	LB137798
	Calcium	234	+/-1000	U	2000	P	11/05/2025	15:21	LB137798
	Chromium	2.12	+/-5	U	10.0	P	11/05/2025	15:21	LB137798
	Cobalt	2.26	+/-15	U	30.0	P	11/05/2025	15:21	LB137798
	Copper	4.60	+/-10	U	20.0	P	11/05/2025	15:21	LB137798
	Iron	23.4	+/-50	U	100	P	11/05/2025	15:21	LB137798
	Lead	2.30	+/-6	U	12.0	P	11/05/2025	15:21	LB137798

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick Engineers

SDG No.: Q3534

Contract: REMI02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	244	+/-1000	U	2000	P	11/05/2025	15:21	LB137798
	Manganese	5.94	+/-10	U	20.0	P	11/05/2025	15:21	LB137798
	Nickel	3.06	+/-20	U	40.0	P	11/05/2025	15:21	LB137798
	Potassium	918	+/-1000	U	2000	P	11/05/2025	15:21	LB137798
	Selenium	9.64	+/-10	U	20.0	P	11/05/2025	15:21	LB137798
	Silver	1.62	+/-5	U	10.0	P	11/05/2025	15:21	LB137798
	Sodium	868	+/-1000	U	2000	P	11/05/2025	15:21	LB137798
	Thallium	4.38	+/-20	U	40.0	P	11/05/2025	15:21	LB137798
	Vanadium	6.26	+/-20	U	40.0	P	11/05/2025	15:21	LB137798
	Zinc	16.7	+/-20	U	40.0	P	11/05/2025	15:21	LB137798
CCB05	Aluminum	11.3	+/-50	U	100	P	11/05/2025	16:25	LB137798
	Antimony	6.76	+/-25	U	50.0	P	11/05/2025	16:25	LB137798
	Arsenic	5.12	+/-10	U	20.0	P	11/05/2025	16:25	LB137798
	Barium	14.6	+/-50	U	100	P	11/05/2025	16:25	LB137798
	Beryllium	0.56	+/-3	U	6.00	P	11/05/2025	16:25	LB137798
	Cadmium	0.50	+/-3	U	6.00	P	11/05/2025	16:25	LB137798
	Calcium	234	+/-1000	U	2000	P	11/05/2025	16:25	LB137798
	Chromium	2.12	+/-5	U	10.0	P	11/05/2025	16:25	LB137798
	Cobalt	2.26	+/-15	U	30.0	P	11/05/2025	16:25	LB137798
	Copper	4.60	+/-10	U	20.0	P	11/05/2025	16:25	LB137798
	Iron	23.4	+/-50	U	100	P	11/05/2025	16:25	LB137798
	Lead	2.30	+/-6	U	12.0	P	11/05/2025	16:25	LB137798
	Magnesium	244	+/-1000	U	2000	P	11/05/2025	16:25	LB137798
	Manganese	5.94	+/-10	U	20.0	P	11/05/2025	16:25	LB137798
	Nickel	3.06	+/-20	U	40.0	P	11/05/2025	16:25	LB137798
	Potassium	918	+/-1000	U	2000	P	11/05/2025	16:25	LB137798
	Selenium	9.64	+/-10	U	20.0	P	11/05/2025	16:25	LB137798
	Silver	1.62	+/-5	U	10.0	P	11/05/2025	16:25	LB137798
	Sodium	868	+/-1000	U	2000	P	11/05/2025	16:25	LB137798
	Thallium	4.38	+/-20	U	40.0	P	11/05/2025	16:25	LB137798
	Vanadium	6.26	+/-20	U	40.0	P	11/05/2025	16:25	LB137798
	Zinc	16.7	+/-20	U	40.0	P	11/05/2025	16:25	LB137798
CCB06	Aluminum	11.3	+/-50	U	100	P	11/05/2025	17:08	LB137798
	Antimony	6.76	+/-25	U	50.0	P	11/05/2025	17:08	LB137798
	Arsenic	5.12	+/-10	U	20.0	P	11/05/2025	17:08	LB137798
	Barium	14.6	+/-50	U	100	P	11/05/2025	17:08	LB137798
	Beryllium	0.56	+/-3	U	6.00	P	11/05/2025	17:08	LB137798
	Cadmium	0.50	+/-3	U	6.00	P	11/05/2025	17:08	LB137798
	Calcium	234	+/-1000	U	2000	P	11/05/2025	17:08	LB137798

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick Engineers

SDG No.: Q3534

Contract: REMI02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Chromium	2.12	+/-5	U	10.0	P	11/05/2025	17:08	LB137798
	Cobalt	2.26	+/-15	U	30.0	P	11/05/2025	17:08	LB137798
	Copper	4.60	+/-10	U	20.0	P	11/05/2025	17:08	LB137798
	Iron	23.4	+/-50	U	100	P	11/05/2025	17:08	LB137798
	Lead	2.30	+/-6	U	12.0	P	11/05/2025	17:08	LB137798
	Magnesium	244	+/-1000	U	2000	P	11/05/2025	17:08	LB137798
	Manganese	5.94	+/-10	U	20.0	P	11/05/2025	17:08	LB137798
	Nickel	3.06	+/-20	U	40.0	P	11/05/2025	17:08	LB137798
	Potassium	1830	+/-1000	J*	2000	P	11/05/2025	17:08	LB137798
	Selenium	9.64	+/-10	U	20.0	P	11/05/2025	17:08	LB137798
	Silver	1.62	+/-5	U	10.0	P	11/05/2025	17:08	LB137798
	Sodium	3440	+/-1000	*	2000	P	11/05/2025	17:08	LB137798
	Thallium	4.38	+/-20	U	40.0	P	11/05/2025	17:08	LB137798
	Vanadium	6.26	+/-20	U	40.0	P	11/05/2025	17:08	LB137798
	Zinc	16.7	+/-20	U	40.0	P	11/05/2025	17:08	LB137798

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick Engineers

SDG No.: Q3534

Contract: REMI02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	25.9	+/-50	J	100	P	11/06/2025	13:25	LB137816
	Antimony	6.76	+/-25	U	50.0	P	11/06/2025	13:25	LB137816
	Arsenic	5.12	+/-10	U	20.0	P	11/06/2025	13:25	LB137816
	Barium	14.6	+/-50	U	100	P	11/06/2025	13:25	LB137816
	Beryllium	0.56	+/-3	U	6.00	P	11/06/2025	13:25	LB137816
	Cadmium	0.50	+/-3	U	6.00	P	11/06/2025	13:25	LB137816
	Calcium	234	+/-1000	U	2000	P	11/06/2025	13:25	LB137816
	Chromium	2.12	+/-5	U	10.0	P	11/06/2025	13:25	LB137816
	Cobalt	2.26	+/-15	U	30.0	P	11/06/2025	13:25	LB137816
	Copper	4.60	+/-10	U	20.0	P	11/06/2025	13:25	LB137816
	Iron	23.4	+/-50	U	100	P	11/06/2025	13:25	LB137816
	Lead	2.30	+/-6	U	12.0	P	11/06/2025	13:25	LB137816
	Magnesium	244	+/-1000	U	2000	P	11/06/2025	13:25	LB137816
	Manganese	5.94	+/-10	U	20.0	P	11/06/2025	13:25	LB137816
	Nickel	3.06	+/-20	U	40.0	P	11/06/2025	13:25	LB137816
	Potassium	918	+/-1000	U	2000	P	11/06/2025	13:25	LB137816
	Selenium	9.64	+/-10	U	20.0	P	11/06/2025	13:25	LB137816
	Silver	1.62	+/-5	U	10.0	P	11/06/2025	13:25	LB137816
	Sodium	868	+/-1000	U	2000	P	11/06/2025	13:25	LB137816
	Thallium	4.38	+/-20	U	40.0	P	11/06/2025	13:25	LB137816
	Vanadium	6.26	+/-20	U	40.0	P	11/06/2025	13:25	LB137816
	Zinc	16.7	+/-20	U	40.0	P	11/06/2025	13:25	LB137816

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick Engineers

SDG No.: Q3534

Contract: REMI02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	13.8	+/-50	J	100	P	11/06/2025	13:54	LB137816
	Antimony	6.76	+/-25	U	50.0	P	11/06/2025	13:54	LB137816
	Arsenic	5.12	+/-10	U	20.0	P	11/06/2025	13:54	LB137816
	Barium	14.6	+/-50	U	100	P	11/06/2025	13:54	LB137816
	Beryllium	0.56	+/-3	U	6.00	P	11/06/2025	13:54	LB137816
	Cadmium	0.50	+/-3	U	6.00	P	11/06/2025	13:54	LB137816
	Calcium	234	+/-1000	U	2000	P	11/06/2025	13:54	LB137816
	Chromium	2.12	+/-5	U	10.0	P	11/06/2025	13:54	LB137816
	Cobalt	2.26	+/-15	U	30.0	P	11/06/2025	13:54	LB137816
	Copper	4.60	+/-10	U	20.0	P	11/06/2025	13:54	LB137816
	Iron	23.4	+/-50	U	100	P	11/06/2025	13:54	LB137816
	Lead	2.30	+/-6	U	12.0	P	11/06/2025	13:54	LB137816
	Magnesium	244	+/-1000	U	2000	P	11/06/2025	13:54	LB137816
	Manganese	5.94	+/-10	U	20.0	P	11/06/2025	13:54	LB137816
	Nickel	3.06	+/-20	U	40.0	P	11/06/2025	13:54	LB137816
	Potassium	918	+/-1000	U	2000	P	11/06/2025	13:54	LB137816
	Selenium	9.64	+/-10	U	20.0	P	11/06/2025	13:54	LB137816
	Silver	1.62	+/-5	U	10.0	P	11/06/2025	13:54	LB137816
	Sodium	868	+/-1000	U	2000	P	11/06/2025	13:54	LB137816
	Thallium	4.38	+/-20	U	40.0	P	11/06/2025	13:54	LB137816
	Vanadium	6.26	+/-20	U	40.0	P	11/06/2025	13:54	LB137816
	Zinc	16.7	+/-20	U	40.0	P	11/06/2025	13:54	LB137816
CCB02	Aluminum	14.7	+/-50	J	100	P	11/06/2025	15:15	LB137816
	Antimony	6.76	+/-25	U	50.0	P	11/06/2025	15:15	LB137816
	Arsenic	5.12	+/-10	U	20.0	P	11/06/2025	15:15	LB137816
	Barium	14.6	+/-50	U	100	P	11/06/2025	15:15	LB137816
	Beryllium	0.56	+/-3	U	6.00	P	11/06/2025	15:15	LB137816
	Cadmium	0.50	+/-3	U	6.00	P	11/06/2025	15:15	LB137816
	Calcium	234	+/-1000	U	2000	P	11/06/2025	15:15	LB137816
	Chromium	2.12	+/-5	U	10.0	P	11/06/2025	15:15	LB137816
	Cobalt	2.26	+/-15	U	30.0	P	11/06/2025	15:15	LB137816
	Copper	4.60	+/-10	U	20.0	P	11/06/2025	15:15	LB137816
	Iron	23.4	+/-50	U	100	P	11/06/2025	15:15	LB137816
	Lead	3.98	+/-6	J	12.0	P	11/06/2025	15:15	LB137816
	Magnesium	244	+/-1000	U	2000	P	11/06/2025	15:15	LB137816
	Manganese	5.94	+/-10	U	20.0	P	11/06/2025	15:15	LB137816
	Nickel	3.06	+/-20	U	40.0	P	11/06/2025	15:15	LB137816
	Potassium	918	+/-1000	U	2000	P	11/06/2025	15:15	LB137816
	Selenium	9.64	+/-10	U	20.0	P	11/06/2025	15:15	LB137816

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick Engineers

SDG No.: Q3534

Contract: REMI02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	1.62	+/-5	U	10.0	P	11/06/2025	15:15	LB137816
	Sodium	868	+/-1000	U	2000	P	11/06/2025	15:15	LB137816
	Thallium	4.38	+/-20	U	40.0	P	11/06/2025	15:15	LB137816
	Vanadium	6.26	+/-20	U	40.0	P	11/06/2025	15:15	LB137816
	Zinc	16.7	+/-20	U	40.0	P	11/06/2025	15:15	LB137816
CCB03	Aluminum	14.2	+/-50	J	100	P	11/06/2025	16:06	LB137816
	Antimony	6.76	+/-25	U	50.0	P	11/06/2025	16:06	LB137816
	Arsenic	6.81	+/-10	J	20.0	P	11/06/2025	16:06	LB137816
	Barium	14.6	+/-50	U	100	P	11/06/2025	16:06	LB137816
	Beryllium	0.56	+/-3	U	6.00	P	11/06/2025	16:06	LB137816
	Cadmium	0.50	+/-3	U	6.00	P	11/06/2025	16:06	LB137816
	Calcium	234	+/-1000	U	2000	P	11/06/2025	16:06	LB137816
	Chromium	2.12	+/-5	U	10.0	P	11/06/2025	16:06	LB137816
	Cobalt	2.26	+/-15	U	30.0	P	11/06/2025	16:06	LB137816
	Copper	4.60	+/-10	U	20.0	P	11/06/2025	16:06	LB137816
	Iron	23.4	+/-50	U	100	P	11/06/2025	16:06	LB137816
	Lead	2.30	+/-6	U	12.0	P	11/06/2025	16:06	LB137816
	Magnesium	244	+/-1000	U	2000	P	11/06/2025	16:06	LB137816
	Manganese	5.94	+/-10	U	20.0	P	11/06/2025	16:06	LB137816
	Nickel	3.06	+/-20	U	40.0	P	11/06/2025	16:06	LB137816
	Potassium	918	+/-1000	U	2000	P	11/06/2025	16:06	LB137816
	Selenium	9.64	+/-10	U	20.0	P	11/06/2025	16:06	LB137816
	Silver	1.62	+/-5	U	10.0	P	11/06/2025	16:06	LB137816
	Sodium	868	+/-1000	U	2000	P	11/06/2025	16:06	LB137816
	Thallium	4.38	+/-20	U	40.0	P	11/06/2025	16:06	LB137816
	Vanadium	6.26	+/-20	U	40.0	P	11/06/2025	16:06	LB137816
	Zinc	16.7	+/-20	U	40.0	P	11/06/2025	16:06	LB137816
CCB04	Aluminum	11.3	+/-50	U	100	P	11/06/2025	16:54	LB137816
	Antimony	6.76	+/-25	U	50.0	P	11/06/2025	16:54	LB137816
	Arsenic	6.57	+/-10	J	20.0	P	11/06/2025	16:54	LB137816
	Barium	14.6	+/-50	U	100	P	11/06/2025	16:54	LB137816
	Beryllium	0.56	+/-3	U	6.00	P	11/06/2025	16:54	LB137816
	Cadmium	0.50	+/-3	U	6.00	P	11/06/2025	16:54	LB137816
	Calcium	234	+/-1000	U	2000	P	11/06/2025	16:54	LB137816
	Chromium	2.12	+/-5	U	10.0	P	11/06/2025	16:54	LB137816
	Cobalt	2.26	+/-15	U	30.0	P	11/06/2025	16:54	LB137816
	Copper	4.60	+/-10	U	20.0	P	11/06/2025	16:54	LB137816
	Iron	23.4	+/-50	U	100	P	11/06/2025	16:54	LB137816
	Lead	2.49	+/-6	J	12.0	P	11/06/2025	16:54	LB137816

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick Engineers

SDG No.: Q3534

Contract: REMI02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	244	+/-1000	U	2000	P	11/06/2025	16:54	LB137816
	Manganese	5.94	+/-10	U	20.0	P	11/06/2025	16:54	LB137816
	Nickel	3.06	+/-20	U	40.0	P	11/06/2025	16:54	LB137816
	Potassium	918	+/-1000	U	2000	P	11/06/2025	16:54	LB137816
	Selenium	9.64	+/-10	U	20.0	P	11/06/2025	16:54	LB137816
	Silver	1.62	+/-5	U	10.0	P	11/06/2025	16:54	LB137816
	Sodium	868	+/-1000	U	2000	P	11/06/2025	16:54	LB137816
	Thallium	4.38	+/-20	U	40.0	P	11/06/2025	16:54	LB137816
	Vanadium	6.26	+/-20	U	40.0	P	11/06/2025	16:54	LB137816
	Zinc	16.7	+/-20	U	40.0	P	11/06/2025	16:54	LB137816
CCB05	Aluminum	31.6	+/-50	J	100	P	11/06/2025	17:44	LB137816
	Antimony	6.76	+/-25	U	50.0	P	11/06/2025	17:44	LB137816
	Arsenic	5.74	+/-10	J	20.0	P	11/06/2025	17:44	LB137816
	Barium	14.6	+/-50	U	100	P	11/06/2025	17:44	LB137816
	Beryllium	0.56	+/-3	U	6.00	P	11/06/2025	17:44	LB137816
	Cadmium	0.50	+/-3	U	6.00	P	11/06/2025	17:44	LB137816
	Calcium	234	+/-1000	U	2000	P	11/06/2025	17:44	LB137816
	Chromium	2.12	+/-5	U	10.0	P	11/06/2025	17:44	LB137816
	Cobalt	2.26	+/-15	U	30.0	P	11/06/2025	17:44	LB137816
	Copper	4.60	+/-10	U	20.0	P	11/06/2025	17:44	LB137816
	Iron	23.4	+/-50	U	100	P	11/06/2025	17:44	LB137816
	Lead	2.30	+/-6	U	12.0	P	11/06/2025	17:44	LB137816
	Magnesium	244	+/-1000	U	2000	P	11/06/2025	17:44	LB137816
	Manganese	5.94	+/-10	U	20.0	P	11/06/2025	17:44	LB137816
	Nickel	3.06	+/-20	U	40.0	P	11/06/2025	17:44	LB137816
	Potassium	918	+/-1000	U	2000	P	11/06/2025	17:44	LB137816
	Selenium	9.64	+/-10	U	20.0	P	11/06/2025	17:44	LB137816
	Silver	1.62	+/-5	U	10.0	P	11/06/2025	17:44	LB137816
	Sodium	868	+/-1000	U	2000	P	11/06/2025	17:44	LB137816
	Thallium	4.38	+/-20	U	40.0	P	11/06/2025	17:44	LB137816
	Vanadium	6.26	+/-20	U	40.0	P	11/06/2025	17:44	LB137816
	Zinc	16.7	+/-20	U	40.0	P	11/06/2025	17:44	LB137816
CCB06	Aluminum	21.1	+/-50	J	100	P	11/06/2025	18:33	LB137816
	Antimony	6.76	+/-25	U	50.0	P	11/06/2025	18:33	LB137816
	Arsenic	6.53	+/-10	J	20.0	P	11/06/2025	18:33	LB137816
	Barium	14.6	+/-50	U	100	P	11/06/2025	18:33	LB137816
	Beryllium	0.56	+/-3	U	6.00	P	11/06/2025	18:33	LB137816
	Cadmium	0.50	+/-3	U	6.00	P	11/06/2025	18:33	LB137816
	Calcium	234	+/-1000	U	2000	P	11/06/2025	18:33	LB137816

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick Engineers

SDG No.: Q3534

Contract: REMI02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Chromium	2.12	+/-5	U	10.0	P	11/06/2025	18:33	LB137816
	Cobalt	2.26	+/-15	U	30.0	P	11/06/2025	18:33	LB137816
	Copper	4.60	+/-10	U	20.0	P	11/06/2025	18:33	LB137816
	Iron	23.4	+/-50	U	100	P	11/06/2025	18:33	LB137816
	Lead	2.30	+/-6	U	12.0	P	11/06/2025	18:33	LB137816
	Magnesium	244	+/-1000	U	2000	P	11/06/2025	18:33	LB137816
	Manganese	5.94	+/-10	U	20.0	P	11/06/2025	18:33	LB137816
	Nickel	3.06	+/-20	U	40.0	P	11/06/2025	18:33	LB137816
	Potassium	918	+/-1000	U	2000	P	11/06/2025	18:33	LB137816
	Selenium	9.64	+/-10	U	20.0	P	11/06/2025	18:33	LB137816
	Silver	1.62	+/-5	U	10.0	P	11/06/2025	18:33	LB137816
	Sodium	868	+/-1000	U	2000	P	11/06/2025	18:33	LB137816
	Thallium	4.38	+/-20	U	40.0	P	11/06/2025	18:33	LB137816
	Vanadium	6.26	+/-20	U	40.0	P	11/06/2025	18:33	LB137816
	Zinc	16.7	+/-20	U	40.0	P	11/06/2025	18:33	LB137816
CCB07	Aluminum	32.5	+/-50	J	100	P	11/06/2025	19:15	LB137816
	Antimony	6.76	+/-25	U	50.0	P	11/06/2025	19:15	LB137816
	Arsenic	5.12	+/-10	U	20.0	P	11/06/2025	19:15	LB137816
	Barium	14.6	+/-50	U	100	P	11/06/2025	19:15	LB137816
	Beryllium	0.56	+/-3	U	6.00	P	11/06/2025	19:15	LB137816
	Cadmium	0.50	+/-3	U	6.00	P	11/06/2025	19:15	LB137816
	Calcium	234	+/-1000	U	2000	P	11/06/2025	19:15	LB137816
	Chromium	2.12	+/-5	U	10.0	P	11/06/2025	19:15	LB137816
	Cobalt	2.26	+/-15	U	30.0	P	11/06/2025	19:15	LB137816
	Copper	4.60	+/-10	U	20.0	P	11/06/2025	19:15	LB137816
	Iron	23.4	+/-50	U	100	P	11/06/2025	19:15	LB137816
	Lead	2.30	+/-6	U	12.0	P	11/06/2025	19:15	LB137816
	Magnesium	244	+/-1000	U	2000	P	11/06/2025	19:15	LB137816
	Manganese	5.94	+/-10	U	20.0	P	11/06/2025	19:15	LB137816
	Nickel	3.06	+/-20	U	40.0	P	11/06/2025	19:15	LB137816
	Potassium	918	+/-1000	U	2000	P	11/06/2025	19:15	LB137816
	Selenium	9.64	+/-10	U	20.0	P	11/06/2025	19:15	LB137816
	Silver	1.62	+/-5	U	10.0	P	11/06/2025	19:15	LB137816
	Sodium	868	+/-1000	U	2000	P	11/06/2025	19:15	LB137816
	Thallium	4.38	+/-20	U	40.0	P	11/06/2025	19:15	LB137816
	Vanadium	6.26	+/-20	U	40.0	P	11/06/2025	19:15	LB137816
	Zinc	16.7	+/-20	U	40.0	P	11/06/2025	19:15	LB137816

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick Engineers

SDG No.: Q3534

Contract: REMI02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	11.3	+/-50	U	100	P	11/07/2025	13:42	LB137834
	Antimony	6.76	+/-25	U	50.0	P	11/07/2025	13:42	LB137834
	Arsenic	5.12	+/-10	U	20.0	P	11/07/2025	13:42	LB137834
	Barium	14.6	+/-50	U	100	P	11/07/2025	13:42	LB137834
	Beryllium	0.56	+/-3	U	6.00	P	11/07/2025	13:42	LB137834
	Cadmium	0.50	+/-3	U	6.00	P	11/07/2025	13:42	LB137834
	Calcium	234	+/-1000	U	2000	P	11/07/2025	13:42	LB137834
	Chromium	2.12	+/-5	U	10.0	P	11/07/2025	13:42	LB137834
	Cobalt	2.26	+/-15	U	30.0	P	11/07/2025	13:42	LB137834
	Copper	4.60	+/-10	U	20.0	P	11/07/2025	13:42	LB137834
	Iron	23.4	+/-50	U	100	P	11/07/2025	13:42	LB137834
	Lead	2.30	+/-6	U	12.0	P	11/07/2025	13:42	LB137834
	Magnesium	244	+/-1000	U	2000	P	11/07/2025	13:42	LB137834
	Manganese	5.94	+/-10	U	20.0	P	11/07/2025	13:42	LB137834
	Nickel	3.06	+/-20	U	40.0	P	11/07/2025	13:42	LB137834
	Potassium	918	+/-1000	U	2000	P	11/07/2025	13:42	LB137834
	Selenium	9.64	+/-10	U	20.0	P	11/07/2025	13:42	LB137834
	Silver	1.62	+/-5	U	10.0	P	11/07/2025	13:42	LB137834
	Sodium	868	+/-1000	U	2000	P	11/07/2025	13:42	LB137834
	Thallium	4.38	+/-20	U	40.0	P	11/07/2025	13:42	LB137834
	Vanadium	8.59	+/-20	J	40.0	P	11/07/2025	13:42	LB137834
	Zinc	16.7	+/-20	U	40.0	P	11/07/2025	13:42	LB137834

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick Engineers

SDG No.: Q3534

Contract: REMI02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	26.8	+/-50	J	100	P	11/07/2025	14:17	LB137834
	Antimony	6.76	+/-25	U	50.0	P	11/07/2025	14:17	LB137834
	Arsenic	5.12	+/-10	U	20.0	P	11/07/2025	14:17	LB137834
	Barium	14.6	+/-50	U	100	P	11/07/2025	14:17	LB137834
	Beryllium	0.56	+/-3	U	6.00	P	11/07/2025	14:17	LB137834
	Cadmium	0.50	+/-3	U	6.00	P	11/07/2025	14:17	LB137834
	Calcium	234	+/-1000	U	2000	P	11/07/2025	14:17	LB137834
	Chromium	2.12	+/-5	U	10.0	P	11/07/2025	14:17	LB137834
	Cobalt	2.26	+/-15	U	30.0	P	11/07/2025	14:17	LB137834
	Copper	4.60	+/-10	U	20.0	P	11/07/2025	14:17	LB137834
	Iron	39.8	+/-50	J	100	P	11/07/2025	14:17	LB137834
	Lead	2.30	+/-6	U	12.0	P	11/07/2025	14:17	LB137834
	Magnesium	244	+/-1000	U	2000	P	11/07/2025	14:17	LB137834
	Manganese	5.94	+/-10	U	20.0	P	11/07/2025	14:17	LB137834
	Nickel	3.06	+/-20	U	40.0	P	11/07/2025	14:17	LB137834
	Potassium	918	+/-1000	U	2000	P	11/07/2025	14:17	LB137834
	Selenium	9.64	+/-10	U	20.0	P	11/07/2025	14:17	LB137834
	Silver	1.62	+/-5	U	10.0	P	11/07/2025	14:17	LB137834
	Sodium	868	+/-1000	U	2000	P	11/07/2025	14:17	LB137834
	Thallium	9.19	+/-20	J	40.0	P	11/07/2025	14:17	LB137834
	Vanadium	12.3	+/-20	J	40.0	P	11/07/2025	14:17	LB137834
	Zinc	16.7	+/-20	U	40.0	P	11/07/2025	14:17	LB137834
CCB02	Aluminum	11.3	+/-50	U	100	P	11/07/2025	15:14	LB137834
	Antimony	6.76	+/-25	U	50.0	P	11/07/2025	15:14	LB137834
	Arsenic	5.12	+/-10	U	20.0	P	11/07/2025	15:14	LB137834
	Barium	14.6	+/-50	U	100	P	11/07/2025	15:14	LB137834
	Beryllium	0.56	+/-3	U	6.00	P	11/07/2025	15:14	LB137834
	Cadmium	0.50	+/-3	U	6.00	P	11/07/2025	15:14	LB137834
	Calcium	234	+/-1000	U	2000	P	11/07/2025	15:14	LB137834
	Chromium	2.12	+/-5	U	10.0	P	11/07/2025	15:14	LB137834
	Cobalt	2.26	+/-15	U	30.0	P	11/07/2025	15:14	LB137834
	Copper	4.60	+/-10	U	20.0	P	11/07/2025	15:14	LB137834
	Iron	23.4	+/-50	U	100	P	11/07/2025	15:14	LB137834
	Lead	2.30	+/-6	U	12.0	P	11/07/2025	15:14	LB137834
	Magnesium	244	+/-1000	U	2000	P	11/07/2025	15:14	LB137834
	Manganese	5.94	+/-10	U	20.0	P	11/07/2025	15:14	LB137834
	Nickel	3.06	+/-20	U	40.0	P	11/07/2025	15:14	LB137834
	Potassium	918	+/-1000	U	2000	P	11/07/2025	15:14	LB137834
	Selenium	9.64	+/-10	U	20.0	P	11/07/2025	15:14	LB137834

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick Engineers

SDG No.: Q3534

Contract: REMI02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	1.62	+/-5	U	10.0	P	11/07/2025	15:14	LB137834
	Sodium	868	+/-1000	U	2000	P	11/07/2025	15:14	LB137834
	Thallium	10.2	+/-20	J	40.0	P	11/07/2025	15:14	LB137834
	Vanadium	9.57	+/-20	J	40.0	P	11/07/2025	15:14	LB137834
	Zinc	16.7	+/-20	U	40.0	P	11/07/2025	15:14	LB137834
CCB03	Aluminum	11.3	+/-50	U	100	P	11/07/2025	16:07	LB137834
	Antimony	6.76	+/-25	U	50.0	P	11/07/2025	16:07	LB137834
	Arsenic	5.12	+/-10	U	20.0	P	11/07/2025	16:07	LB137834
	Barium	14.6	+/-50	U	100	P	11/07/2025	16:07	LB137834
	Beryllium	0.56	+/-3	U	6.00	P	11/07/2025	16:07	LB137834
	Cadmium	0.50	+/-3	U	6.00	P	11/07/2025	16:07	LB137834
	Calcium	234	+/-1000	U	2000	P	11/07/2025	16:07	LB137834
	Chromium	2.12	+/-5	U	10.0	P	11/07/2025	16:07	LB137834
	Cobalt	2.26	+/-15	U	30.0	P	11/07/2025	16:07	LB137834
	Copper	4.60	+/-10	U	20.0	P	11/07/2025	16:07	LB137834
	Iron	23.4	+/-50	U	100	P	11/07/2025	16:07	LB137834
	Lead	2.30	+/-6	U	12.0	P	11/07/2025	16:07	LB137834
	Magnesium	244	+/-1000	U	2000	P	11/07/2025	16:07	LB137834
	Manganese	5.94	+/-10	U	20.0	P	11/07/2025	16:07	LB137834
	Nickel	3.06	+/-20	U	40.0	P	11/07/2025	16:07	LB137834
	Potassium	918	+/-1000	U	2000	P	11/07/2025	16:07	LB137834
	Selenium	9.64	+/-10	U	20.0	P	11/07/2025	16:07	LB137834
	Silver	1.62	+/-5	U	10.0	P	11/07/2025	16:07	LB137834
	Sodium	868	+/-1000	U	2000	P	11/07/2025	16:07	LB137834
	Thallium	4.38	+/-20	U	40.0	P	11/07/2025	16:07	LB137834
	Vanadium	6.26	+/-20	U	40.0	P	11/07/2025	16:07	LB137834
	Zinc	16.7	+/-20	U	40.0	P	11/07/2025	16:07	LB137834

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick Engineers

SDG No.: Q3534

Contract: REMI02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	11.3	+/-50	U	100	P	11/10/2025	12:11	LB137841
	Antimony	6.76	+/-25	U	50.0	P	11/10/2025	12:11	LB137841
	Arsenic	5.12	+/-10	U	20.0	P	11/10/2025	12:11	LB137841
	Barium	14.6	+/-50	U	100	P	11/10/2025	12:11	LB137841
	Beryllium	0.56	+/-3	U	6.00	P	11/10/2025	12:11	LB137841
	Cadmium	0.50	+/-3	U	6.00	P	11/10/2025	12:11	LB137841
	Calcium	234	+/-1000	U	2000	P	11/10/2025	12:11	LB137841
	Chromium	2.12	+/-5	U	10.0	P	11/10/2025	12:11	LB137841
	Cobalt	2.26	+/-15	U	30.0	P	11/10/2025	12:11	LB137841
	Copper	4.60	+/-10	U	20.0	P	11/10/2025	12:11	LB137841
	Iron	32.9	+/-50	J	100	P	11/10/2025	12:11	LB137841
	Lead	2.30	+/-6	U	12.0	P	11/10/2025	12:11	LB137841
	Magnesium	244	+/-1000	U	2000	P	11/10/2025	12:11	LB137841
	Manganese	5.94	+/-10	U	20.0	P	11/10/2025	12:11	LB137841
	Nickel	3.06	+/-20	U	40.0	P	11/10/2025	12:11	LB137841
	Potassium	918	+/-1000	U	2000	P	11/10/2025	12:11	LB137841
	Selenium	9.64	+/-10	U	20.0	P	11/10/2025	12:11	LB137841
	Silver	1.62	+/-5	U	10.0	P	11/10/2025	12:11	LB137841
	Sodium	868	+/-1000	U	2000	P	11/10/2025	12:11	LB137841
	Thallium	4.38	+/-20	U	40.0	P	11/10/2025	12:11	LB137841
	Vanadium	6.26	+/-20	U	40.0	P	11/10/2025	12:11	LB137841
	Zinc	16.7	+/-20	U	40.0	P	11/10/2025	12:11	LB137841

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick Engineers

SDG No.: Q3534

Contract: REMI02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	14.6	+/-50	J	100	P	11/10/2025	12:42	LB137841
	Antimony	6.76	+/-25	U	50.0	P	11/10/2025	12:42	LB137841
	Arsenic	5.12	+/-10	U	20.0	P	11/10/2025	12:42	LB137841
	Barium	14.6	+/-50	U	100	P	11/10/2025	12:42	LB137841
	Beryllium	0.56	+/-3	U	6.00	P	11/10/2025	12:42	LB137841
	Cadmium	0.50	+/-3	U	6.00	P	11/10/2025	12:42	LB137841
	Calcium	234	+/-1000	U	2000	P	11/10/2025	12:42	LB137841
	Chromium	2.12	+/-5	U	10.0	P	11/10/2025	12:42	LB137841
	Cobalt	2.26	+/-15	U	30.0	P	11/10/2025	12:42	LB137841
	Copper	4.60	+/-10	U	20.0	P	11/10/2025	12:42	LB137841
	Iron	23.4	+/-50	U	100	P	11/10/2025	12:42	LB137841
	Lead	2.30	+/-6	U	12.0	P	11/10/2025	12:42	LB137841
	Magnesium	244	+/-1000	U	2000	P	11/10/2025	12:42	LB137841
	Manganese	5.94	+/-10	U	20.0	P	11/10/2025	12:42	LB137841
	Nickel	3.06	+/-20	U	40.0	P	11/10/2025	12:42	LB137841
	Potassium	918	+/-1000	U	2000	P	11/10/2025	12:42	LB137841
	Selenium	9.64	+/-10	U	20.0	P	11/10/2025	12:42	LB137841
	Silver	1.62	+/-5	U	10.0	P	11/10/2025	12:42	LB137841
	Sodium	868	+/-1000	U	2000	P	11/10/2025	12:42	LB137841
	Thallium	4.38	+/-20	U	40.0	P	11/10/2025	12:42	LB137841
	Vanadium	6.26	+/-20	U	40.0	P	11/10/2025	12:42	LB137841
	Zinc	16.7	+/-20	U	40.0	P	11/10/2025	12:42	LB137841
CCB02	Aluminum	20.4	+/-50	J	100	P	11/10/2025	13:54	LB137841
	Antimony	6.76	+/-25	U	50.0	P	11/10/2025	13:54	LB137841
	Arsenic	5.12	+/-10	U	20.0	P	11/10/2025	13:54	LB137841
	Barium	14.6	+/-50	U	100	P	11/10/2025	13:54	LB137841
	Beryllium	0.56	+/-3	U	6.00	P	11/10/2025	13:54	LB137841
	Cadmium	0.50	+/-3	U	6.00	P	11/10/2025	13:54	LB137841
	Calcium	234	+/-1000	U	2000	P	11/10/2025	13:54	LB137841
	Chromium	2.12	+/-5	U	10.0	P	11/10/2025	13:54	LB137841
	Cobalt	2.26	+/-15	U	30.0	P	11/10/2025	13:54	LB137841
	Copper	4.60	+/-10	U	20.0	P	11/10/2025	13:54	LB137841
	Iron	29.8	+/-50	J	100	P	11/10/2025	13:54	LB137841
	Lead	2.30	+/-6	U	12.0	P	11/10/2025	13:54	LB137841
	Magnesium	244	+/-1000	U	2000	P	11/10/2025	13:54	LB137841
	Manganese	5.94	+/-10	U	20.0	P	11/10/2025	13:54	LB137841
	Nickel	3.06	+/-20	U	40.0	P	11/10/2025	13:54	LB137841
	Potassium	918	+/-1000	U	2000	P	11/10/2025	13:54	LB137841
	Selenium	9.64	+/-10	U	20.0	P	11/10/2025	13:54	LB137841

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick Engineers

SDG No.: Q3534

Contract: REMI02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	1.62	+/-5	U	10.0	P	11/10/2025	13:54	LB137841
	Sodium	868	+/-1000	U	2000	P	11/10/2025	13:54	LB137841
	Thallium	4.43	+/-20	J	40.0	P	11/10/2025	13:54	LB137841
	Vanadium	6.26	+/-20	U	40.0	P	11/10/2025	13:54	LB137841
	Zinc	16.7	+/-20	U	40.0	P	11/10/2025	13:54	LB137841
CCB03	Aluminum	11.3	+/-50	U	100	P	11/10/2025	14:50	LB137841
	Antimony	6.76	+/-25	U	50.0	P	11/10/2025	14:50	LB137841
	Arsenic	5.12	+/-10	U	20.0	P	11/10/2025	14:50	LB137841
	Barium	14.6	+/-50	U	100	P	11/10/2025	14:50	LB137841
	Beryllium	0.56	+/-3	U	6.00	P	11/10/2025	14:50	LB137841
	Cadmium	0.50	+/-3	U	6.00	P	11/10/2025	14:50	LB137841
	Calcium	234	+/-1000	U	2000	P	11/10/2025	14:50	LB137841
	Chromium	2.12	+/-5	U	10.0	P	11/10/2025	14:50	LB137841
	Cobalt	2.26	+/-15	U	30.0	P	11/10/2025	14:50	LB137841
	Copper	4.60	+/-10	U	20.0	P	11/10/2025	14:50	LB137841
	Iron	29.3	+/-50	J	100	P	11/10/2025	14:50	LB137841
	Lead	2.30	+/-6	U	12.0	P	11/10/2025	14:50	LB137841
	Magnesium	244	+/-1000	U	2000	P	11/10/2025	14:50	LB137841
	Manganese	5.94	+/-10	U	20.0	P	11/10/2025	14:50	LB137841
	Nickel	3.06	+/-20	U	40.0	P	11/10/2025	14:50	LB137841
	Potassium	918	+/-1000	U	2000	P	11/10/2025	14:50	LB137841
	Selenium	9.64	+/-10	U	20.0	P	11/10/2025	14:50	LB137841
	Silver	1.62	+/-5	U	10.0	P	11/10/2025	14:50	LB137841
	Sodium	868	+/-1000	U	2000	P	11/10/2025	14:50	LB137841
	Thallium	4.38	+/-20	U	40.0	P	11/10/2025	14:50	LB137841
	Vanadium	6.26	+/-20	U	40.0	P	11/10/2025	14:50	LB137841
	Zinc	16.7	+/-20	U	40.0	P	11/10/2025	14:50	LB137841
CCB04	Aluminum	37.9	+/-50	J	100	P	11/10/2025	15:47	LB137841
	Antimony	6.76	+/-25	U	50.0	P	11/10/2025	15:47	LB137841
	Arsenic	5.12	+/-10	U	20.0	P	11/10/2025	15:47	LB137841
	Barium	14.6	+/-50	U	100	P	11/10/2025	15:47	LB137841
	Beryllium	0.56	+/-3	U	6.00	P	11/10/2025	15:47	LB137841
	Cadmium	0.50	+/-3	U	6.00	P	11/10/2025	15:47	LB137841
	Calcium	234	+/-1000	U	2000	P	11/10/2025	15:47	LB137841
	Chromium	2.12	+/-5	U	10.0	P	11/10/2025	15:47	LB137841
	Cobalt	2.26	+/-15	U	30.0	P	11/10/2025	15:47	LB137841
	Copper	4.60	+/-10	U	20.0	P	11/10/2025	15:47	LB137841
	Iron	37.7	+/-50	J	100	P	11/10/2025	15:47	LB137841
	Lead	2.30	+/-6	U	12.0	P	11/10/2025	15:47	LB137841

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick Engineers

SDG No.: Q3534

Contract: REMI02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	244	+/-1000	U	2000	P	11/10/2025	15:47	LB137841
	Manganese	5.94	+/-10	U	20.0	P	11/10/2025	15:47	LB137841
	Nickel	3.06	+/-20	U	40.0	P	11/10/2025	15:47	LB137841
	Potassium	918	+/-1000	U	2000	P	11/10/2025	15:47	LB137841
	Selenium	9.64	+/-10	U	20.0	P	11/10/2025	15:47	LB137841
	Silver	1.62	+/-5	U	10.0	P	11/10/2025	15:47	LB137841
	Sodium	868	+/-1000	U	2000	P	11/10/2025	15:47	LB137841
	Thallium	4.38	+/-20	U	40.0	P	11/10/2025	15:47	LB137841
	Vanadium	6.26	+/-20	U	40.0	P	11/10/2025	15:47	LB137841
	Zinc	16.7	+/-20	U	40.0	P	11/10/2025	15:47	LB137841
CCB05	Aluminum	19.1	+/-50	J	100	P	11/10/2025	16:38	LB137841
	Antimony	6.76	+/-25	U	50.0	P	11/10/2025	16:38	LB137841
	Arsenic	5.12	+/-10	U	20.0	P	11/10/2025	16:38	LB137841
	Barium	14.6	+/-50	U	100	P	11/10/2025	16:38	LB137841
	Beryllium	0.56	+/-3	U	6.00	P	11/10/2025	16:38	LB137841
	Cadmium	0.50	+/-3	U	6.00	P	11/10/2025	16:38	LB137841
	Calcium	234	+/-1000	U	2000	P	11/10/2025	16:38	LB137841
	Chromium	2.12	+/-5	U	10.0	P	11/10/2025	16:38	LB137841
	Cobalt	2.26	+/-15	U	30.0	P	11/10/2025	16:38	LB137841
	Copper	4.60	+/-10	U	20.0	P	11/10/2025	16:38	LB137841
	Iron	24.8	+/-50	J	100	P	11/10/2025	16:38	LB137841
	Lead	2.30	+/-6	U	12.0	P	11/10/2025	16:38	LB137841
	Magnesium	244	+/-1000	U	2000	P	11/10/2025	16:38	LB137841
	Manganese	5.94	+/-10	U	20.0	P	11/10/2025	16:38	LB137841
	Nickel	3.06	+/-20	U	40.0	P	11/10/2025	16:38	LB137841
	Potassium	918	+/-1000	U	2000	P	11/10/2025	16:38	LB137841
	Selenium	9.64	+/-10	U	20.0	P	11/10/2025	16:38	LB137841
	Silver	1.62	+/-5	U	10.0	P	11/10/2025	16:38	LB137841
	Sodium	868	+/-1000	U	2000	P	11/10/2025	16:38	LB137841
	Thallium	4.38	+/-20	U	40.0	P	11/10/2025	16:38	LB137841
	Vanadium	6.26	+/-20	U	40.0	P	11/10/2025	16:38	LB137841
	Zinc	16.7	+/-20	U	40.0	P	11/10/2025	16:38	LB137841
CCB06	Aluminum	23.5	+/-50	J	100	P	11/10/2025	17:31	LB137841
	Antimony	6.76	+/-25	U	50.0	P	11/10/2025	17:31	LB137841
	Arsenic	7.21	+/-10	J	20.0	P	11/10/2025	17:31	LB137841
	Barium	14.6	+/-50	U	100	P	11/10/2025	17:31	LB137841
	Beryllium	0.56	+/-3	U	6.00	P	11/10/2025	17:31	LB137841
	Cadmium	0.50	+/-3	U	6.00	P	11/10/2025	17:31	LB137841
	Calcium	234	+/-1000	U	2000	P	11/10/2025	17:31	LB137841

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick Engineers

SDG No.: Q3534

Contract: REMI02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Chromium	2.12	+/-5	U	10.0	P	11/10/2025	17:31	LB137841
	Cobalt	2.26	+/-15	U	30.0	P	11/10/2025	17:31	LB137841
	Copper	4.60	+/-10	U	20.0	P	11/10/2025	17:31	LB137841
	Iron	24.6	+/-50	J	100	P	11/10/2025	17:31	LB137841
	Lead	2.30	+/-6	U	12.0	P	11/10/2025	17:31	LB137841
	Magnesium	244	+/-1000	U	2000	P	11/10/2025	17:31	LB137841
	Manganese	5.94	+/-10	U	20.0	P	11/10/2025	17:31	LB137841
	Nickel	3.06	+/-20	U	40.0	P	11/10/2025	17:31	LB137841
	Potassium	918	+/-1000	U	2000	P	11/10/2025	17:31	LB137841
	Selenium	9.64	+/-10	U	20.0	P	11/10/2025	17:31	LB137841
	Silver	1.62	+/-5	U	10.0	P	11/10/2025	17:31	LB137841
	Sodium	868	+/-1000	U	2000	P	11/10/2025	17:31	LB137841
	Thallium	4.38	+/-20	U	40.0	P	11/10/2025	17:31	LB137841
	Vanadium	6.26	+/-20	U	40.0	P	11/10/2025	17:31	LB137841
	Zinc	16.7	+/-20	U	40.0	P	11/10/2025	17:31	LB137841
CCB07	Aluminum	18.8	+/-50	J	100	P	11/10/2025	19:18	LB137841
	Antimony	6.76	+/-25	U	50.0	P	11/10/2025	19:18	LB137841
	Arsenic	5.12	+/-10	U	20.0	P	11/10/2025	19:18	LB137841
	Barium	14.6	+/-50	U	100	P	11/10/2025	19:18	LB137841
	Beryllium	0.56	+/-3	U	6.00	P	11/10/2025	19:18	LB137841
	Cadmium	0.50	+/-3	U	6.00	P	11/10/2025	19:18	LB137841
	Calcium	234	+/-1000	U	2000	P	11/10/2025	19:18	LB137841
	Chromium	2.12	+/-5	U	10.0	P	11/10/2025	19:18	LB137841
	Cobalt	2.26	+/-15	U	30.0	P	11/10/2025	19:18	LB137841
	Copper	4.60	+/-10	U	20.0	P	11/10/2025	19:18	LB137841
	Iron	38.0	+/-50	J	100	P	11/10/2025	19:18	LB137841
	Lead	2.30	+/-6	U	12.0	P	11/10/2025	19:18	LB137841
	Magnesium	244	+/-1000	U	2000	P	11/10/2025	19:18	LB137841
	Manganese	5.94	+/-10	U	20.0	P	11/10/2025	19:18	LB137841
	Nickel	3.06	+/-20	U	40.0	P	11/10/2025	19:18	LB137841
	Potassium	918	+/-1000	U	2000	P	11/10/2025	19:18	LB137841
	Selenium	9.64	+/-10	U	20.0	P	11/10/2025	19:18	LB137841
	Silver	1.62	+/-5	U	10.0	P	11/10/2025	19:18	LB137841
	Sodium	868	+/-1000	U	2000	P	11/10/2025	19:18	LB137841
	Thallium	4.38	+/-20	U	40.0	P	11/10/2025	19:18	LB137841
	Vanadium	6.26	+/-20	U	40.0	P	11/10/2025	19:18	LB137841
	Zinc	16.7	+/-20	U	40.0	P	11/10/2025	19:18	LB137841
CCB08	Aluminum	49.0	+/-50	J	100	P	11/10/2025	20:26	LB137841
	Antimony	6.76	+/-25	U	50.0	P	11/10/2025	20:26	LB137841

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick Engineers

SDG No.: Q3534

Contract: REMI02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB08	Arsenic	5.12	+/-10	U	20.0	P	11/10/2025	20:26	LB137841
	Barium	14.6	+/-50	U	100	P	11/10/2025	20:26	LB137841
	Beryllium	0.56	+/-3	U	6.00	P	11/10/2025	20:26	LB137841
	Cadmium	0.50	+/-3	U	6.00	P	11/10/2025	20:26	LB137841
	Calcium	234	+/-1000	U	2000	P	11/10/2025	20:26	LB137841
	Chromium	2.12	+/-5	U	10.0	P	11/10/2025	20:26	LB137841
	Cobalt	2.26	+/-15	U	30.0	P	11/10/2025	20:26	LB137841
	Copper	4.60	+/-10	U	20.0	P	11/10/2025	20:26	LB137841
	Iron	27.2	+/-50	J	100	P	11/10/2025	20:26	LB137841
	Lead	2.30	+/-6	U	12.0	P	11/10/2025	20:26	LB137841
	Magnesium	244	+/-1000	U	2000	P	11/10/2025	20:26	LB137841
	Manganese	5.94	+/-10	U	20.0	P	11/10/2025	20:26	LB137841
	Nickel	3.06	+/-20	U	40.0	P	11/10/2025	20:26	LB137841
	Potassium	918	+/-1000	U	2000	P	11/10/2025	20:26	LB137841
	Selenium	9.64	+/-10	U	20.0	P	11/10/2025	20:26	LB137841
	Silver	1.62	+/-5	U	10.0	P	11/10/2025	20:26	LB137841
	Sodium	868	+/-1000	U	2000	P	11/10/2025	20:26	LB137841
	Thallium	4.38	+/-20	U	40.0	P	11/10/2025	20:26	LB137841
	Vanadium	6.26	+/-20	U	40.0	P	11/10/2025	20:26	LB137841
	Zinc	16.7	+/-20	U	40.0	P	11/10/2025	20:26	LB137841
CCB09	Aluminum	11.3	+/-50	U	100	P	11/10/2025	21:41	LB137841
	Antimony	6.76	+/-25	U	50.0	P	11/10/2025	21:41	LB137841
	Arsenic	5.12	+/-10	U	20.0	P	11/10/2025	21:41	LB137841
	Barium	14.6	+/-50	U	100	P	11/10/2025	21:41	LB137841
	Beryllium	0.56	+/-3	U	6.00	P	11/10/2025	21:41	LB137841
	Cadmium	0.50	+/-3	U	6.00	P	11/10/2025	21:41	LB137841
	Calcium	234	+/-1000	U	2000	P	11/10/2025	21:41	LB137841
	Chromium	2.12	+/-5	U	10.0	P	11/10/2025	21:41	LB137841
	Cobalt	2.26	+/-15	U	30.0	P	11/10/2025	21:41	LB137841
	Copper	4.60	+/-10	U	20.0	P	11/10/2025	21:41	LB137841
	Iron	23.4	+/-50	U	100	P	11/10/2025	21:41	LB137841
	Lead	2.30	+/-6	U	12.0	P	11/10/2025	21:41	LB137841
	Magnesium	244	+/-1000	U	2000	P	11/10/2025	21:41	LB137841
	Manganese	5.94	+/-10	U	20.0	P	11/10/2025	21:41	LB137841
	Nickel	3.06	+/-20	U	40.0	P	11/10/2025	21:41	LB137841
	Potassium	918	+/-1000	U	2000	P	11/10/2025	21:41	LB137841
	Selenium	9.64	+/-10	U	20.0	P	11/10/2025	21:41	LB137841
	Silver	1.62	+/-5	U	10.0	P	11/10/2025	21:41	LB137841
	Sodium	868	+/-1000	U	2000	P	11/10/2025	21:41	LB137841

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick Engineers

SDG No.: Q3534

Contract: REMI02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB09	Thallium	4.38	+/-20	U	40.0	P	11/10/2025	21:41	LB137841
	Vanadium	6.26	+/-20	U	40.0	P	11/10/2025	21:41	LB137841
	Zinc	16.7	+/-20	U	40.0	P	11/10/2025	21:41	LB137841
CCB10	Aluminum	21.4	+/-50	J	100	P	11/10/2025	22:47	LB137841
	Antimony	6.76	+/-25	U	50.0	P	11/10/2025	22:47	LB137841
	Arsenic	5.12	+/-10	U	20.0	P	11/10/2025	22:47	LB137841
	Barium	14.6	+/-50	U	100	P	11/10/2025	22:47	LB137841
	Beryllium	0.56	+/-3	U	6.00	P	11/10/2025	22:47	LB137841
	Cadmium	0.50	+/-3	U	6.00	P	11/10/2025	22:47	LB137841
	Calcium	234	+/-1000	U	2000	P	11/10/2025	22:47	LB137841
	Chromium	2.12	+/-5	U	10.0	P	11/10/2025	22:47	LB137841
	Cobalt	2.26	+/-15	U	30.0	P	11/10/2025	22:47	LB137841
	Copper	4.60	+/-10	U	20.0	P	11/10/2025	22:47	LB137841
	Iron	23.4	+/-50	U	100	P	11/10/2025	22:47	LB137841
	Lead	2.30	+/-6	U	12.0	P	11/10/2025	22:47	LB137841
	Magnesium	244	+/-1000	U	2000	P	11/10/2025	22:47	LB137841
	Manganese	5.94	+/-10	U	20.0	P	11/10/2025	22:47	LB137841
	Nickel	3.06	+/-20	U	40.0	P	11/10/2025	22:47	LB137841
	Potassium	918	+/-1000	U	2000	P	11/10/2025	22:47	LB137841
	Selenium	9.64	+/-10	U	20.0	P	11/10/2025	22:47	LB137841
	Silver	1.62	+/-5	U	10.0	P	11/10/2025	22:47	LB137841
	Sodium	868	+/-1000	U	2000	P	11/10/2025	22:47	LB137841
	Thallium	4.38	+/-20	U	40.0	P	11/10/2025	22:47	LB137841
	Vanadium	6.26	+/-20	U	40.0	P	11/10/2025	22:47	LB137841
	Zinc	16.7	+/-20	U	40.0	P	11/10/2025	22:47	LB137841
	Aluminum	39.4	+/-50	J	100	P	11/10/2025	23:34	LB137841
	Antimony	6.76	+/-25	U	50.0	P	11/10/2025	23:34	LB137841
	Arsenic	5.12	+/-10	U	20.0	P	11/10/2025	23:34	LB137841
	Barium	14.6	+/-50	U	100	P	11/10/2025	23:34	LB137841
	Beryllium	0.56	+/-3	U	6.00	P	11/10/2025	23:34	LB137841
	Cadmium	0.50	+/-3	U	6.00	P	11/10/2025	23:34	LB137841
	Calcium	234	+/-1000	U	2000	P	11/10/2025	23:34	LB137841
	Chromium	2.12	+/-5	U	10.0	P	11/10/2025	23:34	LB137841
	Cobalt	2.26	+/-15	U	30.0	P	11/10/2025	23:34	LB137841
	Copper	4.60	+/-10	U	20.0	P	11/10/2025	23:34	LB137841
	Iron	23.4	+/-50	U	100	P	11/10/2025	23:34	LB137841
	Lead	2.30	+/-6	U	12.0	P	11/10/2025	23:34	LB137841
	Magnesium	244	+/-1000	U	2000	P	11/10/2025	23:34	LB137841
	Manganese	5.94	+/-10	U	20.0	P	11/10/2025	23:34	LB137841

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick Engineers

SDG No.: Q3534

Contract: REMI02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB11	Nickel	3.06	+/-20	U	40.0	P	11/10/2025	23:34	LB137841
	Potassium	918	+/-1000	U	2000	P	11/10/2025	23:34	LB137841
	Selenium	9.64	+/-10	U	20.0	P	11/10/2025	23:34	LB137841
	Silver	1.62	+/-5	U	10.0	P	11/10/2025	23:34	LB137841
	Sodium	868	+/-1000	U	2000	P	11/10/2025	23:34	LB137841
	Thallium	4.38	+/-20	U	40.0	P	11/10/2025	23:34	LB137841
	Vanadium	6.26	+/-20	U	40.0	P	11/10/2025	23:34	LB137841
	Zinc	16.7	+/-20	U	40.0	P	11/10/2025	23:34	LB137841