

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Remington & Vernick Engineers

SDG No.: Q3534

Instrument: CV1

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB170407BL		WATER		Batch Number:	PB170407		Prep Date:	11/05/2025	
	Mercury	0.076	<0.2	U	0.20	CV	11/06/2025	08:25	LB137791
Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB170409BL		WATER		Batch Number:	PB170409		Prep Date:	11/06/2025	
	Mercury	0.076	<0.2	U	0.20	CV	11/06/2025	14:25	LB137799

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Remington & Vernick Engineers

SDG No.: Q3534

Instrument: P4

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB170392BL	WATER			Batch Number:	PB170392		Prep Date:	11/05/2025	
	Aluminum	15.0	<25	J	50.0	P	11/05/2025	15:25	LB137798
	Antimony	3.38	<12.5	U	25.0	P	11/05/2025	15:25	LB137798
	Arsenic	2.56	<5	U	10.0	P	11/05/2025	15:25	LB137798
	Barium	7.28	<25	U	50.0	P	11/05/2025	15:25	LB137798
	Beryllium	0.28	<1.5	U	3.00	P	11/05/2025	15:25	LB137798
	Cadmium	0.25	<1.5	U	3.00	P	11/05/2025	15:25	LB137798
	Calcium	117	<500	U	1000	P	11/05/2025	15:25	LB137798
	Chromium	1.06	<2.5	U	5.00	P	11/05/2025	15:25	LB137798
	Cobalt	1.13	<7.5	U	15.0	P	11/05/2025	15:25	LB137798
	Copper	2.30	<5	U	10.0	P	11/05/2025	15:25	LB137798
	Iron	21.3	<25	J	50.0	P	11/05/2025	15:25	LB137798
	Lead	1.15	<3	U	6.00	P	11/05/2025	15:25	LB137798
	Magnesium	122	<500	U	1000	P	11/05/2025	15:25	LB137798
	Manganese	2.97	<5	U	10.0	P	11/05/2025	15:25	LB137798
	Nickel	1.53	<10	U	20.0	P	11/05/2025	15:25	LB137798
	Potassium	459	<500	U	1000	P	11/05/2025	15:25	LB137798
	Selenium	4.82	<5	U	10.0	P	11/05/2025	15:25	LB137798
	Silver	0.81	<2.5	U	5.00	P	11/05/2025	15:25	LB137798
	Sodium	434	<500	U	1000	P	11/05/2025	15:25	LB137798
	Thallium	2.19	<10	U	20.0	P	11/05/2025	15:25	LB137798
	Vanadium	3.13	<10	U	20.0	P	11/05/2025	15:25	LB137798
	Zinc	8.33	<10	U	20.0	P	11/05/2025	15:25	LB137798
Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB170411BL	WATER			Batch Number:	PB170411		Prep Date:	11/06/2025	
	Aluminum	5.67	<25	U	50.0	P	11/07/2025	14:52	LB137834
	Antimony	3.38	<12.5	U	25.0	P	11/07/2025	14:52	LB137834
	Arsenic	2.56	<5	U	10.0	P	11/07/2025	14:52	LB137834
	Barium	7.28	<25	U	50.0	P	11/07/2025	14:52	LB137834
	Beryllium	0.28	<1.5	U	3.00	P	11/07/2025	14:52	LB137834
	Cadmium	0.25	<1.5	U	3.00	P	11/07/2025	14:52	LB137834
	Calcium	117	<500	U	1000	P	11/07/2025	14:52	LB137834
	Chromium	1.06	<2.5	U	5.00	P	11/07/2025	14:52	LB137834
	Cobalt	1.13	<7.5	U	15.0	P	11/07/2025	14:52	LB137834
	Copper	2.30	<5	U	10.0	P	11/07/2025	14:52	LB137834
	Iron	11.7	<25	U	50.0	P	11/07/2025	14:52	LB137834
	Lead	1.15	<3	U	6.00	P	11/07/2025	14:52	LB137834
	Magnesium	122	<500	U	1000	P	11/07/2025	14:52	LB137834
	Manganese	2.97	<5	U	10.0	P	11/07/2025	14:52	LB137834
	Nickel	1.53	<10	U	20.0	P	11/07/2025	14:52	LB137834

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Remington & Vernick Engineers

SDG No.: Q3534

Instrument: P4

Potassium	459	<500	U	1000	P	11/07/2025	14:52	LB137834
Selenium	4.82	<5	U	10.0	P	11/07/2025	14:52	LB137834
Silver	0.81	<2.5	U	5.00	P	11/07/2025	14:52	LB137834
Sodium	434	<500	U	1000	P	11/07/2025	14:52	LB137834
Thallium	6.65	<10	J	20.0	P	11/07/2025	14:52	LB137834
Vanadium	3.13	<10	U	20.0	P	11/07/2025	14:52	LB137834
Zinc	8.33	<10	U	20.0	P	11/07/2025	14:52	LB137834