

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI02
Lab Code: ACE **SDG NO.:** Q3534
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL01 **Date Analyzed:** 11/06/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114936.D **Time Analyzed:** 14:00

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.487	5.387	5.587	505.510	500.000	1.1
Aroclor-1016-2	5.509	5.408	5.608	500.610	500.000	0.1
Aroclor-1016-3	5.570	5.470	5.670	519.930	500.000	4.0
Aroclor-1016-4	5.666	5.566	5.766	511.720	500.000	2.3
Aroclor-1016-5	5.958	5.858	6.058	486.570	500.000	-2.7
Aroclor-1260-1	7.072	6.971	7.171	485.660	500.000	-2.9
Aroclor-1260-2	7.325	7.225	7.425	513.500	500.000	2.7
Aroclor-1260-3	7.683	7.582	7.782	513.020	500.000	2.6
Aroclor-1260-4	7.904	7.803	8.003	499.360	500.000	-0.1
Aroclor-1260-5	8.210	8.110	8.310	502.550	500.000	0.5
Decachlorobiphenyl	9.930	9.830	10.030	49.940	50.000	-0.1
Tetrachloro-m-xylene	4.348	4.248	4.448	51.460	50.000	2.9

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI02
Lab Code: ACE **SDG NO.:** Q3534
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL01 **Date Analyzed:** 11/06/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114936.D **Time Analyzed:** 14:00

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.726	4.628	4.828	491.610	500.000	-1.7
Aroclor-1016-2	4.745	4.647	4.847	494.270	500.000	-1.1
Aroclor-1016-3	4.920	4.821	5.021	500.530	500.000	0.1
Aroclor-1016-4	4.962	4.864	5.064	496.850	500.000	-0.6
Aroclor-1016-5	5.174	5.076	5.276	502.840	500.000	0.6
Aroclor-1260-1	6.205	6.106	6.306	483.630	500.000	-3.3
Aroclor-1260-2	6.392	6.294	6.494	488.750	500.000	-2.3
Aroclor-1260-3	6.545	6.446	6.646	489.890	500.000	-2.0
Aroclor-1260-4	7.016	6.917	7.117	498.910	500.000	-0.2
Aroclor-1260-5	7.257	7.158	7.358	493.840	500.000	-1.2
Decachlorobiphenyl	8.639	8.541	8.741	52.450	50.000	4.9
Tetrachloro-m-xylene	3.648	3.548	3.748	49.210	50.000	-1.6