

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI02
Lab Code: ACE **SDG NO.:** Q3534
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL02 **Date Analyzed:** 11/06/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114951.D **Time Analyzed:** 19:28

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.486	5.387	5.587	512.950	500.000	2.6
Aroclor-1016-2	5.508	5.408	5.608	504.850	500.000	1.0
Aroclor-1016-3	5.569	5.470	5.670	531.690	500.000	6.3
Aroclor-1016-4	5.665	5.566	5.766	520.950	500.000	4.2
Aroclor-1016-5	5.957	5.858	6.058	497.090	500.000	-0.6
Aroclor-1260-1	7.070	6.971	7.171	491.430	500.000	-1.7
Aroclor-1260-2	7.324	7.225	7.425	525.920	500.000	5.2
Aroclor-1260-3	7.682	7.582	7.782	524.330	500.000	4.9
Aroclor-1260-4	7.903	7.803	8.003	511.630	500.000	2.3
Aroclor-1260-5	8.210	8.110	8.310	510.830	500.000	2.2
Decachlorobiphenyl	9.928	9.830	10.030	50.820	50.000	1.6
Tetrachloro-m-xylene	4.347	4.248	4.448	52.160	50.000	4.3

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI02
Lab Code: ACE **SDG NO.:** Q3534
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL02 **Date Analyzed:** 11/06/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114951.D **Time Analyzed:** 19:28

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.726	4.628	4.828	493.530	500.000	-1.3
Aroclor-1016-2	4.745	4.647	4.847	495.870	500.000	-0.8
Aroclor-1016-3	4.920	4.821	5.021	502.760	500.000	0.6
Aroclor-1016-4	4.961	4.864	5.064	498.680	500.000	-0.3
Aroclor-1016-5	5.174	5.076	5.276	501.150	500.000	0.2
Aroclor-1260-1	6.204	6.106	6.306	481.910	500.000	-3.6
Aroclor-1260-2	6.392	6.294	6.494	484.950	500.000	-3.0
Aroclor-1260-3	6.544	6.446	6.646	488.220	500.000	-2.4
Aroclor-1260-4	7.015	6.917	7.117	496.560	500.000	-0.7
Aroclor-1260-5	7.256	7.158	7.358	494.130	500.000	-1.2
Decachlorobiphenyl	8.639	8.541	8.741	52.800	50.000	5.6
Tetrachloro-m-xylene	3.647	3.548	3.748	49.810	50.000	-0.4