

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI02
Lab Code: ACE **SDG NO.:** Q3534
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL03 **Date Analyzed:** 11/07/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114957.D **Time Analyzed:** 09:49

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.486	5.387	5.587	500.540	500.000	0.1
Aroclor-1016-2	5.508	5.408	5.608	499.430	500.000	-0.1
Aroclor-1016-3	5.569	5.470	5.670	516.380	500.000	3.3
Aroclor-1016-4	5.665	5.566	5.766	504.650	500.000	0.9
Aroclor-1016-5	5.957	5.858	6.058	491.820	500.000	-1.6
Aroclor-1260-1	7.070	6.971	7.171	482.130	500.000	-3.6
Aroclor-1260-2	7.324	7.225	7.425	515.470	500.000	3.1
Aroclor-1260-3	7.682	7.582	7.782	515.080	500.000	3.0
Aroclor-1260-4	7.904	7.803	8.003	501.190	500.000	0.2
Aroclor-1260-5	8.210	8.110	8.310	498.760	500.000	-0.2
Decachlorobiphenyl	9.932	9.830	10.030	49.920	50.000	-0.2
Tetrachloro-m-xylene	4.346	4.248	4.448	52.000	50.000	4.0

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI02
Lab Code: ACE **SDG NO.:** Q3534
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL03 **Date Analyzed:** 11/07/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114957.D **Time Analyzed:** 09:49

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.725	4.628	4.828	491.660	500.000	-1.7
Aroclor-1016-2	4.745	4.647	4.847	488.160	500.000	-2.4
Aroclor-1016-3	4.919	4.821	5.021	496.200	500.000	-0.8
Aroclor-1016-4	4.961	4.864	5.064	493.920	500.000	-1.2
Aroclor-1016-5	5.174	5.076	5.276	496.130	500.000	-0.8
Aroclor-1260-1	6.204	6.106	6.306	474.390	500.000	-5.1
Aroclor-1260-2	6.391	6.294	6.494	480.080	500.000	-4.0
Aroclor-1260-3	6.544	6.446	6.646	480.920	500.000	-3.8
Aroclor-1260-4	7.015	6.917	7.117	487.810	500.000	-2.4
Aroclor-1260-5	7.255	7.158	7.358	486.450	500.000	-2.7
Decachlorobiphenyl	8.639	8.541	8.741	52.000	50.000	4.0
Tetrachloro-m-xylene	3.646	3.548	3.748	49.430	50.000	-1.1