

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI02
Lab Code: ACE **SDG NO.:** Q3534
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL04 **Date Analyzed:** 11/07/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114974.D **Time Analyzed:** 16:14

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.487	5.387	5.587	497.780	500.000	-0.4
Aroclor-1016-2	5.509	5.408	5.608	488.030	500.000	-2.4
Aroclor-1016-3	5.570	5.470	5.670	510.240	500.000	2.0
Aroclor-1016-4	5.666	5.566	5.766	498.750	500.000	-0.3
Aroclor-1016-5	5.959	5.858	6.058	477.030	500.000	-4.6
Aroclor-1260-1	7.072	6.971	7.171	473.000	500.000	-5.4
Aroclor-1260-2	7.327	7.225	7.425	506.660	500.000	1.3
Aroclor-1260-3	7.684	7.582	7.782	499.360	500.000	-0.1
Aroclor-1260-4	7.905	7.803	8.003	490.650	500.000	-1.9
Aroclor-1260-5	8.213	8.110	8.310	493.530	500.000	-1.3
Decachlorobiphenyl	9.935	9.830	10.030	49.200	50.000	-1.6
Tetrachloro-m-xylene	4.347	4.248	4.448	51.230	50.000	2.5

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI02
Lab Code: ACE **SDG NO.:** Q3534
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL04 **Date Analyzed:** 11/07/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114974.D **Time Analyzed:** 16:14

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.726	4.628	4.828	493.590	500.000	-1.3
Aroclor-1016-2	4.745	4.647	4.847	495.330	500.000	-0.9
Aroclor-1016-3	4.920	4.821	5.021	504.380	500.000	0.9
Aroclor-1016-4	4.962	4.864	5.064	495.530	500.000	-0.9
Aroclor-1016-5	5.174	5.076	5.276	498.560	500.000	-0.3
Aroclor-1260-1	6.206	6.106	6.306	478.330	500.000	-4.3
Aroclor-1260-2	6.392	6.294	6.494	477.760	500.000	-4.4
Aroclor-1260-3	6.546	6.446	6.646	481.330	500.000	-3.7
Aroclor-1260-4	7.017	6.917	7.117	489.610	500.000	-2.1
Aroclor-1260-5	7.258	7.158	7.358	489.610	500.000	-2.1
Decachlorobiphenyl	8.641	8.541	8.741	52.650	50.000	5.3
Tetrachloro-m-xylene	3.647	3.548	3.748	50.340	50.000	0.7