

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI02
Lab Code: ACE **SDG NO.:** Q3534
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL05 **Date Analyzed:** 11/06/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076352.D **Time Analyzed:** 14:02

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.793	5.695	5.895	500.170	500.000	0.0
Aroclor-1016-2	5.814	5.717	5.917	500.320	500.000	0.1
Aroclor-1016-3	5.877	5.780	5.980	505.820	500.000	1.2
Aroclor-1016-4	5.974	5.877	6.077	509.170	500.000	1.8
Aroclor-1016-5	6.266	6.169	6.369	501.660	500.000	0.3
Aroclor-1260-1	7.384	7.287	7.487	529.390	500.000	5.9
Aroclor-1260-2	7.637	7.540	7.740	518.180	500.000	3.6
Aroclor-1260-3	7.995	7.899	8.099	536.210	500.000	7.2
Aroclor-1260-4	8.223	8.127	8.327	546.660	500.000	9.3
Aroclor-1260-5	8.548	8.452	8.652	507.390	500.000	1.5
Decachlorobiphenyl	10.411	10.319	10.519	47.980	50.000	-4.0
Tetrachloro-m-xylene	4.641	4.542	4.742	49.360	50.000	-1.3

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI02
Lab Code: ACE **SDG NO.:** Q3534
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL05 **Date Analyzed:** 11/06/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076352.D **Time Analyzed:** 14:02

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.872	4.778	4.978	487.590	500.000	-2.5
Aroclor-1016-2	4.931	4.836	5.036	466.830	500.000	-6.6
Aroclor-1016-3	5.050	4.954	5.154	458.550	500.000	-8.3
Aroclor-1016-4	5.091	4.996	5.196	523.780	500.000	4.8
Aroclor-1016-5	5.306	5.210	5.410	466.790	500.000	-6.6
Aroclor-1260-1	6.332	6.238	6.438	534.840	500.000	7.0
Aroclor-1260-2	6.519	6.425	6.625	531.060	500.000	6.2
Aroclor-1260-3	6.673	6.578	6.778	529.430	500.000	5.9
Aroclor-1260-4	7.142	7.047	7.247	594.100	500.000	18.8
Aroclor-1260-5	7.382	7.287	7.487	538.080	500.000	7.6
Decachlorobiphenyl	8.774	8.682	8.882	44.670	50.000	-10.7
Tetrachloro-m-xylene	3.778	3.681	3.881	51.150	50.000	2.3