

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI02
Lab Code: ACE **SDG NO.:** Q3534
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/17/2025 10/17/2025

Client Sample No.: CCAL01 **Date Analyzed:** 11/06/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD090992.D **Time Analyzed:** 14:05

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.699	6.600	6.800	50.060	50.000	0.1
4,4'-DDE	6.188	6.090	6.290	46.420	50.000	-7.2
4,4'-DDT	7.014	6.915	7.115	50.740	50.000	1.5
Aldrin	5.262	5.164	5.364	47.580	50.000	-4.8
alpha-BHC	3.992	3.894	4.094	48.420	50.000	-3.2
alpha-Chlordane	6.019	5.921	6.121	46.670	50.000	-6.7
beta-BHC	4.511	4.412	4.612	46.700	50.000	-6.6
Decachlorobiphenyl	9.063	8.964	9.164	43.690	50.000	-12.6
delta-BHC	4.759	4.660	4.860	47.810	50.000	-4.4
Dieldrin	6.339	6.240	6.440	47.710	50.000	-4.6
Endosulfan I	6.066	5.968	6.168	47.450	50.000	-5.1
Endosulfan II	6.779	6.681	6.881	48.160	50.000	-3.7
Endosulfan sulfate	7.143	7.045	7.245	47.600	50.000	-4.8
Endrin	6.566	6.468	6.668	47.030	50.000	-5.9
Endrin aldehyde	6.907	6.810	7.010	47.990	50.000	-4.0
Endrin ketone	7.624	7.525	7.725	49.920	50.000	-0.2
gamma-BHC (Lindane)	4.323	4.225	4.425	47.790	50.000	-4.4
gamma-Chlordane	5.938	5.840	6.040	46.860	50.000	-6.3
Heptachlor	4.921	4.823	5.023	58.110	50.000	16.2
Heptachlor epoxide	5.683	5.584	5.784	48.310	50.000	-3.4
Methoxychlor	7.487	7.388	7.588	51.410	50.000	2.8
Tetrachloro-m-xylene	3.543	3.445	3.645	46.870	50.000	-6.3

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI02
Lab Code: ACE **SDG NO.:** Q3534
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/17/2025 10/17/2025

Client Sample No.: CCAL01 **Date Analyzed:** 11/06/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD090992.D **Time Analyzed:** 14:05

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.918	5.821	6.021	46.250	50.000	-7.5
4,4'-DDE	5.363	5.266	5.466	48.570	50.000	-2.9
4,4'-DDT	6.171	6.074	6.274	50.560	50.000	1.1
Aldrin	4.358	4.261	4.461	48.070	50.000	-3.9
alpha-BHC	3.385	3.287	3.487	48.530	50.000	-2.9
alpha-Chlordane	5.179	5.081	5.281	47.750	50.000	-4.5
beta-BHC	4.017	3.919	4.119	47.620	50.000	-4.8
Decachlorobiphenyl	8.059	7.962	8.162	50.400	50.000	0.8
delta-BHC	4.253	4.155	4.355	48.290	50.000	-3.4
Dieldrin	5.501	5.404	5.604	48.320	50.000	-3.4
Endosulfan I	5.235	5.138	5.338	45.880	50.000	-8.2
Endosulfan II	6.069	5.972	6.172	46.380	50.000	-7.2
Endosulfan sulfate	6.471	6.374	6.574	49.280	50.000	-1.4
Endrin	5.778	5.681	5.881	47.430	50.000	-5.1
Endrin aldehyde	6.247	6.150	6.350	49.950	50.000	-0.1
Endrin ketone	6.979	6.883	7.083	53.660	50.000	7.3
gamma-BHC (Lindane)	3.721	3.623	3.823	48.300	50.000	-3.4
gamma-Chlordane	5.114	5.017	5.217	47.820	50.000	-4.4
Heptachlor	4.073	3.975	4.175	50.580	50.000	1.2
Heptachlor epoxide	4.861	4.764	4.964	47.020	50.000	-6.0
Methoxychlor	6.742	6.645	6.845	51.500	50.000	3.0
Tetrachloro-m-xylene	2.874	2.775	2.975	48.060	50.000	-3.9