

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI02
Lab Code: ACE **SDG NO.:** Q3534
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/17/2025 10/17/2025

Client Sample No.: CCAL02 **Date Analyzed:** 11/06/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD091013.D **Time Analyzed:** 19:10

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.698	6.600	6.800	51.110	50.000	2.2
4,4'-DDE	6.188	6.090	6.290	46.760	50.000	-6.5
4,4'-DDT	7.013	6.915	7.115	49.590	50.000	-0.8
Aldrin	5.262	5.164	5.364	47.720	50.000	-4.6
alpha-BHC	3.992	3.894	4.094	48.960	50.000	-2.1
alpha-Chlordane	6.018	5.921	6.121	46.050	50.000	-7.9
beta-BHC	4.510	4.412	4.612	47.120	50.000	-5.8
Decachlorobiphenyl	9.061	8.964	9.164	44.140	50.000	-11.7
delta-BHC	4.759	4.660	4.860	48.580	50.000	-2.8
Dieldrin	6.338	6.240	6.440	48.280	50.000	-3.4
Endosulfan I	6.066	5.968	6.168	47.530	50.000	-4.9
Endosulfan II	6.779	6.681	6.881	48.290	50.000	-3.4
Endosulfan sulfate	7.142	7.045	7.245	47.730	50.000	-4.5
Endrin	6.566	6.468	6.668	47.810	50.000	-4.4
Endrin aldehyde	6.908	6.810	7.010	48.240	50.000	-3.5
Endrin ketone	7.623	7.525	7.725	50.520	50.000	1.0
gamma-BHC (Lindane)	4.323	4.225	4.425	48.200	50.000	-3.6
gamma-Chlordane	5.937	5.840	6.040	47.100	50.000	-5.8
Heptachlor	4.920	4.823	5.023	57.740	50.000	15.5
Heptachlor epoxide	5.682	5.584	5.784	48.530	50.000	-2.9
Methoxychlor	7.486	7.388	7.588	49.400	50.000	-1.2
Tetrachloro-m-xylene	3.543	3.445	3.645	46.980	50.000	-6.0

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI02
Lab Code: ACE **SDG NO.:** Q3534
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/17/2025 10/17/2025

Client Sample No.: CCAL02 **Date Analyzed:** 11/06/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD091013.D **Time Analyzed:** 19:10

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.917	5.821	6.021	49.590	50.000	-0.8
4,4'-DDE	5.363	5.266	5.466	48.230	50.000	-3.5
4,4'-DDT	6.171	6.074	6.274	48.920	50.000	-2.2
Aldrin	4.358	4.261	4.461	48.180	50.000	-3.6
alpha-BHC	3.385	3.287	3.487	49.080	50.000	-1.8
alpha-Chlordane	5.179	5.081	5.281	47.690	50.000	-4.6
beta-BHC	4.017	3.919	4.119	47.850	50.000	-4.3
Decachlorobiphenyl	8.058	7.962	8.162	50.040	50.000	0.1
delta-BHC	4.253	4.155	4.355	48.770	50.000	-2.5
Dieldrin	5.501	5.404	5.604	48.370	50.000	-3.3
Endosulfan I	5.236	5.138	5.338	45.880	50.000	-8.2
Endosulfan II	6.069	5.972	6.172	48.660	50.000	-2.7
Endosulfan sulfate	6.471	6.374	6.574	48.950	50.000	-2.1
Endrin	5.778	5.681	5.881	48.460	50.000	-3.1
Endrin aldehyde	6.247	6.150	6.350	48.910	50.000	-2.2
Endrin ketone	6.978	6.883	7.083	52.410	50.000	4.8
gamma-BHC (Lindane)	3.721	3.623	3.823	48.750	50.000	-2.5
gamma-Chlordane	5.114	5.017	5.217	47.800	50.000	-4.4
Heptachlor	4.072	3.975	4.175	50.280	50.000	0.6
Heptachlor epoxide	4.861	4.764	4.964	46.940	50.000	-6.1
Methoxychlor	6.742	6.645	6.845	49.470	50.000	-1.1
Tetrachloro-m-xylene	2.873	2.775	2.975	48.520	50.000	-3.0