

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI02
Lab Code: ACE **SDG NO.:** Q3534
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/17/2025 10/17/2025

Client Sample No.: CCAL03 **Date Analyzed:** 11/07/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD091017.D **Time Analyzed:** 10:10

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.698	6.600	6.800	53.320	50.000	6.6
4,4'-DDE	6.189	6.090	6.290	48.830	50.000	-2.3
4,4'-DDT	7.014	6.915	7.115	52.380	50.000	4.8
Aldrin	5.263	5.164	5.364	50.120	50.000	0.2
alpha-BHC	3.993	3.894	4.094	51.440	50.000	2.9
alpha-Chlordane	6.019	5.921	6.121	48.970	50.000	-2.1
beta-BHC	4.511	4.412	4.612	49.180	50.000	-1.6
Decachlorobiphenyl	9.063	8.964	9.164	46.610	50.000	-6.8
delta-BHC	4.759	4.660	4.860	50.630	50.000	1.3
Dieldrin	6.339	6.240	6.440	50.510	50.000	1.0
Endosulfan I	6.066	5.968	6.168	49.850	50.000	-0.3
Endosulfan II	6.779	6.681	6.881	50.530	50.000	1.1
Endosulfan sulfate	7.143	7.045	7.245	49.710	50.000	-0.6
Endrin	6.567	6.468	6.668	49.360	50.000	-1.3
Endrin aldehyde	6.908	6.810	7.010	50.950	50.000	1.9
Endrin ketone	7.624	7.525	7.725	51.990	50.000	4.0
gamma-BHC (Lindane)	4.324	4.225	4.425	50.510	50.000	1.0
gamma-Chlordane	5.938	5.840	6.040	49.310	50.000	-1.4
Heptachlor	4.920	4.823	5.023	59.830	50.000	19.7
Heptachlor epoxide	5.683	5.584	5.784	51.140	50.000	2.3
Methoxychlor	7.487	7.388	7.588	52.780	50.000	5.6
Tetrachloro-m-xylene	3.544	3.445	3.645	49.160	50.000	-1.7

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI02
Lab Code: ACE **SDG NO.:** Q3534
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/17/2025 10/17/2025

Client Sample No.: CCAL03 **Date Analyzed:** 11/07/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD091017.D **Time Analyzed:** 10:10

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.917	5.821	6.021	53.730	50.000	7.5
4,4'-DDE	5.363	5.266	5.466	52.370	50.000	4.7
4,4'-DDT	6.171	6.074	6.274	53.960	50.000	7.9
Aldrin	4.358	4.261	4.461	52.300	50.000	4.6
alpha-BHC	3.385	3.287	3.487	52.850	50.000	5.7
alpha-Chlordane	5.179	5.081	5.281	51.750	50.000	3.5
beta-BHC	4.017	3.919	4.119	51.540	50.000	3.1
Decachlorobiphenyl	8.058	7.962	8.162	54.040	50.000	8.1
delta-BHC	4.253	4.155	4.355	52.270	50.000	4.5
Dieldrin	5.501	5.404	5.604	52.230	50.000	4.5
Endosulfan I	5.236	5.138	5.338	49.320	50.000	-1.4
Endosulfan II	6.069	5.972	6.172	52.760	50.000	5.5
Endosulfan sulfate	6.471	6.374	6.574	52.730	50.000	5.5
Endrin	5.778	5.681	5.881	51.270	50.000	2.5
Endrin aldehyde	6.247	6.150	6.350	53.490	50.000	7.0
Endrin ketone	6.979	6.883	7.083	57.790	50.000	15.6
gamma-BHC (Lindane)	3.721	3.623	3.823	52.450	50.000	4.9
gamma-Chlordane	5.114	5.017	5.217	51.890	50.000	3.8
Heptachlor	4.073	3.975	4.175	54.080	50.000	8.2
Heptachlor epoxide	4.862	4.764	4.964	50.810	50.000	1.6
Methoxychlor	6.742	6.645	6.845	54.430	50.000	8.9
Tetrachloro-m-xylene	2.874	2.775	2.975	52.180	50.000	4.4