

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI02
Lab Code: ACE **SDG NO.:** Q3534
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/17/2025 10/17/2025

Client Sample No.: CCAL04 **Date Analyzed:** 11/07/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD091026.D **Time Analyzed:** 13:32

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.700	6.600	6.800	54.820	50.000	9.6
4,4'-DDE	6.190	6.090	6.290	46.700	50.000	-6.6
4,4'-DDT	7.015	6.915	7.115	51.990	50.000	4.0
Aldrin	5.263	5.164	5.364	48.620	50.000	-2.8
alpha-BHC	3.994	3.894	4.094	49.560	50.000	-0.9
alpha-Chlordane	6.020	5.921	6.121	46.400	50.000	-7.2
beta-BHC	4.512	4.412	4.612	48.130	50.000	-3.7
Decachlorobiphenyl	9.065	8.964	9.164	45.770	50.000	-8.5
delta-BHC	4.760	4.660	4.860	49.210	50.000	-1.6
Dieldrin	6.340	6.240	6.440	49.040	50.000	-1.9
Endosulfan I	6.068	5.968	6.168	48.230	50.000	-3.5
Endosulfan II	6.779	6.681	6.881	47.870	50.000	-4.3
Endosulfan sulfate	7.144	7.045	7.245	48.110	50.000	-3.8
Endrin	6.568	6.468	6.668	49.730	50.000	-0.5
Endrin aldehyde	6.908	6.810	7.010	48.360	50.000	-3.3
Endrin ketone	7.626	7.525	7.725	50.160	50.000	0.3
gamma-BHC (Lindane)	4.324	4.225	4.425	48.920	50.000	-2.2
gamma-Chlordane	5.939	5.840	6.040	47.870	50.000	-4.3
Heptachlor	4.922	4.823	5.023	58.580	50.000	17.2
Heptachlor epoxide	5.684	5.584	5.784	51.710	50.000	3.4
Methoxychlor	7.489	7.388	7.588	52.740	50.000	5.5
Tetrachloro-m-xylene	3.544	3.445	3.645	47.770	50.000	-4.5

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI02
Lab Code: ACE **SDG NO.:** Q3534
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/17/2025 10/17/2025

Client Sample No.: CCAL04 **Date Analyzed:** 11/07/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD091026.D **Time Analyzed:** 13:32

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.918	5.821	6.021	52.760	50.000	5.5
4,4'-DDE	5.362	5.266	5.466	50.500	50.000	1.0
4,4'-DDT	6.172	6.074	6.274	53.640	50.000	7.3
Aldrin	4.359	4.261	4.461	51.740	50.000	3.5
alpha-BHC	3.385	3.287	3.487	51.700	50.000	3.4
alpha-Chlordane	5.180	5.081	5.281	50.320	50.000	0.6
beta-BHC	4.018	3.919	4.119	49.800	50.000	-0.4
Decachlorobiphenyl	8.059	7.962	8.162	54.870	50.000	9.7
delta-BHC	4.254	4.155	4.355	51.510	50.000	3.0
Dieldrin	5.502	5.404	5.604	51.530	50.000	3.1
Endosulfan I	5.236	5.138	5.338	41.290	50.000	-17.4
Endosulfan II	6.070	5.972	6.172	51.980	50.000	4.0
Endosulfan sulfate	6.471	6.374	6.574	52.020	50.000	4.0
Endrin	5.779	5.681	5.881	53.110	50.000	6.2
Endrin aldehyde	6.248	6.150	6.350	52.670	50.000	5.3
Endrin ketone	6.979	6.883	7.083	58.870	50.000	17.7
gamma-BHC (Lindane)	3.721	3.623	3.823	51.380	50.000	2.8
gamma-Chlordane	5.115	5.017	5.217	50.670	50.000	1.3
Heptachlor	4.073	3.975	4.175	53.310	50.000	6.6
Heptachlor epoxide	4.862	4.764	4.964	49.440	50.000	-1.1
Methoxychlor	6.742	6.645	6.845	54.070	50.000	8.1
Tetrachloro-m-xylene	2.874	2.775	2.975	51.110	50.000	2.2