

PESTICIDE CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: REMI02

Lab Code: ACE

SDG NO.: Q3534

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/17/2025 10/17/2025

Client Sample No. (PEM): PEM - PD090981.D Date Analyzed: 11/06/2025

Lab Sample No.(PEM): PEM Time Analyzed: 09:34

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.062	8.960	9.160	21.660	20.000	8.3
Tetrachloro-m-xylene	3.544	3.490	3.590	22.290	20.000	11.5
alpha-BHC	3.993	3.940	4.040	9.490	10.000	-5.1
beta-BHC	4.511	4.460	4.560	10.760	10.000	7.6
gamma-BHC (Lindane)	4.324	4.270	4.370	9.720	10.000	-2.8
Endrin	6.567	6.500	6.640	51.620	50.000	3.2
4,4'-DDT	7.015	6.940	7.090	110.870	100.000	10.9
Methoxychlor	7.488	7.420	7.560	262.650	250.000	5.1

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/17/2025 10/17/2025

Client Sample No. (PEM): PEM - PD090981.D Date Analyzed: 11/06/2025

Lab Sample No.(PEM): PEM Time Analyzed: 09:34

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	8.059	7.960	8.160	23.650	20.000	18.3
Tetrachloro-m-xylene	2.874	2.820	2.920	23.040	20.000	15.2
alpha-BHC	3.385	3.330	3.440	11.220	10.000	12.2
beta-BHC	4.017	3.970	4.070	11.440	10.000	14.4
gamma-BHC (Lindane)	3.721	3.670	3.770	11.300	10.000	13.0
Endrin	5.779	5.710	5.850	51.010	50.000	2.0
4,4'-DDT	6.172	6.100	6.240	108.570	100.000	8.6
Methoxychlor	6.742	6.670	6.810	224.730	250.000	-10.1