

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110725\
 Data File : BN038147.D
 Acq On : 07 Nov 2025 09:39
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Nov 07 10:03:14 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 29 13:09:46 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	91	0.00
2	1,4-Dioxane	0.413	0.407	1.5	87	0.00
3	n-Nitrosodimethylamine	0.532	0.514	3.4	87	0.00
4 S	2-Fluorophenol	0.942	0.847	10.1	80	0.00
5 S	Phenol-d6	1.148	1.036	9.8	83	0.00
6	bis(2-Chloroethyl)ether	1.128	1.082	4.1	87	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	89	0.00
8 S	Nitrobenzene-d5	0.323	0.310	4.0	88	0.00
9	Naphthalene	1.102	1.094	0.7	90	0.00
10	Hexachlorobutadiene	0.186	0.197	-5.9	94	0.00
11 SURR	2-Methylnaphthalene-d10	0.571	0.545	4.6	88	0.00
12	2-Methylnaphthalene	0.735	0.704	4.2	89	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	87	0.00
14 S	2,4,6-Tribromophenol	0.165	0.133	19.4	78	0.00
15 S	2-Fluorobiphenyl	1.602	1.557	2.8	86	0.00
16	Acenaphthylene	1.811	1.729	4.5	88	0.00
17	Acenaphthene	1.267	1.227	3.2	88	0.00
18	Fluorene	1.661	1.637	1.4	89	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	88	0.00
20	4,6-Dinitro-2-methylphenol	0.055	0.037	32.7#	72	0.00
21	4-Bromophenyl-phenylether	0.262	0.258	1.5	90	0.00
22	Hexachlorobenzene	0.298	0.318	-6.7	93	0.00
23	Atrazine	0.192	0.163	15.1	81	0.00
24	Pentachlorophenol	0.128	0.095	25.8#	77	0.00
25	Phenanthrene	1.270	1.262	0.6	90	0.00
26	Anthracene	1.111	1.046	5.9	87	0.00
27 SURR	Fluoranthene-d10	1.009	0.947	6.1	87	0.00
28	Fluoranthene	1.419	1.359	4.2	88	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	89	0.00
30	Pyrene	1.806	1.806	0.0	87	0.00
31 S	Terphenyl-d14	0.869	0.848	2.4	87	0.00
32	Benzo(a)anthracene	1.427	1.322	7.4	87	0.00
33	Chrysene	1.549	1.583	-2.2	91	0.00
34	Bis(2-ethylhexyl)phthalate	0.752	0.669	11.0	83	0.00
35 I	Perylene-d12	1.000	1.000	0.0	90	0.00
36	Indeno(1,2,3-cd)pyrene	1.643	1.607	2.2	90	0.00
37	Benzo(b)fluoranthene	1.710	1.686	1.4	93	0.00
38	Benzo(k)fluoranthene	1.721	1.737	-0.9	97	0.00
39 C	Benzo(a)pyrene	1.369	1.294	5.5	89	0.00
40	Dibenzo(a,h)anthracene	1.265	1.215	4.0	92	0.00
41	Benzo(g,h,i)perylene	1.444	1.432	0.8	91	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4