

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110725\
 Data File : BN038149.D
 Acq On : 07 Nov 2025 10:52
 Operator : RC/JU
 Sample : Q3534-01
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 MW-6

Quant Time: Nov 07 11:15:31 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 29 13:09:46 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.775	152	10192	0.400	ng	0.00
7) Naphthalene-d8	10.573	136	31652	0.400	ng	# 0.00
13) Acenaphthene-d10	14.430	164	28745	0.400	ng	0.00
19) Phenanthrene-d10	17.186	188	36418	0.400	ng	# 0.00
29) Chrysene-d12	21.375	240	24102	0.400	ng	# 0.00
35) Perylene-d12	23.680	264	20067	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.348	112	5382	0.224	ng	0.00
5) Phenol-d6	6.937	99	3656	0.125	ng	0.00
8) Nitrobenzene-d5	8.929	82	8623	0.338	ng	0.00
11) 2-Methylnaphthalene-d10	12.172	152	17265	0.382	ng	0.00
14) 2,4,6-Tribromophenol	15.932	330	4308	0.364	ng	0.00
15) 2-Fluorobiphenyl	13.051	172	27151	0.236	ng	0.00
27) Fluoranthene-d10	19.220	212	33915	0.369	ng	0.00
31) Terphenyl-d14	19.819	244	27365	0.523	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.247	88	148894	14.139	ng	98
3) n-Nitrosodimethylamine	3.535	42	1436	0.106	ng	# 1
6) bis(2-Chloroethyl)ether	7.240	93	2020	0.070	ng	# 6
9) Naphthalene	10.626	128	17492	0.201	ng	# 92
12) 2-Methylnaphthalene	12.248	142	91422	1.573	ng	100
16) Acenaphthylene	14.152	152	3724	0.029	ng	# 83
17) Acenaphthene	14.494	154	58144	0.638	ng	98
18) Fluorene	15.489	166	67952	0.569	ng	# 94
20) 4,6-Dinitro-2-methylph...	15.489	198	68	0.193	ng	# 1
23) Atrazine	16.652	200	387	0.022	ng	# 1
24) Pentachlorophenol	16.838	266	2835	0.242	ng	99
25) Phenanthrene	17.223	178	71457	0.618	ng	99
26) Anthracene	17.223	178	71457	0.707	ng	97
28) Fluoranthene	19.248	202	15796	0.122	ng	# 86
30) Pyrene	19.610	202	12169	0.112	ng	# 80
32) Benzo(a)anthracene	21.357	228	6339	0.074	ng	# 80
33) Chrysene	21.411	228	4095	0.044	ng	# 67
34) Bis(2-ethylhexyl)phtha...	21.295	149	16812	0.371	ng	# 90
36) Indeno(1,2,3-cd)pyrene	26.034	276	4958	0.060	ng	99
37) Benzo(b)fluoranthene	22.984	252	5785	0.067	ng	# 39
38) Benzo(k)fluoranthene	23.028	252	6151	0.071	ng	# 45
39) Benzo(a)pyrene	23.578	252	2762	0.040	ng	# 5
40) Dibenzo(a,h)anthracene	26.051	278	4107	0.065	ng	# 67
41) Benzo(g,h,i)perylene	26.750	276	4286	0.059	ng	# 83


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110725\
Data File : BN038149.D
Acq On : 07 Nov 2025 10:52
Operator : RC/JU
Sample : Q3534-01
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
MW-6

Quant Time: Nov 07 11:15:31 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 29 13:09:46 2025
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.775 min Scan# 61
 Delta R.T. -0.000 min
 Lab File: BN038149.D
 Acq: 07 Nov 2025 10:52




Instrument :
 BNA_N
 ClientSampleId :
 MW-6


Tgt Ion:152 Resp: 10192
 Ion Ratio Lower Upper
 152 100
 150 153.4 122.3 183.5
 115 58.3 47.4 71.0



#2
 1,4-Dioxane
 Concen: 14.139 ng
 RT: 3.247 min Scan# 23
 Delta R.T. -0.007 min
 Lab File: BN038149.D
 Acq: 07 Nov 2025 10:52

Tgt Ion: 88 Resp: 148894
 Ion Ratio Lower Upper
 88 100
 43 33.4 24.3 36.5
 58 75.6 60.4 90.6









#3
n-Nitrosodimethylamine
Concen: 0.106 ng
RT: 3.535 min Scan# 61
Delta R.T. -0.022 min
Lab File: BN038149.D
Acq: 07 Nov 2025 10:52




Instrument :
BNA_N
ClientSampleId :
MW-6


Tgt Ion: 42 Resp: 1436
Ion Ratio Lower Upper
42 100
74 0.0 96.8 145.2#
44 0.0 5.3 7.9#



#4
2-Fluorophenol
Concen: 0.224 ng
RT: 5.348 min Scan# 314
Delta R.T. -0.000 min
Lab File: BN038149.D
Acq: 07 Nov 2025 10:52

Tgt Ion: 112 Resp: 5382
Ion Ratio Lower Upper
112 100
64 42.9 45.4 68.0#
63 21.6 23.2 34.8#









#5
Phenol-d6
Concen: 0.125 ng
RT: 6.937 min Scan# 511
Delta R.T. -0.000 min
Lab File: BN038149.D
Acq: 07 Nov 2025 10:52




Instrument :
BNA_N
ClientSampleId :
MW-6


Tgt Ion: 99 Resp: 3656
Ion Ratio Lower Upper
99 100
42 54.3 16.6 24.8#
71 0.0 25.4 38.2#



#6
bis(2-Chloroethyl)ether
Concen: 0.070 ng
RT: 7.240 min Scan# 576
Delta R.T. 0.043 min
Lab File: BN038149.D
Acq: 07 Nov 2025 10:52

Tgt Ion: 93 Resp: 2020
Ion Ratio Lower Upper
93 100
63 41.1 59.3 88.9#
95 155.3 25.2 37.8#









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.573 min Scan# 91
Delta R.T. -0.000 min
Lab File: BN038149.D
Acq: 07 Nov 2025 10:52




Instrument :
BNA_N
ClientSampleId :
MW-6


Tgt Ion:	136	Resp:	31652
Ion	Ratio	Lower	Upper
136	100		
137	11.8	9.0	13.4
54	12.2	5.2	7.8#
68	11.7	4.1	6.1#



#8
Nitrobenzene-d5
Concen: 0.338 ng
RT: 8.929 min Scan# 782
Delta R.T. -0.000 min
Lab File: BN038149.D
Acq: 07 Nov 2025 10:52

Tgt Ion:	82	Resp:	8623
Ion	Ratio	Lower	Upper
82	100		
128	44.6	34.1	51.1
54	53.3	43.5	65.3









#9
Naphthalene
Concen: 0.201 ng
RT: 10.626 min Scan# 941
Delta R.T. -0.000 min
Lab File: BN038149.D
Acq: 07 Nov 2025 10:52




Instrument :
BNA_N
ClientSampleId :
MW-6


Tgt Ion:128 Resp: 17492
Ion Ratio Lower Upper
128 100
129 14.9 9.1 13.7#
127 16.1 10.5 15.7#



#11
2-Methylnaphthalene-d10
Concen: 0.382 ng
RT: 12.172 min Scan# 1154
Delta R.T. -0.005 min
Lab File: BN038149.D
Acq: 07 Nov 2025 10:52

Tgt Ion:152 Resp: 17265
Ion Ratio Lower Upper
152 100
151 24.7 16.8 25.2









#12
2-Methylnaphthalene
Concen: 1.573 ng
RT: 12.248 min Scan# 1170
Delta R.T. -0.000 min
Lab File: BN038149.D
Acq: 07 Nov 2025 10:52




Instrument :
BNA_N
ClientSampleId :
MW-6


Tgt Ion:142 Resp: 91422
Ion Ratio Lower Upper
142 100
141 88.1 70.3 105.5
115 35.3 27.8 41.6



#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.430 min Scan# 1452
Delta R.T. -0.000 min
Lab File: BN038149.D
Acq: 07 Nov 2025 10:52

Tgt Ion:164 Resp: 28745
Ion Ratio Lower Upper
164 100
162 95.3 83.0 124.4
160 48.3 40.7 61.1









#14
2,4,6-Tribromophenol
Concen: 0.364 ng
RT: 15.932 min Scan# 1587
Delta R.T. 0.000 min
Lab File: BN038149.D
Acq: 07 Nov 2025 10:52




Instrument :
BNA_N
ClientSampleId :
MW-6


Tgt Ion:330 Resp: 4308
Ion Ratio Lower Upper
330 100
332 95.5 76.6 114.8
141 134.6 34.1 51.1#



#15
2-Fluorobiphenyl
Concen: 0.236 ng
RT: 13.051 min Scan# 1323
Delta R.T. -0.000 min
Lab File: BN038149.D
Acq: 07 Nov 2025 10:52

Tgt Ion:172 Resp: 27151
Ion Ratio Lower Upper
172 100
171 35.6 27.8 41.6
170 24.3 18.1 27.1









#16
Acenaphthylene
Concen: 0.029 ng
RT: 14.152 min Scan# 1426
Delta R.T. -0.000 min
Lab File: BN038149.D
Acq: 07 Nov 2025 10:52




Instrument :
BNA_N
ClientSampleId :
MW-6


Tgt Ion:152 Resp: 3724
Ion Ratio Lower Upper
152 100
151 18.4 15.8 23.8
153 28.5 10.2 15.2#



#17
Acenaphthene
Concen: 0.638 ng
RT: 14.494 min Scan# 1458
Delta R.T. -0.000 min
Lab File: BN038149.D
Acq: 07 Nov 2025 10:52

Tgt Ion:154 Resp: 58144
Ion Ratio Lower Upper
154 100
153 112.8 90.0 135.0
152 55.0 47.3 70.9









#18
Fluorene
Concen: 0.569 ng
RT: 15.489 min Scan# 1551
Delta R.T. -0.000 min
Lab File: BN038149.D
Acq: 07 Nov 2025 10:52




Instrument :
BNA_N
ClientSampleId :
MW-6


Tgt Ion:166 Resp: 67952
Ion Ratio Lower Upper
166 100
165 96.4 79.0 118.4
167 25.8 10.7 16.1#



#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.186 min Scan# 1688
Delta R.T. -0.000 min
Lab File: BN038149.D
Acq: 07 Nov 2025 10:52

Tgt Ion:188 Resp: 36418
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 17.1 8.6 13.0#









#20
4,6-Dinitro-2-methylphenol
Concen: 0.193 ng
RT: 15.489 min Scan# 1551
Delta R.T. -0.071 min
Lab File: BN038149.D
Acq: 07 Nov 2025 10:52




Instrument :
BNA_N
ClientSampleId :
MW-6


Tgt Ion:198 Resp: 68
Ion Ratio Lower Upper
198 100
51 2351.5 88.9 133.3#
105 2836.4 49.2 73.8#



#23
Atrazine
Concen: 0.022 ng
RT: 16.652 min Scan# 1645
Delta R.T. -0.000 min
Lab File: BN038149.D
Acq: 07 Nov 2025 10:52

Tgt Ion:200 Resp: 387
Ion Ratio Lower Upper
200 100
173 309.2 21.4 32.2#
215 37.9 36.7 55.1









#24
Pentachlorophenol
Concen: 0.242 ng
RT: 16.838 min Scan# 1660
Delta R.T. -0.000 min
Lab File: BN038149.D
Acq: 07 Nov 2025 10:52




Instrument :
BNA_N
ClientSampleId :
MW-6


Tgt Ion:	266	Resp:	2835
Ion	Ratio	Lower	Upper
266	100		
264	60.7	49.3	73.9
268	62.8	50.6	75.8



#25
Phenanthrene
Concen: 0.618 ng
RT: 17.223 min Scan# 1691
Delta R.T. -0.000 min
Lab File: BN038149.D
Acq: 07 Nov 2025 10:52

Tgt Ion:	178	Resp:	71457
Ion	Ratio	Lower	Upper
178	100		
176	19.7	15.5	23.3
179	16.1	12.2	18.4









#26
Anthracene
Concen: 0.707 ng
RT: 17.223 min Scan# 1691
Delta R.T. -0.099 min
Lab File: BN038149.D
Acq: 07 Nov 2025 10:52




Instrument :
BNA_N
ClientSampleId :
MW-6


Tgt Ion:178 Resp: 71457
Ion Ratio Lower Upper
178 100
176 19.7 14.9 22.3
179 16.1 11.8 17.8



#27
Fluoranthene-d10
Concen: 0.369 ng
RT: 19.220 min Scan# 1947
Delta R.T. -0.000 min
Lab File: BN038149.D
Acq: 07 Nov 2025 10:52

Tgt Ion:212 Resp: 33915
Ion Ratio Lower Upper
212 100
106 17.3 12.6 19.0
104 6.7 7.1 10.7#









#28
Fluoranthene
Concen: 0.122 ng
RT: 19.248 min Scan# 1953
Delta R.T. -0.005 min
Lab File: BN038149.D
Acq: 07 Nov 2025 10:52




Instrument :
BNA_N
ClientSampleId :
MW-6


Tgt Ion:202 Resp: 15796
Ion Ratio Lower Upper
202 100
101 23.6 9.4 14.2#
203 15.4 13.4 20.2



#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.375 min Scan# 2269
Delta R.T. -0.000 min
Lab File: BN038149.D
Acq: 07 Nov 2025 10:52

Tgt Ion:240 Resp: 24102
Ion Ratio Lower Upper
240 100
120 19.3 10.7 16.1#
236 30.5 23.1 34.7









#30
 Pyrene
 Concen: 0.112 ng
 RT: 19.610 min Scan# 2032
 Delta R.T. -0.000 min
 Lab File: BN038149.D
 Acq: 07 Nov 2025 10:52




Instrument :
 BNA_N
 ClientSampleId :
 MW-6


Tgt Ion:202 Resp: 12169
 Ion Ratio Lower Upper
 202 100
 200 26.0 16.7 25.1#
 203 30.8 14.2 21.2#



#31
 Terphenyl-d14
 Concen: 0.523 ng
 RT: 19.819 min Scan# 2076
 Delta R.T. -0.000 min
 Lab File: BN038149.D
 Acq: 07 Nov 2025 10:52

Tgt Ion:244 Resp: 27365
 Ion Ratio Lower Upper
 244 100
 212 11.2 7.8 11.6
 122 26.1 15.2 22.8#









#32
Benzo(a)anthracene
Concen: 0.074 ng
RT: 21.357 min Scan# 2135
Delta R.T. -0.000 min
Lab File: BN038149.D
Acq: 07 Nov 2025 10:52




Instrument :
BNA_N
ClientSampleId :
MW-6


Tgt Ion:228 Resp: 6339
Ion Ratio Lower Upper
228 100
226 34.7 22.2 33.4#
229 33.1 15.8 23.8#



#33
Chrysene
Concen: 0.044 ng
RT: 21.411 min Scan# 2273
Delta R.T. -0.000 min
Lab File: BN038149.D
Acq: 07 Nov 2025 10:52

Tgt Ion:228 Resp: 4095
Ion Ratio Lower Upper
228 100
226 43.0 24.3 36.5#
229 41.8 15.7 23.5#









#34
Bis(2-ethylhexyl)phthalate
Concen: 0.371 ng
RT: 21.295 min Scan# 21
Delta R.T. -0.000 min
Lab File: BN038149.D
Acq: 07 Nov 2025 10:52




Instrument :
BNA_N
ClientSampleId :
MW-6


Tgt Ion:149 Resp: 16812
Ion Ratio Lower Upper
149 100
167 20.9 21.1 31.7#
279 2.8 2.9 4.3#



#35
Perylene-d12
Concen: 0.400 ng
RT: 23.680 min Scan# 2820
Delta R.T. -0.000 min
Lab File: BN038149.D
Acq: 07 Nov 2025 10:52

Tgt Ion:264 Resp: 20067
Ion Ratio Lower Upper
264 100
260 27.8 21.8 32.8
265 74.7 68.6 102.8









#36
Indeno(1,2,3-cd)pyrene
Concen: 0.060 ng
RT: 26.034 min Scan# 30
Delta R.T. -0.000 min
Lab File: BN038149.D
Acq: 07 Nov 2025 10:52




Instrument :
BNA_N
ClientSampleId :
MW-6


Tgt Ion:276 Resp: 4958
Ion Ratio Lower Upper
276 100
138 27.2 21.4 32.0
277 24.0 19.2 28.8



#37
Benzo(b)fluoranthene
Concen: 0.067 ng
RT: 22.984 min Scan# 2582
Delta R.T. 0.003 min
Lab File: BN038149.D
Acq: 07 Nov 2025 10:52


Tgt Ion:252 Resp: 5785
Ion Ratio Lower Upper
252 100
253 50.2 20.9 31.3#
125 53.2 14.1 21.1#







#38
Benzo(k)fluoranthene
Concen: 0.071 ng
RT: 23.028 min Scan# 21
Delta R.T. -0.000 min
Lab File: BN038149.D
Acq: 07 Nov 2025 10:52

Instrument :
BNA_N
ClientSampleId :
MW-6







Tgt Ion:252 Resp: 6151
Ion Ratio Lower Upper
252 100
253 48.4 21.2 31.8#
125 50.8 14.6 22.0#



#39
Benzo(a)pyrene
Concen: 0.040 ng
RT: 23.578 min Scan# 2785
Delta R.T. -0.000 min
Lab File: BN038149.D
Acq: 07 Nov 2025 10:52

Tgt Ion:252 Resp: 2762
Ion Ratio Lower Upper
252 100
253 71.1 23.1 34.7#
125 78.9 18.6 28.0#









#40
Dibenzo(a,h)anthracene
Concen: 0.065 ng
RT: 26.051 min Scan# 3070
Delta R.T. 0.003 min
Lab File: BN038149.D
Acq: 07 Nov 2025 10:52

Instrument :
BNA_N
ClientSampleId :
MW-6

Tgt Ion:278 Resp: 4107
Ion Ratio Lower Upper
278 100
139 40.9 17.6 26.4#
279 42.4 22.6 33.8#



#41
Benzo(g,h,i)perylene
Concen: 0.059 ng
RT: 26.750 min Scan# 3870
Delta R.T. -0.000 min
Lab File: BN038149.D
Acq: 07 Nov 2025 10:52

Tgt Ion:276 Resp: 4286
Ion Ratio Lower Upper
276 100
277 30.6 19.7 29.5#
138 35.8 20.2 30.4#

