

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3534

Client ID : SSTDCCC0.4

Date Analyzed: 11/07/2025

Lab File ID: BN038165.D

Time Analyzed: 22:18

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	9349	7.775	29936	10.57	17702	14.43
UPPER LIMIT	18698	8.275	59872	11.073	35404	14.93
LOWER LIMIT	4674.5	7.275	14968	10.073	8851	13.93
EPA SAMPLE NO.						
01 MW-6DL	9036	7.78	29109	10.57	18416	14.43
02 MW-1DL	8096	7.78	25457	10.57	14977	14.43
03 MW-11DL	8387	7.78	25419	10.57	14582	14.43
04 FB-1	14967	7.78	45140	10.57	24883	14.43
05 EB-1	7465	7.78	23044	10.57	12718	14.43

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.