

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3534

Client ID : SSTDCCC0.4

Date Analyzed: 11/10/2025

Lab File ID: BN038174.D

Time Analyzed: 10:14

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	8757	7.782	27774	10.58	15906	14.44
UPPER LIMIT	17514	8.282	55548	11.084	31812	14.941
LOWER LIMIT	4378.5	7.282	13887	10.084	7953	13.941
EPA SAMPLE NO.						
01 PB170427BL	8905	7.78	26771	10.57	15040	14.43
02 MW-EB-1DL	9993	7.78	33147	10.57	39685 *	14.43

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.