

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3534

Client ID : SSTDCCC040

Date Analyzed: 11/06/2025

Lab File ID: BF144180.D

Time Analyzed: 11:13

Instrument ID: BNA_F

GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	77162	6.875	299960	8.16	175155	9.92
UPPER LIMIT	154324	7.375	599920	8.657	350310	10.416
LOWER LIMIT	38581	6.375	149980	7.657	87577.5	9.416
EPA SAMPLE NO.						
01 MW-1	63903	6.88	245141	8.16	136285	9.91
02 MW-EB-1	62761	6.88	229494	8.16	120957	9.91

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.