

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3534

Client ID : SSTDCCC040

Date Analyzed: 11/06/2025

Lab File ID: BG064699.D

Time Analyzed: 12:13

Instrument ID: BNA_G

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	39621	8.206	163730	11.03	142094	14.82
UPPER LIMIT	79242	8.706	327460	11.526	284188	15.322
LOWER LIMIT	19810.5	7.706	81865	10.526	71047	14.322
EPA SAMPLE NO.						
01 MW-6	30988	8.20	121821	11.02	100117	14.82
02 MW-EB-3	33315	8.20	133129	11.02	109224	14.82

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.