

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: REMI02
 Lab Code: ACE SDG No.: Q3534
 Instrument ID: MSVOA_V Calibration Date/Time: 11/05/2025 11:34
 Lab File ID: VV039328.D Init. Calib. Date(s): 10/07/2025 10/07/2025
 Heated Purge: (Y/N) N Init. Calib. Time(s): 09:40 11:32
 GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	Amount	Conc.	%Dev	MAX%D
Chloroethane	50.000	52.07	-4.129997	20
Acetone	250.000	282.34	-12.9364	20
Methylene Chloride	50.000	57.85	-15.7	20