

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: REMI02
Lab Code: ACE SDG NO.: Q3534
Lab File ID: VV039328.D Date Analyzed: 11/05/2025
Instrument ID: MSVOA_V Time Analyzed: 11:34
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	644517	4.60	1006970	5.53	993503	8.77
UPPER LIMIT	1289030	5.096	2013950	6.029	1987010	9.273
LOWER LIMIT	322259	4.096	503487	5.029	496752	8.273
EPA SAMPLE NO.						
MW-11	938043	4.59	1744872	5.53	1594094	8.77
VV1105WBL01	856812	4.60	1628447	5.53	1466262	8.77
VV1105WBS01	708659	4.60	1110235	5.53	1080434	8.77

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.