

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: REMI02
Lab Code: ACE SDG NO.: Q3534
Lab File ID: VV039306.D Date Analyzed: 11/04/2025
Instrument ID: MSVOA_V Time Analyzed: 11:21
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	572200	11.172				
UPPER LIMIT	1144400	11.672				
LOWER LIMIT	286100	10.672				
EPA SAMPLE NO.						
MW-6	779349	11.17				
MW-1	725582	11.17				
MW-5	731965	11.17				
MW-EB-1	851194	11.17				
MW-EB-2	497981	11.17				
MW-EB-3	687973	11.17				
FB-1	578801	11.17				
EB-1	472586	11.17				
VV1104WBL01	828938	11.17				
VV1104WBS01	636116	11.17				
VV1104WBSD01	618050	11.17				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.