

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: REMI02
 Lab Code: ACE SDG NO.: Q3534
 Lab File ID: VV039328.D Date Analyzed: 11/05/2025
 Instrument ID: MSVOA_V Time Analyzed: 11:34
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	575078	11.172				
UPPER LIMIT	1150160	11.672				
LOWER LIMIT	287539	10.672				
EPA SAMPLE NO.						
MW-11	693574	11.17				
VV1105WBL01	623984	11.17				
VV1105WBS01	614058	11.17				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.