

DATA PACKAGE

SUB - DATA

PROJECT NAME : EDISON LANDFILL

REMINGTON & VERNICK ENGINEERS

2059 Springdale Road

Cherry Hill, NJ - 08003

Phone No: 856-795-9595

ORDER ID : Q3535

ATTENTION : Kyle Carlson



Cover Page

Order ID : Q3535

Project ID : Edison Landfill

Client : Remington & Vernick Engineers

Lab Sample Number

Q3535-01
Q3535-02
Q3535-03
Q3535-04
Q3535-05
Q3535-06
Q3535-07
Q3535-08
Q3535-09

Client Sample Number

MW-6
MW-1
MW-11
MW-5
MW-EB-1
MW-EB-2
MW-EB-3
FB-1
EB-1

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : _____

Date: 11/19/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012



*Alliance Technical Group - Akron
3310 Win St.
Cuyahoga Falls, Ohio 44223
TEL: (330) 253-8211 FAX: (330) 253-4489
Website: <http://www.settek.com>*

Yazmeen Gomez
ATG - NEWARK LAB
284 Sheffield Street
Mountainside, NJ 07092
TEL:
FAX:
RE: Q3535

Order No.: 25110242

Dear Yazmeen Gomez:

Alliance Technical Group - Akron received 9 sample(s) on 11/5/2025 for the analyses presented in the following report.

There were no problems with the analytical events associated with this report unless noted in the Case Narrative.

Quality control data is within laboratory defined or method specified acceptance limits except where noted.

If you have any questions regarding these tests results, please feel free to call the laboratory.

Sincerely,

A handwritten signature in black ink, appearing to read "Jennifer Woolf", is written over a large, light gray, semi-transparent watermark that says "PRELIMINARY".

Jennifer Woolf

Project Manager

3310 Win St.
Cuyahoga Falls, Ohio 44223

Arkansas 88-0735, California 2943, Colorado, Connecticut PH-0828, Florida NELAC E87688, Idaho OH00923, Illinois 200061, Indiana C-OH-13, ISO/IEC 17025:2017 119125 L22-544, Kansas E-10347, Kentucky (Underground Storage Tank) 3, Kentucky 90146, Maryland 339, Michigan 9988, Minnesota 1780279, Nevada OH009232020-1, New Hampshire 2996, New Jersey OH006, New York 11777, North Carolina 39705 and 631, North Dakota R-201, Ohio DW, Ohio VAP CL0052, Oklahoma 2019-155, Oregon OH200001, Pennsylvania 68-01335, Rhode Island LA000317, South Carolina 92016001, Texas T104704466-19-16, Utah OH009232020-12, Virginia VELAP 10381, West Virginia 9957C



Alliance Technical Group - Akron
3310 Win St.
Cuyahoga Falls, Ohio 44223
TEL: (330) 253-8211 FAX: (330) 253-4489
Website: <http://www.settek.com>

Case Narrative

WO#: 25110242

Date:

CLIENT: ATG - NEWARK LAB

Project: Q3535

WorkOrder Narrative:

25110242: This report in its entirety consists of the following documents: Cover Letter, Case Narrative, Analytical Results, QC Summary Report, Applicable Accreditation Information, Chain-of-Custody, Cooler Receipt Form, and other applicable forms as necessary. All documents contain the Alliance Technical Group Work Order Number assigned to this report.

Alliance Technical Group holds the accreditations/certifications listed at the bottom of the cover letter that may or may not pertain to this report. Please refer to the "Accreditation Program Analytes Report" for accredited analytes list.

The information contained in this analytical report is the sole property of Alliance Technical Group and that of the customer. It cannot be reproduced in any form without the consent of Alliance Technical Group or the customer for which this report was issued. The results contained in this report are only representative of the samples received. Conditions can vary at different times and at different sampling conditions. Alliance Technical Group is not responsible for use or interpretation of the data included herein.

All results for solid samples are reported on an "as received" or "wet weight" basis unless indicated as "dry weight" using the "-dry" designation on the reporting units.

This report is believed to meet all of the requirements of the accrediting agency, where applicable. Any comments or problems with the analytical events associated with this report are noted below.

Analytical Sequence Sample Notes:

25110242-001A SVOC-EPA1633_NPW: CCV exhibited high recoveries for ADONA and PFEESA. Potential high bias in bracketed samples. These analytes were ND in the sample. Sample exhibited a low recovery for IPS 13C3-PFBA. Potential high bias in associated IDAs. Potential matrix effects. Confirmed via extract dilution analysis.

25110242-002A SVOC-EPA1633_NPW: CCV exhibited high recoveries for ADONA and PFEESA. Potential high bias in bracketed samples. These analytes were ND in the sample.

25110242-002A SVOC-EPA1633_NPW: Sample exhibited a high recovery for IDA 13C4-PFHpA. Potential low bias in associated analytes. Confirmed via extract dilution analysis.

25110242-003A SVOC-EPA1633_NPW: CCV exhibited high recoveries for ADONA and PFEESA. Potential high bias in bracketed samples. These analytes were ND in the sample. Sample exhibited a

PRELIMINARY

Original



Alliance Technical Group - Akron
3310 Win St.
Cuyahoga Falls, Ohio 44223
TEL: (330) 253-8211 FAX: (330) 253-4489
Website: <http://www.settek.com>

Case Narrative

WO#: 25110242

Date:

CLIENT: ATG - NEWARK LAB

Project: Q3535

low recovery for IPS 13C3-PFBA. Potential high bias in associated IDAs. Potential matrix effects. Confirmed via extract dilution analysis.

25110242-006A SVOC-EPA1633_NPW: CCV exhibited high recoveries for ADONA and PFEESA. Potential high bias in bracketed samples. These analytes were ND in the sample. Sample exhibited a low recovery for IPS 13C3-PFBA. Potential high bias in associated IDAs. Potential matrix effects. Confirmed via extract dilution analysis.

25110242-007A SVOC-EPA1633_NPW: CCV exhibited high recoveries for ADONA and PFEESA. Potential high bias in bracketed samples. These analytes were ND in the sample.

25110242-008A SVOC-EPA1633_NPW: CCV exhibited high recoveries for ADONA and PFEESA. Potential high bias in bracketed samples. These analytes were ND in the sample.

25110242-009A SVOC-EPA1633_NPW: CCV exhibited high recoveries for ADONA and PFEESA. Potential high bias in bracketed samples. These analytes were ND in the sample.

PRELIMINARY

Original

These commonly used Qualifiers and Acronyms may or may not be present in this report.

Qualifiers

U	The compound was analyzed for but was not detected above the MDL.
J	The reported value is greater than the Method Detection Limit but less than the Reporting Limit.
H	The hold time for sample preparation and/or analysis was exceeded. Not Clean Water Act compliant.
D	The result is reported from a dilution.
E	The result exceeded the linear range of the calibration or is estimated due to interference.
MC	The result is below the Minimum Compound Limit.
*	The result exceeds the Regulatory Limit or Maximum Contamination Limit.
m	Manual integration was used to determine the area response.
d	Manual integration in which peak was deleted
N	The result is presumptive based on a Mass Spectral library search assuming a 1:1 response.
P	The second column confirmation exceeded 25% difference.
C	The result has been confirmed by GC/MS.
X	The result was not confirmed when GC/MS Analysis was performed.
B	The analyte was detected in the Method Blank at a concentration greater than the RL.
MB+	The analyte was detected in the Method Blank at a concentration greater than the MDL.
G	The ICB or CCB contained reportable amounts of analyte.
QC-/+	The CCV recovery failed low (-) or high (+).
R/QDR	The RPD was outside of accepted recovery limits.
QL-/+	The LCS or LCSD recovery failed low (-) or high (+).
QLR	The LCS/LCSD RPD was outside of accepted recovery limits.
QM-/+	The MS or MSD recovery failed low (-) or high (+).
QMR	The MS/MSD RPD was outside of accepted recovery limits.
QV-/+	The ICV recovery failed low (-) or high (+).
S	The spike result was outside of accepted recovery limits.
W	Samples were received outside temperature limits (0° – 6° C). Not Clean Water Act compliant.
Z	Deviation; A deviation from the method was performed; Please refer to the Case Narrative for additional information

Acronyms

ND	Not Detected	RL	Reporting Limit
QC	Quality Control	MDL	Method Detection Limit
MB	Method Blank	LOD	Level of Detection
LCS	Laboratory Control Sample	LOQ	Level of Quantitation
LCSD	Laboratory Control Sample Duplicate	PQL	Practical Quantitation Limit
QCS	Quality Control Sample	CRQL	Contract Required Quantitation Limit
DUP	Duplicate	PL	Permit Limit
MS	Matrix Spike	RegLvl	Regulatory Limit
MSD	Matrix Spike Duplicate	MCL	Maximum Contamination Limit
RPD	Relative Percent Different	MinCL	Minimum Compound Limit
ICV	Initial Calibration Verification	RA	Reanalysis
ICB	Initial Calibration Blank	RE	Reextraction
CCV	Continuing Calibration Verification	TIC	Tentatively Identified Compound
CCB	Continuing Calibration Blank	RT	Retention Time
RLC	Reporting Limit Check	CF	Calibration Factor

This list of Qualifiers and Acronyms reflects the most commonly utilized Qualifiers and Acronyms for reporting. Please refer to the Analytical Notes in the Case Narrative for any Qualifiers or Acronyms that do not appear in this list or for additional information regarding the use of these Qualifiers on reported data.

Original

PRELIMINARY



Alliance Technical Group - Akron
3310 Win St.
Cuyahoga Falls, Ohio 44223
TEL: (330) 253-8211 FAX: (330) 253-4489
Website: <http://www.settek.com>

Workorder Sample Summary

WO#: 25110242
19-Nov-25

CLIENT: ATG - NEWARK LAB
Project: Q3535

Lab SampleID	Client Sample ID	Tag No	Date Collected	Date Received	Matrix
25110242-001	MW-6		11/3/2025 10:45:00 AM	11/5/2025 10:30:00 AM	Non-Potable Water
25110242-002	MW-1		11/3/2025 11:50:00 AM	11/5/2025 10:30:00 AM	Non-Potable Water
25110242-003	MW-11		11/3/2025 12:00:00 PM	11/5/2025 10:30:00 AM	Non-Potable Water
25110242-004	MW-5		11/4/2025 10:20:00 AM	11/5/2025 10:30:00 AM	Non-Potable Water
25110242-005	MW-EB-1		11/3/2025 3:00:00 PM	11/5/2025 10:30:00 AM	Non-Potable Water
25110242-006	MW-EB-2		11/4/2025 10:05:00 AM	11/5/2025 10:30:00 AM	Non-Potable Water
25110242-007	MW-EB-3		11/3/2025 2:30:00 PM	11/5/2025 10:30:00 AM	Non-Potable Water
25110242-008	FB-1		11/4/2025 12:00:00 PM	11/5/2025 10:30:00 AM	Non-Potable Water
25110242-009	EB-1		11/4/2025 12:00:00 PM	11/5/2025 10:30:00 AM	Non-Potable Water

PRELIMINARY



Alliance Technical Group - Akron
3310 Win St.
Cuyahoga Falls, Ohio 44223
TEL: (330) 253-8211 FAX: (330) 253-4489
Website: <http://www.settek.com>

DATES REPORT

WO#: 25110242

19-Nov-25

Client: ATG - NEWARK LAB

Project: Q3535

Sample ID	Client Sample ID	Collection Date	Matrix	Test Name	Leachate Date	Prep Date	Analysis Date
25110242-001A	MW-6	11/3/2025 10:45:00 AM	Non-Potable Water	PFAS in Aqueous Samples (EPA Meth 1633Mod)		11/6/2025 7:49:02 AM	11/11/2025 8:21:00 PM
				PFAS in Aqueous Samples (EPA Meth 1633Mod)		11/6/2025 7:49:02 AM	11/11/2025 4:13:00 AM
				TSS for 1633			11/6/2025 3:49:00 PM
25110242-002A	MW-1	11/3/2025 11:50:00 AM		PFAS in Aqueous Samples (EPA Meth 1633Mod)		11/6/2025 7:49:02 AM	11/11/2025 8:35:00 PM
				PFAS in Aqueous Samples (EPA Meth 1633Mod)		11/6/2025 7:49:02 AM	11/11/2025 4:28:00 AM
				TSS for 1633			11/6/2025 3:49:00 PM
25110242-003A	MW-11	11/3/2025 12:00:00 PM		PFAS in Aqueous Samples (EPA Meth 1633Mod)		11/6/2025 7:49:02 AM	11/11/2025 4:42:00 AM
				PFAS in Aqueous Samples (EPA Meth 1633Mod)		11/6/2025 7:49:02 AM	11/11/2025 8:50:00 PM
				TSS for 1633			11/6/2025 3:49:00 PM
25110242-004A	MW-5	11/4/2025 10:20:00 AM		TSS for 1633			11/6/2025 3:49:00 PM
25110242-005A	MW-EB-1	11/3/2025 3:00:00 PM		TSS for 1633			11/6/2025 3:49:00 PM
25110242-006A	MW-EB-2	11/4/2025 10:05:00 AM		PFAS in Aqueous Samples (EPA Meth 1633Mod)		11/6/2025 7:49:02 AM	11/11/2025 9:34:00 PM
				PFAS in Aqueous Samples (EPA Meth 1633Mod)		11/6/2025 7:49:02 AM	11/11/2025 5:26:00 AM
				TSS for 1633			11/6/2025 3:49:00 PM
25110242-007A	MW-EB-3	11/3/2025 2:30:00 PM		PFAS in Aqueous Samples (EPA Meth 1633Mod)		11/6/2025 7:49:02 AM	11/11/2025 9:49:00 PM
				PFAS in Aqueous Samples (EPA Meth 1633Mod)		11/6/2025 7:49:02 AM	11/11/2025 5:41:00 AM
				TSS for 1633			11/6/2025 3:49:00 PM
25110242-008A	FB-1	11/4/2025 12:00:00 PM		PFAS in Aqueous Samples (EPA Meth 1633Mod)		11/6/2025 7:49:02 AM	11/11/2025 5:56:00 AM
				TSS for 1633			11/10/2025 3:00:00 PM

Original

PRELIMINARY



Alliance Technical Group - Akron
3310 Win St.
Cuyahoga Falls, Ohio 44223
TEL: (330) 253-8211 FAX: (330) 253-4489
Website: <http://www.settek.com>

DATES REPORT

WO#: 25110242

19-Nov-25

Client: ATG - NEWARK LAB

Project: Q3535

Sample ID	Client Sample ID	Collection Date	Matrix	Test Name	Leachate Date	Prep Date	Analysis Date
25110242-009A	EB-1	11/4/2025 12:00:00 PM	Non-Potable Water	PFAS in Aqueous Samples (EPA Meth 1633Mod) TSS for 1633		11/6/2025 7:49:02 AM	11/11/2025 6:10:00 AM 11/10/2025 3:00:00 PM

PRELIMINARY

Page 7 of 45

Original



Alliance Technical Group - Akron
 3310 Win St.
 Cuyahoga Falls, Ohio 44223
 TEL: (330) 253-8211 FAX: (330) 253-4489
 Website: <http://www.settek.com>

WO#: 25110242
 Date Reported:
 Company: ATG - NEWARK LAB
 Address: 284 Sheffield Street
 Mountainside NJ 07092
 Received: 11/5/2025
 Project#: Q3535

Client ID#	Lab ID#	Collected	Analyte	Result Units	Qual	Matrix	Method	DF	MDL	PQL	Run	Analyst
MW-6	001	11/3/2025	11CI-PF3OUdS	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.226	8.28	11/11/2025	AEH
MW-6	001	11/3/2025	3:3 FTCA	2.79 ng/L	J	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.476	10.3	11/11/2025	AEH
MW-6	001	11/3/2025	4:2 FTS	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	1.77	8.28	11/11/2025	AEH
MW-6	001	11/3/2025	5:3 FTCA	40.3 ng/L	J	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	1.79	51.7	11/11/2025	AEH
MW-6	001	11/3/2025	6:2 FTS	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	2.71	8.28	11/11/2025	AEH
MW-6	001	11/3/2025	7:3 FTCA	7.28 ng/L	J	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	3.16	51.7	11/11/2025	AEH
MW-6	001	11/3/2025	8:2 FTS	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	3.18	8.28	11/11/2025	AEH
MW-6	001	11/3/2025	9CL-PF3ONS	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.311	8.28	11/11/2025	AEH
MW-6	001	11/3/2025	ADONA	ND ng/L	UQC+	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.353	8.28	11/11/2025	AEH
MW-6	001	11/3/2025	HFPO-DA	11.0 ng/L		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.431	8.28	11/11/2025	AEH
MW-6	001	11/3/2025	NEtFOSA	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.378	2.07	11/11/2025	AEH
MW-6	001	11/3/2025	NEtFOSAA	18.8 ng/L		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	1.05	2.07	11/11/2025	AEH
MW-6	001	11/3/2025	NEtFOSE	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	3.64	20.7	11/11/2025	AEH
MW-6	001	11/3/2025	NFDHA	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.301	4.14	11/11/2025	AEH



Alliance Technical Group - Akron
 3310 Win St.
 Cuyahoga Falls, Ohio 44223
 TEL: (330) 253-8211 FAX: (330) 253-4489
 Website: <http://www.settek.com>

WO#: 25110242
 Date Reported:
 Company: ATG - NEWARK LAB
 Address: 284 Sheffield Street
 Mountainside NJ 07092
 Received: 11/5/2025
 Project#: Q3535

Client ID#	Lab ID#	Collected	Analyte	Result	Units	Qual	Matrix	Method	DF	MDL	PQL	Run	Analyst
MW-6	001	11/3/2025	NMeFOSA	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.634	2.07	11/11/2025	AEH
MW-6	001	11/3/2025	NMeFOSAA	9.20	ng/L		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.922	2.07	11/11/2025	AEH
MW-6	001	11/3/2025	NMeFOSE	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	3.90	20.7	11/11/2025	AEH
MW-6	001	11/3/2025	PFBA	73.9	ng/L		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.850	8.28	11/11/2025	AEH
MW-6	001	11/3/2025	PFBS	9.41	ng/L		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.504	2.07	11/11/2025	AEH
MW-6	001	11/3/2025	PFDA	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.661	2.07	11/11/2025	AEH
MW-6	001	11/3/2025	PFDaA	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.594	2.07	11/11/2025	AEH
MW-6	001	11/3/2025	PFDoS	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.413	2.07	11/11/2025	AEH
MW-6	001	11/3/2025	PFDS	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.426	2.07	11/11/2025	AEH
MW-6	001	11/3/2025	PFEESA	ND	ng/L	UQC+	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.250	4.14	11/11/2025	AEH
MW-6	001	11/3/2025	PFHpA	53.0	ng/L		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.693	2.07	11/11/2025	AEH
MW-6	001	11/3/2025	PFHpS	4.91	ng/L		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.660	2.07	11/11/2025	AEH
MW-6	001	11/3/2025	PFHxA	49.9	ng/L		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.419	2.07	11/11/2025	AEH
MW-6	001	11/3/2025	PFHxS	26.4	ng/L		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	1.26	2.07	11/11/2025	AEH



Alliance Technical Group - Akron
 3310 Win St.
 Cuyahoga Falls, Ohio 44223
 TEL: (330) 253-8211 FAX: (330) 253-4489
 Website: <http://www.settek.com>

WO#: 25110242
 Date Reported:
 Company: ATG - NEWARK LAB
 Address: 284 Sheffield Street
 Mountainside NJ 07092
 Received: 11/5/2025
 Project#: Q3535

Client ID#	Lab ID#	Collected	Analyte	Result Units	Qual	Matrix	Method	DF	MDL	PQL	Run	Analyst
MW-6	001	11/3/2025	PFMBA	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.287	4.14	11/11/2025	AEH
MW-6	001	11/3/2025	PFMPA	0.352 ng/L	J	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.138	4.14	11/11/2025	AEH
MW-6	001	11/3/2025	PFNA	7.76 ng/L		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.391	2.07	11/11/2025	AEH
MW-6	001	11/3/2025	PFNS	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.842	2.07	11/11/2025	AEH
MW-6	001	11/3/2025	PFOA	251 ng/L		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.553	2.07	11/11/2025	AEH
MW-6	001	11/3/2025	PFOS	118 ng/L		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	1.36	2.07	11/11/2025	AEH
MW-6	001	11/3/2025	PFOSA	1.82 ng/L	J	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.306	2.07	11/11/2025	AEH
MW-6	001	11/3/2025	PFPeA	16.3 ng/L		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.603	4.14	11/11/2025	AEH
MW-6	001	11/3/2025	PFPeS	6.86 ng/L		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.692	2.07	11/11/2025	AEH
MW-6	001	11/3/2025	PFTeDA	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.448	2.07	11/11/2025	AEH
MW-6	001	11/3/2025	PFTTrDA	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.344	2.07	11/11/2025	AEH
MW-6	001	11/3/2025	PFUnA	1.27 ng/L	J	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.741	2.07	11/11/2025	AEH
MW-6	001	11/3/2025	Surr: 13C2-PFDA	78.4 %Rec		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0		11/11/2025	AEH
MW-6	001	11/3/2025	Surr: 13C2-PFHxA	62.4 %Rec		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0		11/11/2025	AEH



Alliance Technical Group - Akron
3310 Win St.
Cuyahoga Falls, Ohio 44223
TEL: (330) 253-8211 FAX: (330) 253-4489
Website: <http://www.settek.com>

WO#: 25110242
Date Reported:
Company: ATG - NEWARK LAB
Address: 284 Sheffield Street
Mountainside NJ 07092
Received: 11/5/2025
Project#: Q3535

Client ID#	Lab ID#	Collected	Analyte	Result Units	Qual	Matrix	Method	DF	MDL	PQL	Run	Analyst
MW-6	001	11/3/2025	Surr: 13C3-PFBA	49.0 %Rec	S	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0		11/11/2025	AEH
MW-6	001	11/3/2025	Surr: 13C4-PFOA	79.2 %Rec		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0		11/11/2025	AEH
MW-6	001	11/3/2025	Surr: 13C4-PFOS	84.7 %Rec		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0		11/11/2025	AEH
MW-6	001	11/3/2025	Surr: 13C5-PFNA	82.4 %Rec		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0		11/11/2025	AEH
MW-6	001	11/3/2025	Surr: 18O2-PFHxS	81.3 %Rec		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0		11/11/2025	AEH



Alliance Technical Group - Akron
 3310 Win St.
 Cuyahoga Falls, Ohio 44223
 TEL: (330) 253-8211 FAX: (330) 253-4489
 Website: <http://www.settek.com>

WO#: 25110242
 Date Reported:
 Company: ATG - NEWARK LAB
 Address: 284 Sheffield Street
 Mountainside NJ 07092
 Received: 11/5/2025
 Project#: Q3535

Client ID#	Lab ID#	Collected	Analyte	Result Units	Qual	Matrix	Method	DF	MDL	PQL	Run	Analyst
MW-1	002	11/3/2025	11CI-PF3OUdS	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.219	8.03	11/11/2025	AEH
MW-1	002	11/3/2025	3:3 FTCA	6.67 ng/L	J	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.462	10.0	11/11/2025	AEH
MW-1	002	11/3/2025	4:2 FTS	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	1.71	8.03	11/11/2025	AEH
MW-1	002	11/3/2025	5:3 FTCA	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	1.73	50.2	11/11/2025	AEH
MW-1	002	11/3/2025	6:2 FTS	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	2.63	8.03	11/11/2025	AEH
MW-1	002	11/3/2025	7:3 FTCA	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	3.06	50.2	11/11/2025	AEH
MW-1	002	11/3/2025	8:2 FTS	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	3.09	8.03	11/11/2025	AEH
MW-1	002	11/3/2025	9CL-PF3ONS	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.301	8.03	11/11/2025	AEH
MW-1	002	11/3/2025	ADONA	ND ng/L	UQC+	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.342	8.03	11/11/2025	AEH
MW-1	002	11/3/2025	HFPO-DA	1.19 ng/L	J	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.418	8.03	11/11/2025	AEH
MW-1	002	11/3/2025	NEtFOSA	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.367	2.01	11/11/2025	AEH
MW-1	002	11/3/2025	NEtFOSAA	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	1.02	2.01	11/11/2025	AEH
MW-1	002	11/3/2025	NEtFOSE	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	3.53	20.1	11/11/2025	AEH
MW-1	002	11/3/2025	NFDHA	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.292	4.01	11/11/2025	AEH



Alliance Technical Group - Akron
 3310 Win St.
 Cuyahoga Falls, Ohio 44223
 TEL: (330) 253-8211 FAX: (330) 253-4489
 Website: <http://www.settek.com>

WO#: 25110242
 Date Reported:
 Company: ATG - NEWARK LAB
 Address: 284 Sheffield Street
 Mountainside NJ 07092
 Received: 11/5/2025
 Project#: Q3535

Client ID#	Lab ID#	Collected	Analyte	Result	Units	Qual	Matrix	Method	DF	MDL	PQL	Run	Analyst
MW-1	002	11/3/2025	NMeFOSA	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.614	2.01	11/11/2025	AEH
MW-1	002	11/3/2025	NMeFOSAA	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.894	2.01	11/11/2025	AEH
MW-1	002	11/3/2025	NMeFOSE	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	3.79	20.1	11/11/2025	AEH
MW-1	002	11/3/2025	PFBA	218	ng/L		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.824	8.03	11/11/2025	AEH
MW-1	002	11/3/2025	PFBS	12.0	ng/L		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.489	2.01	11/11/2025	AEH
MW-1	002	11/3/2025	PFDA	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.641	2.01	11/11/2025	AEH
MW-1	002	11/3/2025	PFDaA	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.576	2.01	11/11/2025	AEH
MW-1	002	11/3/2025	PFDoS	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.401	2.01	11/11/2025	AEH
MW-1	002	11/3/2025	PFDS	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.413	2.01	11/11/2025	AEH
MW-1	002	11/3/2025	PFEESA	ND	ng/L	UQC+	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.242	4.01	11/11/2025	AEH
MW-1	002	11/3/2025	PFHpA	67.1	ng/L		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.672	2.01	11/11/2025	AEH
MW-1	002	11/3/2025	PFHpS	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.640	2.01	11/11/2025	AEH
MW-1	002	11/3/2025	PFHxA	139	ng/L		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.407	2.01	11/11/2025	AEH
MW-1	002	11/3/2025	PFHxS	34.3	ng/L		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	1.23	2.01	11/11/2025	AEH



Alliance Technical Group - Akron
 3310 Win St.
 Cuyahoga Falls, Ohio 44223
 TEL: (330) 253-8211 FAX: (330) 253-4489
 Website: <http://www.settek.com>

WO#: 25110242
 Date Reported:
 Company: ATG - NEWARK LAB
 Address: 284 Sheffield Street
 Mountainside NJ 07092
 Received: 11/5/2025
 Project#: Q3535

Client ID#	Lab ID#	Collected	Analyte	Result Units	Qual	Matrix	Method	DF	MDL	PQL	Run	Analyst
MW-1	002	11/3/2025	PFMBA	0.381 ng/L	J	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.278	4.01	11/11/2025	AEH
MW-1	002	11/3/2025	PFMPA	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.133	4.01	11/11/2025	AEH
MW-1	002	11/3/2025	PFNA	1.39 ng/L	J	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.380	2.01	11/11/2025	AEH
MW-1	002	11/3/2025	PFNS	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.816	2.01	11/11/2025	AEH
MW-1	002	11/3/2025	PFOA	223 ng/L		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.536	2.01	11/11/2025	AEH
MW-1	002	11/3/2025	PFOS	3.23 ng/L		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	1.32	2.01	11/11/2025	AEH
MW-1	002	11/3/2025	PFOSA	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.296	2.01	11/11/2025	AEH
MW-1	002	11/3/2025	PFPeA	84.8 ng/L		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.585	4.01	11/11/2025	AEH
MW-1	002	11/3/2025	PFPeS	6.67 ng/L		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.671	2.01	11/11/2025	AEH
MW-1	002	11/3/2025	PFTeDA	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.434	2.01	11/11/2025	AEH
MW-1	002	11/3/2025	PFTTrDA	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.333	2.01	11/11/2025	AEH
MW-1	002	11/3/2025	PFUnA	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.718	2.01	11/11/2025	AEH
MW-1	002	11/3/2025	Surr: 13C2-PFDA	85.0 %Rec		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0		11/11/2025	AEH
MW-1	002	11/3/2025	Surr: 13C2-PFHxA	61.6 %Rec		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0		11/11/2025	AEH



Alliance Technical Group - Akron
3310 Win St.
Cuyahoga Falls, Ohio 44223
TEL: (330) 253-8211 FAX: (330) 253-4489
Website: <http://www.settek.com>

WO#: 25110242
Date Reported:
Company: ATG - NEWARK LAB
Address: 284 Sheffield Street
Mountainside NJ 07092
Received: 11/5/2025
Project#: Q3535

Client ID#	Lab ID#	Collected	Analyte	Result	Units	Qual	Matrix	Method	DF	MDL	PQL	Run	Analyst
MW-1	002	11/3/2025	Surr: 13C3-PFBA	53.4	%Rec		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0		11/11/2025	AEH
MW-1	002	11/3/2025	Surr: 13C4-PFOA	58.8	%Rec		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0		11/11/2025	AEH
MW-1	002	11/3/2025	Surr: 13C4-PFOS	93.7	%Rec		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0		11/11/2025	AEH
MW-1	002	11/3/2025	Surr: 13C5-PFNA	84.1	%Rec		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0		11/11/2025	AEH
MW-1	002	11/3/2025	Surr: 18O2-PFHxS	96.3	%Rec		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0		11/11/2025	AEH



Alliance Technical Group - Akron
 3310 Win St.
 Cuyahoga Falls, Ohio 44223
 TEL: (330) 253-8211 FAX: (330) 253-4489
 Website: <http://www.settek.com>

WO#: 25110242
 Date Reported:
 Company: ATG - NEWARK LAB
 Address: 284 Sheffield Street
 Mountainside NJ 07092
 Received: 11/5/2025
 Project#: Q3535

Client ID#	Lab ID#	Collected	Analyte	Result	Units	Qual	Matrix	Method	DF	MDL	PQL	Run	Analyst
MW-11	003	11/3/2025	11CI-PF3OUdS	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.203	7.45	11/11/2025	AEH
MW-11	003	11/3/2025	3:3 FTCA	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.429	9.32	11/11/2025	AEH
MW-11	003	11/3/2025	4:2 FTS	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	1.59	7.45	11/11/2025	AEH
MW-11	003	11/3/2025	5:3 FTCA	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	1.61	46.6	11/11/2025	AEH
MW-11	003	11/3/2025	6:2 FTS	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	2.44	7.45	11/11/2025	AEH
MW-11	003	11/3/2025	7:3 FTCA	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	2.84	46.6	11/11/2025	AEH
MW-11	003	11/3/2025	8:2 FTS	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	2.87	7.45	11/11/2025	AEH
MW-11	003	11/3/2025	9CL-PF3ONS	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.280	7.45	11/11/2025	AEH
MW-11	003	11/3/2025	ADONA	ND	ng/L	UQC+	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.318	7.45	11/11/2025	AEH
MW-11	003	11/3/2025	HFPO-DA	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.388	7.45	11/11/2025	AEH
MW-11	003	11/3/2025	NEtFOSA	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.341	1.86	11/11/2025	AEH
MW-11	003	11/3/2025	NEtFOSAA	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.946	1.86	11/11/2025	AEH
MW-11	003	11/3/2025	NEtFOSE	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	3.28	18.6	11/11/2025	AEH
MW-11	003	11/3/2025	NFDHA	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.271	3.73	11/11/2025	AEH



Alliance Technical Group - Akron
 3310 Win St.
 Cuyahoga Falls, Ohio 44223
 TEL: (330) 253-8211 FAX: (330) 253-4489
 Website: <http://www.settek.com>

WO#: 25110242
 Date Reported:
 Company: ATG - NEWARK LAB
 Address: 284 Sheffield Street
 Mountainside NJ 07092
 Received: 11/5/2025
 Project#: Q3535

Client ID#	Lab ID#	Collected	Analyte	Result	Units	Qual	Matrix	Method	DF	MDL	PQL	Run	Analyst
MW-11	003	11/3/2025	NMeFOSA	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.571	1.86	11/11/2025	AEH
MW-11	003	11/3/2025	NMeFOSAA	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.830	1.86	11/11/2025	AEH
MW-11	003	11/3/2025	NMeFOSE	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	3.52	18.6	11/11/2025	AEH
MW-11	003	11/3/2025	PFBA	9.61	ng/L		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.765	7.45	11/11/2025	AEH
MW-11	003	11/3/2025	PFBS	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.454	1.86	11/11/2025	AEH
MW-11	003	11/3/2025	PFDA	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.595	1.86	11/11/2025	AEH
MW-11	003	11/3/2025	PFDaA	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.535	1.86	11/11/2025	AEH
MW-11	003	11/3/2025	PFDoS	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.372	1.86	11/11/2025	AEH
MW-11	003	11/3/2025	PFDS	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.384	1.86	11/11/2025	AEH
MW-11	003	11/3/2025	PFEESA	ND	ng/L	UQC+	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.225	3.73	11/11/2025	AEH
MW-11	003	11/3/2025	PFHpA	0.745	ng/L	J	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.624	1.86	11/11/2025	AEH
MW-11	003	11/3/2025	PFHpS	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.595	1.86	11/11/2025	AEH
MW-11	003	11/3/2025	PFHxA	3.01	ng/L		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.378	1.86	11/11/2025	AEH
MW-11	003	11/3/2025	PFHxS	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	1.14	1.86	11/11/2025	AEH



Alliance Technical Group - Akron
 3310 Win St.
 Cuyahoga Falls, Ohio 44223
 TEL: (330) 253-8211 FAX: (330) 253-4489
 Website: <http://www.settek.com>

WO#: 25110242
 Date Reported:
 Company: ATG - NEWARK LAB
 Address: 284 Sheffield Street
 Mountainside NJ 07092
 Received: 11/5/2025
 Project#: Q3535

Client ID#	Lab ID#	Collected	Analyte	Result Units	Qual	Matrix	Method	DF	MDL	PQL	Run	Analyst
MW-11	003	11/3/2025	PFMBA	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.258	3.73	11/11/2025	AEH
MW-11	003	11/3/2025	PFMPA	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.124	3.73	11/11/2025	AEH
MW-11	003	11/3/2025	PFNA	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.353	1.86	11/11/2025	AEH
MW-11	003	11/3/2025	PFNS	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.758	1.86	11/11/2025	AEH
MW-11	003	11/3/2025	PFOA	1.36 ng/L	J	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.498	1.86	11/11/2025	AEH
MW-11	003	11/3/2025	PFOS	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	1.22	1.86	11/11/2025	AEH
MW-11	003	11/3/2025	PFOSA	1.77 ng/L	J	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.275	1.86	11/11/2025	AEH
MW-11	003	11/3/2025	PFPeA	2.74 ng/L	J	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.543	3.73	11/11/2025	AEH
MW-11	003	11/3/2025	PFPeS	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.623	1.86	11/11/2025	AEH
MW-11	003	11/3/2025	PFTeDA	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.403	1.86	11/11/2025	AEH
MW-11	003	11/3/2025	PFTTrDA	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.309	1.86	11/11/2025	AEH
MW-11	003	11/3/2025	PFUnA	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.667	1.86	11/11/2025	AEH
MW-11	003	11/3/2025	Surr: 13C2-PFDA	74.0 %Rec		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0		11/11/2025	AEH
MW-11	003	11/3/2025	Surr: 13C2-PFHxA	65.9 %Rec		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0		11/11/2025	AEH



Alliance Technical Group - Akron
3310 Win St.
Cuyahoga Falls, Ohio 44223
TEL: (330) 253-8211 FAX: (330) 253-4489
Website: <http://www.settek.com>

WO#: 25110242
Date Reported:
Company: ATG - NEWARK LAB
Address: 284 Sheffield Street
Mountainside NJ 07092
Received: 11/5/2025
Project#: Q3535

Client ID#	Lab ID#	Collected	Analyte	Result Units	Qual	Matrix	Method	DF	MDL	PQL	Run	Analyst
MW-11	003	11/3/2025	Surr: 13C3-PFBA	49.5 %Rec	S	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0		11/11/2025	AEH
MW-11	003	11/3/2025	Surr: 13C4-PFOA	67.3 %Rec		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0		11/11/2025	AEH
MW-11	003	11/3/2025	Surr: 13C4-PFOS	74.0 %Rec		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0		11/11/2025	AEH
MW-11	003	11/3/2025	Surr: 13C5-PFNA	80.7 %Rec		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0		11/11/2025	AEH
MW-11	003	11/3/2025	Surr: 18O2-PFHxS	75.4 %Rec		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0		11/11/2025	AEH
Client ID#	Lab ID#	Collected	Analyte	Result Units	Qual	Matrix	Method	DF	MDL	PQL	Run	Analyst
Client ID#	Lab ID#	Collected	Analyte	Result Units	Qual	Matrix	Method	DF	MDL	PQL	Run	Analyst



Alliance Technical Group - Akron
 3310 Win St.
 Cuyahoga Falls, Ohio 44223
 TEL: (330) 253-8211 FAX: (330) 253-4489
 Website: <http://www.settek.com>

WO#: 25110242
 Date Reported:
 Company: ATG - NEWARK LAB
 Address: 284 Sheffield Street
 Mountainside NJ 07092
 Received: 11/5/2025
 Project#: Q3535

Client ID#	Lab ID#	Collected	Analyte	Result Units	Qual	Matrix	Method	DF	MDL	PQL	Run	Analyst
MW-EB-2	006	11/4/2025	11CI-PF3OUdS	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.217	7.96	11/11/2025	AEH
MW-EB-2	006	11/4/2025	3:3 FTCA	0.757 ng/L	J	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.458	9.96	11/11/2025	AEH
MW-EB-2	006	11/4/2025	4:2 FTS	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	1.70	7.96	11/11/2025	AEH
MW-EB-2	006	11/4/2025	5:3 FTCA	2.97 ng/L	J	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	1.72	49.8	11/11/2025	AEH
MW-EB-2	006	11/4/2025	6:2 FTS	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	2.61	7.96	11/11/2025	AEH
MW-EB-2	006	11/4/2025	7:3 FTCA	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	3.04	49.8	11/11/2025	AEH
MW-EB-2	006	11/4/2025	8:2 FTS	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	3.06	7.96	11/11/2025	AEH
MW-EB-2	006	11/4/2025	9CL-PF3ONS	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.299	7.96	11/11/2025	AEH
MW-EB-2	006	11/4/2025	ADONA	ND ng/L	UQC+	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.340	7.96	11/11/2025	AEH
MW-EB-2	006	11/4/2025	HFPO-DA	9.50 ng/L		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.415	7.96	11/11/2025	AEH
MW-EB-2	006	11/4/2025	NEtFOSA	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.364	1.99	11/11/2025	AEH
MW-EB-2	006	11/4/2025	NEtFOSAA	10.1 ng/L		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	1.01	1.99	11/11/2025	AEH
MW-EB-2	006	11/4/2025	NEtFOSE	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	3.50	19.9	11/11/2025	AEH
MW-EB-2	006	11/4/2025	NFDHA	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.290	3.98	11/11/2025	AEH



Alliance Technical Group - Akron
 3310 Win St.
 Cuyahoga Falls, Ohio 44223
 TEL: (330) 253-8211 FAX: (330) 253-4489
 Website: <http://www.settek.com>

WO#: 25110242
 Date Reported:
 Company: ATG - NEWARK LAB
 Address: 284 Sheffield Street
 Mountainside NJ 07092
 Received: 11/5/2025
 Project#: Q3535

Client ID#	Lab ID#	Collected	Analyte	Result Units	Qual	Matrix	Method	DF	MDL	PQL	Run	Analyst
MW-EB-2	006	11/4/2025	NMeFOSA	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.610	1.99	11/11/2025	AEH
MW-EB-2	006	11/4/2025	NMeFOSAA	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.887	1.99	11/11/2025	AEH
MW-EB-2	006	11/4/2025	NMeFOSE	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	3.76	19.9	11/11/2025	AEH
MW-EB-2	006	11/4/2025	PFBA	25.4 ng/L		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.818	7.96	11/11/2025	AEH
MW-EB-2	006	11/4/2025	PFBS	14.4 ng/L		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.485	1.99	11/11/2025	AEH
MW-EB-2	006	11/4/2025	PFDA	0.787 ng/L	J	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.636	1.99	11/11/2025	AEH
MW-EB-2	006	11/4/2025	PFDaA	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.572	1.99	11/11/2025	AEH
MW-EB-2	006	11/4/2025	PFDoS	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.398	1.99	11/11/2025	AEH
MW-EB-2	006	11/4/2025	PFDS	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.410	1.99	11/11/2025	AEH
MW-EB-2	006	11/4/2025	PFEESA	ND ng/L	UQC+	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.241	3.98	11/11/2025	AEH
MW-EB-2	006	11/4/2025	PFHpA	20.6 ng/L		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.667	1.99	11/11/2025	AEH
MW-EB-2	006	11/4/2025	PFHpS	3.77 ng/L		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.636	1.99	11/11/2025	AEH
MW-EB-2	006	11/4/2025	PFHxA	29.0 ng/L		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.404	1.99	11/11/2025	AEH
MW-EB-2	006	11/4/2025	PFHxS	35.0 ng/L		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	1.22	1.99	11/11/2025	AEH



Alliance Technical Group - Akron
 3310 Win St.
 Cuyahoga Falls, Ohio 44223
 TEL: (330) 253-8211 FAX: (330) 253-4489
 Website: <http://www.settek.com>

WO#: 25110242
 Date Reported:
 Company: ATG - NEWARK LAB
 Address: 284 Sheffield Street
 Mountainside NJ 07092
 Received: 11/5/2025
 Project#: Q3535

Client ID#	Lab ID#	Collected	Analyte	Result Units	Qual	Matrix	Method	DF	MDL	PQL	Run	Analyst
MW-EB-2	006	11/4/2025	PFMBA	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.276	3.98	11/11/2025	AEH
MW-EB-2	006	11/4/2025	PFMPA	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.132	3.98	11/11/2025	AEH
MW-EB-2	006	11/4/2025	PFNA	5.20 ng/L		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.377	1.99	11/11/2025	AEH
MW-EB-2	006	11/4/2025	PFNS	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.810	1.99	11/11/2025	AEH
MW-EB-2	006	11/4/2025	PFOA	97.6 ng/L		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.532	1.99	11/11/2025	AEH
MW-EB-2	006	11/4/2025	PFOS	168 ng/L		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	1.31	1.99	11/11/2025	AEH
MW-EB-2	006	11/4/2025	PFOSA	2.34 ng/L		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.294	1.99	11/11/2025	AEH
MW-EB-2	006	11/4/2025	PFPeA	23.7 ng/L		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.580	3.98	11/11/2025	AEH
MW-EB-2	006	11/4/2025	PFPeS	8.93 ng/L		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.666	1.99	11/11/2025	AEH
MW-EB-2	006	11/4/2025	PFTeDA	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.431	1.99	11/11/2025	AEH
MW-EB-2	006	11/4/2025	PFTrDA	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.331	1.99	11/11/2025	AEH
MW-EB-2	006	11/4/2025	PFUnA	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.713	1.99	11/11/2025	AEH
MW-EB-2	006	11/4/2025	Surr: 13C2-PFDA	69.4 %Rec		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0		11/11/2025	AEH
MW-EB-2	006	11/4/2025	Surr: 13C2-PFHxA	60.4 %Rec		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0		11/11/2025	AEH



Alliance Technical Group - Akron
3310 Win St.
Cuyahoga Falls, Ohio 44223
TEL: (330) 253-8211 FAX: (330) 253-4489
Website: <http://www.settek.com>

WO#: 25110242
Date Reported:
Company: ATG - NEWARK LAB
Address: 284 Sheffield Street
Mountainside NJ 07092
Received: 11/5/2025
Project#: Q3535

Client ID#	Lab ID#	Collected	Analyte	Result Units	Qual	Matrix	Method	DF	MDL	PQL	Run	Analyst
MW-EB-2	006	11/4/2025	Surr: 13C3-PFBA	46.2 %Rec	S	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0		11/11/2025	AEH
MW-EB-2	006	11/4/2025	Surr: 13C4-PFOA	73.6 %Rec		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0		11/11/2025	AEH
MW-EB-2	006	11/4/2025	Surr: 13C4-PFOS	90.3 %Rec		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0		11/11/2025	AEH
MW-EB-2	006	11/4/2025	Surr: 13C5-PFNA	64.0 %Rec		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0		11/11/2025	AEH
MW-EB-2	006	11/4/2025	Surr: 18O2-PFHxS	86.2 %Rec		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0		11/11/2025	AEH



Alliance Technical Group - Akron
3310 Win St.
Cuyahoga Falls, Ohio 44223
TEL: (330) 253-8211 FAX: (330) 253-4489
Website: <http://www.settek.com>

WO#: 25110242
Date Reported:
Company: ATG - NEWARK LAB
Address: 284 Sheffield Street
Mountainside NJ 07092
Received: 11/5/2025
Project#: Q3535

Client ID#	Lab ID#	Collected	Analyte	Result	Units	Qual	Matrix	Method	DF	MDL	PQL	Run	Analyst
MW-EB-3	007	11/3/2025	11CI-PF3OUdS	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.225	8.27	11/11/2025	AEH
MW-EB-3	007	11/3/2025	3:3 FTCA	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.476	10.3	11/11/2025	AEH
MW-EB-3	007	11/3/2025	4:2 FTS	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	1.77	8.27	11/11/2025	AEH
MW-EB-3	007	11/3/2025	5:3 FTCA	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	1.79	51.7	11/11/2025	AEH
MW-EB-3	007	11/3/2025	6:2 FTS	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	2.71	8.27	11/11/2025	AEH
MW-EB-3	007	11/3/2025	7:3 FTCA	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	3.16	51.7	11/11/2025	AEH
MW-EB-3	007	11/3/2025	8:2 FTS	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	3.18	8.27	11/11/2025	AEH
MW-EB-3	007	11/3/2025	9CL-PF3ONS	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.311	8.27	11/11/2025	AEH
MW-EB-3	007	11/3/2025	ADONA	ND	ng/L	UQC+	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.353	8.27	11/11/2025	AEH
MW-EB-3	007	11/3/2025	HFPO-DA	4.65	ng/L	J	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.431	8.27	11/11/2025	AEH
MW-EB-3	007	11/3/2025	NEtFOSA	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.378	2.07	11/11/2025	AEH
MW-EB-3	007	11/3/2025	NEtFOSAA	11.7	ng/L		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	1.05	2.07	11/11/2025	AEH
MW-EB-3	007	11/3/2025	NEtFOSE	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	3.64	20.7	11/11/2025	AEH
MW-EB-3	007	11/3/2025	NFDHA	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.301	4.14	11/11/2025	AEH



Alliance Technical Group - Akron
 3310 Win St.
 Cuyahoga Falls, Ohio 44223
 TEL: (330) 253-8211 FAX: (330) 253-4489
 Website: <http://www.settek.com>

WO#: 25110242
 Date Reported:
 Company: ATG - NEWARK LAB
 Address: 284 Sheffield Street
 Mountainside NJ 07092
 Received: 11/5/2025
 Project#: Q3535

Client ID#	Lab ID#	Collected	Analyte	Result	Units	Qual	Matrix	Method	DF	MDL	PQL	Run	Analyst
MW-EB-3	007	11/3/2025	NMeFOSA	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.633	2.07	11/11/2025	AEH
MW-EB-3	007	11/3/2025	NMeFOSAA	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.922	2.07	11/11/2025	AEH
MW-EB-3	007	11/3/2025	NMeFOSE	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	3.90	20.7	11/11/2025	AEH
MW-EB-3	007	11/3/2025	PFBA	21.3	ng/L		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.849	8.27	11/11/2025	AEH
MW-EB-3	007	11/3/2025	PFBS	15.7	ng/L		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.504	2.07	11/11/2025	AEH
MW-EB-3	007	11/3/2025	PFDA	0.827	ng/L	J	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.660	2.07	11/11/2025	AEH
MW-EB-3	007	11/3/2025	PFDaA	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.594	2.07	11/11/2025	AEH
MW-EB-3	007	11/3/2025	PFDoS	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.413	2.07	11/11/2025	AEH
MW-EB-3	007	11/3/2025	PFDS	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.426	2.07	11/11/2025	AEH
MW-EB-3	007	11/3/2025	PFEESA	ND	ng/L	UQC+	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.250	4.14	11/11/2025	AEH
MW-EB-3	007	11/3/2025	PFHpA	17.8	ng/L		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.693	2.07	11/11/2025	AEH
MW-EB-3	007	11/3/2025	PFHpS	4.91	ng/L		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.660	2.07	11/11/2025	AEH
MW-EB-3	007	11/3/2025	PFHxA	24.0	ng/L		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.419	2.07	11/11/2025	AEH
MW-EB-3	007	11/3/2025	PFHxS	36.9	ng/L		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	1.26	2.07	11/11/2025	AEH



Alliance Technical Group - Akron
 3310 Win St.
 Cuyahoga Falls, Ohio 44223
 TEL: (330) 253-8211 FAX: (330) 253-4489
 Website: <http://www.settek.com>

WO#: 25110242
 Date Reported:
 Company: ATG - NEWARK LAB
 Address: 284 Sheffield Street
 Mountainside NJ 07092
 Received: 11/5/2025
 Project#: Q3535

Client ID#	Lab ID#	Collected	Analyte	Result Units	Qual	Matrix	Method	DF	MDL	PQL	Run	Analyst
MW-EB-3	007	11/3/2025	PFMBA	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.287	4.14	11/11/2025	AEH
MW-EB-3	007	11/3/2025	PFMPA	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.138	4.14	11/11/2025	AEH
MW-EB-3	007	11/3/2025	PFNA	4.81 ng/L		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.391	2.07	11/11/2025	AEH
MW-EB-3	007	11/3/2025	PFNS	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.841	2.07	11/11/2025	AEH
MW-EB-3	007	11/3/2025	PFOA	130 ng/L		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.553	2.07	11/11/2025	AEH
MW-EB-3	007	11/3/2025	PFOS	172 ng/L		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	1.36	2.07	11/11/2025	AEH
MW-EB-3	007	11/3/2025	PFOSA	2.25 ng/L		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.305	2.07	11/11/2025	AEH
MW-EB-3	007	11/3/2025	PFPeA	17.2 ng/L		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.603	4.14	11/11/2025	AEH
MW-EB-3	007	11/3/2025	PFPeS	10.3 ng/L		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.692	2.07	11/11/2025	AEH
MW-EB-3	007	11/3/2025	PFTeDA	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.448	2.07	11/11/2025	AEH
MW-EB-3	007	11/3/2025	PFTrDA	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.343	2.07	11/11/2025	AEH
MW-EB-3	007	11/3/2025	PFUnA	0.837 ng/L	J	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.740	2.07	11/11/2025	AEH
MW-EB-3	007	11/3/2025	Surr: 13C2-PFDA	97.0 %Rec		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0		11/11/2025	AEH
MW-EB-3	007	11/3/2025	Surr: 13C2-PFHxA	79.2 %Rec		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0		11/11/2025	AEH



Alliance Technical Group - Akron
3310 Win St.
Cuyahoga Falls, Ohio 44223
TEL: (330) 253-8211 FAX: (330) 253-4489
Website: <http://www.settek.com>

WO#: 25110242
Date Reported:
Company: ATG - NEWARK LAB
Address: 284 Sheffield Street
Mountainside NJ 07092
Received: 11/5/2025
Project#: Q3535

Client ID#	Lab ID#	Collected	Analyte	Result	Units	Qual	Matrix	Method	DF	MDL	PQL	Run	Analyst
MW-EB-3	007	11/3/2025	Surr: 13C3-PFBA	59.5	%Rec		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0		11/11/2025	AEH
MW-EB-3	007	11/3/2025	Surr: 13C4-PFOA	59.0	%Rec		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0		11/11/2025	AEH
MW-EB-3	007	11/3/2025	Surr: 13C4-PFOS	105	%Rec		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0		11/11/2025	AEH
MW-EB-3	007	11/3/2025	Surr: 13C5-PFNA	79.3	%Rec		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0		11/11/2025	AEH
MW-EB-3	007	11/3/2025	Surr: 18O2-PFHxS	103	%Rec		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0		11/11/2025	AEH



Alliance Technical Group - Akron
 3310 Win St.
 Cuyahoga Falls, Ohio 44223
 TEL: (330) 253-8211 FAX: (330) 253-4489
 Website: <http://www.settek.com>

WO#: 25110242
 Date Reported:
 Company: ATG - NEWARK LAB
 Address: 284 Sheffield Street
 Mountainside NJ 07092
 Received: 11/5/2025
 Project#: Q3535

Client ID#	Lab ID#	Collected	Analyte	Result	Units	Qual	Matrix	Method	DF	MDL	PQL	Run	Analyst
FB-1	008	11/4/2025	11CI-PF3OUdS	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.217	7.97	11/11/2025	AEH
FB-1	008	11/4/2025	3:3 FTCA	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.458	9.96	11/11/2025	AEH
FB-1	008	11/4/2025	4:2 FTS	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	1.70	7.97	11/11/2025	AEH
FB-1	008	11/4/2025	5:3 FTCA	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	1.72	49.8	11/11/2025	AEH
FB-1	008	11/4/2025	6:2 FTS	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	2.61	7.97	11/11/2025	AEH
FB-1	008	11/4/2025	7:3 FTCA	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	3.04	49.8	11/11/2025	AEH
FB-1	008	11/4/2025	8:2 FTS	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	3.06	7.97	11/11/2025	AEH
FB-1	008	11/4/2025	9CL-PF3ONS	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.299	7.97	11/11/2025	AEH
FB-1	008	11/4/2025	ADONA	ND	ng/L	UQC+	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.340	7.97	11/11/2025	AEH
FB-1	008	11/4/2025	HFPO-DA	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.415	7.97	11/11/2025	AEH
FB-1	008	11/4/2025	NEtFOSA	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.364	1.99	11/11/2025	AEH
FB-1	008	11/4/2025	NEtFOSAA	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	1.01	1.99	11/11/2025	AEH
FB-1	008	11/4/2025	NEtFOSE	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	3.50	19.9	11/11/2025	AEH
FB-1	008	11/4/2025	NFDHA	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.290	3.98	11/11/2025	AEH



Alliance Technical Group - Akron
 3310 Win St.
 Cuyahoga Falls, Ohio 44223
 TEL: (330) 253-8211 FAX: (330) 253-4489
 Website: <http://www.settek.com>

WO#: 25110242
 Date Reported:
 Company: ATG - NEWARK LAB
 Address: 284 Sheffield Street
 Mountainside NJ 07092
 Received: 11/5/2025
 Project#: Q3535

Client ID#	Lab ID#	Collected	Analyte	Result	Units	Qual	Matrix	Method	DF	MDL	PQL	Run	Analyst
FB-1	008	11/4/2025	NMeFOSA	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.610	1.99	11/11/2025	AEH
FB-1	008	11/4/2025	NMeFOSAA	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.888	1.99	11/11/2025	AEH
FB-1	008	11/4/2025	NMeFOSE	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	3.76	19.9	11/11/2025	AEH
FB-1	008	11/4/2025	PFBA	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.818	7.97	11/11/2025	AEH
FB-1	008	11/4/2025	PFBS	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.485	1.99	11/11/2025	AEH
FB-1	008	11/4/2025	PFDA	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.636	1.99	11/11/2025	AEH
FB-1	008	11/4/2025	PFDaA	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.572	1.99	11/11/2025	AEH
FB-1	008	11/4/2025	PFDoS	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.398	1.99	11/11/2025	AEH
FB-1	008	11/4/2025	PFDS	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.410	1.99	11/11/2025	AEH
FB-1	008	11/4/2025	PFEESA	ND	ng/L	UQC+	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.241	3.98	11/11/2025	AEH
FB-1	008	11/4/2025	PFHpA	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.667	1.99	11/11/2025	AEH
FB-1	008	11/4/2025	PFHpS	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.636	1.99	11/11/2025	AEH
FB-1	008	11/4/2025	PFHxA	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.404	1.99	11/11/2025	AEH
FB-1	008	11/4/2025	PFHxS	ND	ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	1.22	1.99	11/11/2025	AEH



Alliance Technical Group - Akron
 3310 Win St.
 Cuyahoga Falls, Ohio 44223
 TEL: (330) 253-8211 FAX: (330) 253-4489
 Website: <http://www.settek.com>

WO#: 25110242
 Date Reported:
 Company: ATG - NEWARK LAB
 Address: 284 Sheffield Street
 Mountainside NJ 07092
 Received: 11/5/2025
 Project#: Q3535

Client ID#	Lab ID#	Collected	Analyte	Result Units	Qual	Matrix	Method	DF	MDL	PQL	Run	Analyst
FB-1	008	11/4/2025	PFMBA	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.276	3.98	11/11/2025	AEH
FB-1	008	11/4/2025	PFMPA	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.132	3.98	11/11/2025	AEH
FB-1	008	11/4/2025	PFNA	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.377	1.99	11/11/2025	AEH
FB-1	008	11/4/2025	PFNS	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.810	1.99	11/11/2025	AEH
FB-1	008	11/4/2025	PFOA	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.533	1.99	11/11/2025	AEH
FB-1	008	11/4/2025	PFOS	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	1.31	1.99	11/11/2025	AEH
FB-1	008	11/4/2025	PFOSA	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.294	1.99	11/11/2025	AEH
FB-1	008	11/4/2025	PFPeA	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.581	3.98	11/11/2025	AEH
FB-1	008	11/4/2025	PFPeS	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.666	1.99	11/11/2025	AEH
FB-1	008	11/4/2025	PFTeDA	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.431	1.99	11/11/2025	AEH
FB-1	008	11/4/2025	PFTTrDA	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.331	1.99	11/11/2025	AEH
FB-1	008	11/4/2025	PFUnA	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.713	1.99	11/11/2025	AEH
FB-1	008	11/4/2025	Surr: 13C2-PFDA	75.2 %Rec		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0		11/11/2025	AEH
FB-1	008	11/4/2025	Surr: 13C2-PFHxA	70.4 %Rec		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0		11/11/2025	AEH



Alliance Technical Group - Akron
3310 Win St.
Cuyahoga Falls, Ohio 44223
TEL: (330) 253-8211 FAX: (330) 253-4489
Website: <http://www.settek.com>

WO#: 25110242
Date Reported:
Company: ATG - NEWARK LAB
Address: 284 Sheffield Street
Mountainside NJ 07092
Received: 11/5/2025
Project#: Q3535

Client ID#	Lab ID#	Collected	Analyte	Result	Units	Qual	Matrix	Method	DF	MDL	PQL	Run	Analyst
FB-1	008	11/4/2025	Surr: 13C3-PFBA	55.8	%Rec		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0		11/11/2025	AEH
FB-1	008	11/4/2025	Surr: 13C4-PFOA	70.3	%Rec		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0		11/11/2025	AEH
FB-1	008	11/4/2025	Surr: 13C4-PFOS	78.9	%Rec		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0		11/11/2025	AEH
FB-1	008	11/4/2025	Surr: 13C5-PFNA	79.0	%Rec		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0		11/11/2025	AEH
FB-1	008	11/4/2025	Surr: 18O2-PFHxS	71.2	%Rec		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0		11/11/2025	AEH



Alliance Technical Group - Akron
3310 Win St.
Cuyahoga Falls, Ohio 44223
TEL: (330) 253-8211 FAX: (330) 253-4489
Website: <http://www.settek.com>

WO#: 25110242
Date Reported:
Company: ATG - NEWARK LAB
Address: 284 Sheffield Street
Mountainside NJ 07092
Received: 11/5/2025
Project#: Q3535

Client ID#	Lab ID#	Collected	Analyte	Result Units	Qual	Matrix	Method	DF	MDL	PQL	Run	Analyst
EB-1	009	11/4/2025	11CI-PF3OUdS	0.728 ng/L	J	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.221	8.09	11/11/2025	AEH
EB-1	009	11/4/2025	3:3 FTCA	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.466	10.1	11/11/2025	AEH
EB-1	009	11/4/2025	4:2 FTS	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	1.73	8.09	11/11/2025	AEH
EB-1	009	11/4/2025	5:3 FTCA	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	1.75	50.6	11/11/2025	AEH
EB-1	009	11/4/2025	6:2 FTS	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	2.65	8.09	11/11/2025	AEH
EB-1	009	11/4/2025	7:3 FTCA	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	3.09	50.6	11/11/2025	AEH
EB-1	009	11/4/2025	8:2 FTS	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	3.11	8.09	11/11/2025	AEH
EB-1	009	11/4/2025	9CL-PF3ONS	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.304	8.09	11/11/2025	AEH
EB-1	009	11/4/2025	ADONA	ND ng/L	UQC+	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.345	8.09	11/11/2025	AEH
EB-1	009	11/4/2025	HFPO-DA	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.422	8.09	11/11/2025	AEH
EB-1	009	11/4/2025	NEtFOSA	0.577 ng/L	J	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.370	2.02	11/11/2025	AEH
EB-1	009	11/4/2025	NEtFOSAA	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	1.03	2.02	11/11/2025	AEH
EB-1	009	11/4/2025	NEtFOSE	5.67 ng/L	J	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	3.56	20.2	11/11/2025	AEH
EB-1	009	11/4/2025	NFDHA	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.295	4.05	11/11/2025	AEH



Alliance Technical Group - Akron
3310 Win St.
Cuyahoga Falls, Ohio 44223
TEL: (330) 253-8211 FAX: (330) 253-4489
Website: <http://www.settek.com>

WO#: 25110242
Date Reported:
Company: ATG - NEWARK LAB
Address: 284 Sheffield Street
Mountainside NJ 07092
Received: 11/5/2025
Project#: Q3535

Client ID#	Lab ID#	Collected	Analyte	Result Units	Qual	Matrix	Method	DF	MDL	PQL	Run	Analyst
EB-1	009	11/4/2025	NMeFOSA	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.619	2.02	11/11/2025	AEH
EB-1	009	11/4/2025	NMeFOSAA	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.902	2.02	11/11/2025	AEH
EB-1	009	11/4/2025	NMeFOSE	4.70 ng/L	J	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	3.82	20.2	11/11/2025	AEH
EB-1	009	11/4/2025	PFBA	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.831	8.09	11/11/2025	AEH
EB-1	009	11/4/2025	PFBS	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.493	2.02	11/11/2025	AEH
EB-1	009	11/4/2025	PFDA	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.646	2.02	11/11/2025	AEH
EB-1	009	11/4/2025	PFDaA	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.581	2.02	11/11/2025	AEH
EB-1	009	11/4/2025	PFDoS	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.404	2.02	11/11/2025	AEH
EB-1	009	11/4/2025	PFDS	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.417	2.02	11/11/2025	AEH
EB-1	009	11/4/2025	PFEESA	ND ng/L	UQC+	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.244	4.05	11/11/2025	AEH
EB-1	009	11/4/2025	PFHpA	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.678	2.02	11/11/2025	AEH
EB-1	009	11/4/2025	PFHpS	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.646	2.02	11/11/2025	AEH
EB-1	009	11/4/2025	PFHxA	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.410	2.02	11/11/2025	AEH
EB-1	009	11/4/2025	PFHxS	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	1.24	2.02	11/11/2025	AEH



Alliance Technical Group - Akron
 3310 Win St.
 Cuyahoga Falls, Ohio 44223
 TEL: (330) 253-8211 FAX: (330) 253-4489
 Website: <http://www.settek.com>

WO#: 25110242
 Date Reported:
 Company: ATG - NEWARK LAB
 Address: 284 Sheffield Street
 Mountainside NJ 07092
 Received: 11/5/2025
 Project#: Q3535

Client ID#	Lab ID#	Collected	Analyte	Result Units	Qual	Matrix	Method	DF	MDL	PQL	Run	Analyst
EB-1	009	11/4/2025	PFMBA	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.281	4.05	11/11/2025	AEH
EB-1	009	11/4/2025	PFMPA	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.135	4.05	11/11/2025	AEH
EB-1	009	11/4/2025	PFNA	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.383	2.02	11/11/2025	AEH
EB-1	009	11/4/2025	PFNS	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.823	2.02	11/11/2025	AEH
EB-1	009	11/4/2025	PFOA	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.541	2.02	11/11/2025	AEH
EB-1	009	11/4/2025	PFOS	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	1.33	2.02	11/11/2025	AEH
EB-1	009	11/4/2025	PFOSA	0.354 ng/L	J	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.299	2.02	11/11/2025	AEH
EB-1	009	11/4/2025	PFPeA	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.590	4.05	11/11/2025	AEH
EB-1	009	11/4/2025	PFPeS	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.677	2.02	11/11/2025	AEH
EB-1	009	11/4/2025	PFTeDA	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.438	2.02	11/11/2025	AEH
EB-1	009	11/4/2025	PFTrDA	0.445 ng/L	J	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.336	2.02	11/11/2025	AEH
EB-1	009	11/4/2025	PFUnA	ND ng/L	U	Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0.724	2.02	11/11/2025	AEH
EB-1	009	11/4/2025	Surr: 13C2-PFDA	97.0 %Rec		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0		11/11/2025	AEH
EB-1	009	11/4/2025	Surr: 13C2-PFHxA	91.9 %Rec		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0		11/11/2025	AEH



Alliance Technical Group - Akron
3310 Win St.
Cuyahoga Falls, Ohio 44223
TEL: (330) 253-8211 FAX: (330) 253-4489
Website: <http://www.settek.com>

WO#: 25110242
Date Reported:
Company: ATG - NEWARK LAB
Address: 284 Sheffield Street
Mountainside NJ 07092
Received: 11/5/2025
Project#: Q3535

Client ID#	Lab ID#	Collected	Analyte	Result	Units	Qual	Matrix	Method	DF	MDL	PQL	Run	Analyst
EB-1	009	11/4/2025	Surr: 13C3-PFBA	72.2	%Rec		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0		11/11/2025	AEH
EB-1	009	11/4/2025	Surr: 13C4-PFOA	92.4	%Rec		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0		11/11/2025	AEH
EB-1	009	11/4/2025	Surr: 13C4-PFOS	91.5	%Rec		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0		11/11/2025	AEH
EB-1	009	11/4/2025	Surr: 13C5-PFNA	99.6	%Rec		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0		11/11/2025	AEH
EB-1	009	11/4/2025	Surr: 18O2-PFHxS	107	%Rec		Non-Potable Water	EPA 1633 Draft - Rev. 4 EPA 1633	1	0		11/11/2025	AEH



Alliance Technical Group - Akron
3310 Win St.
Cuyahoga Falls, Ohio 44223
TEL: (330) 253-8211 FAX: (330) 253-4489
Website: <http://www.settek.com>

QC SUMMARY REPORT

WO#: 25110242

19-Nov-25

Client: ATG - NEWARK LAB

Project: Q3535

BatchID: 89674

Sample ID: MB-89674	SampType: MBLK	TestCode: SVOC-EPA16		Units: ng/L	Prep Date: 11/6/2025				RunNo: 220639		
Client ID: BatchQC	Batch ID: 89674	TestNo: E1633		E1633	Analysis Date: 11/11/2025				SeqNo: 5883613		
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual
11CI-PF3OUdS	ND	8.00									U
3:3 FTCA	ND	10.0									U
4:2 FTS	ND	8.00									U
5:3 FTCA	ND	50.0									U
6:2 FTS	ND	8.00									U
7:3 FTCA	ND	50.0									U
8:2 FTS	ND	8.00									U
9CL-PF3ONS	ND	8.00									U
ADONA	ND	8.00									U
HFPO-DA	ND	8.00									U
NEtFOSA	ND	2.00									U
NEtFOSAA	ND	2.00									U
NEtFOSE	ND	20.0									U
NFDHA	ND	4.00									U
NMeFOSA	ND	2.00									U
NMeFOSAA	ND	2.00									U
NMeFOSE	ND	20.0									U
PFBA	ND	8.00									U
PFBS	ND	2.00									U
PFDA	ND	2.00									U
PFDaA	ND	2.00									U
PFDoS	ND	2.00									U
PFDS	ND	2.00									U

Qualifiers: H Holding times for preparation or analysis exceeded
ND Not Detected
S Spike Recovery outside accepted recovery limits

J Analyte detected below quantitation limits
PL Permit Limit
U Samples with CalcVal < MDL

M Manual Integration used to determine area response
RL Reporting Detection Limit
W Sample container temperature is out of limit as specified

PRELIMINARY

Original



Alliance Technical Group - Akron
3310 Win St.
Cuyahoga Falls, Ohio 44223
TEL: (330) 253-8211 FAX: (330) 253-4489
Website: <http://www.settek.com>

QC SUMMARY REPORT

WO#: 25110242

19-Nov-25

Client: ATG - NEWARK LAB

Project: Q3535

BatchID: 89674

Sample ID: MB-89674	SampType: MBLK	TestCode: SVOC-EPA16 Units: ng/L			Prep Date: 11/6/2025			RunNo: 220639			
Client ID: BatchQC	Batch ID: 89674	TestNo: E1633		E1633	Analysis Date: 11/11/2025			SeqNo: 5883613			
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual
PFEESA	ND	4.00									U
PFHpA	ND	2.00									U
PFHpS	ND	2.00									U
PFHxA	ND	2.00									U
PFHxS	ND	2.00									U
PFMBA	ND	4.00									U
PFMPA	ND	4.00									U
PFNA	ND	2.00									U
PFNS	ND	2.00									U
PFOA	ND	2.00									U
PFOS	ND	2.00									U
PFOSA	ND	2.00									U
PFPeA	ND	4.00									U
PFPeS	ND	2.00									U
PFTeDA	ND	2.00									U
PFTrDA	ND	2.00									U
PFOA	ND	2.00									U
PFUnA	ND	2.00									U
Surr: 13C2-PFDA	0.0193		0.02500		77.2	50	200				
Surr: 13C2-PFHxA	0.0357		0.05000		71.3	50	200				
Surr: 13C3-PFBA	0.0598		0.1000		59.8	50	200				
Surr: 13C4-PFOA	0.0340		0.05000		67.9	50	200				
Surr: 13C4-PFOS	0.0340		0.04790		70.9	50	200				
Surr: 13C5-PFNA	0.0165		0.02500		66.2	50	200				

Qualifiers: H Holding times for preparation or analysis exceeded
ND Not Detected
S Spike Recovery outside accepted recovery limits

J Analyte detected below quantitation limits
PL Permit Limit
U Samples with CalcVal < MDL

M Manual Integration used to determine area response
RL Reporting Detection Limit
W Sample container temperature is out of limit as specified

PRELIMINARY

Original



Alliance Technical Group - Akron
3310 Win St.
Cuyahoga Falls, Ohio 44223
TEL: (330) 253-8211 FAX: (330) 253-4489
Website: <http://www.settek.com>

QC SUMMARY REPORT

WO#: 25110242

19-Nov-25

Client: ATG - NEWARK LAB

Project: Q3535

BatchID: 89674

Sample ID: MB-89674	SampType: MBLK	TestCode: SVOC-EPA16 Units: ng/L			Prep Date: 11/6/2025			RunNo: 220639			
Client ID: BatchQC	Batch ID: 89674	TestNo: E1633		E1633	Analysis Date: 11/11/2025			SeqNo: 5883613			
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual
Surr: 18O2-PFHxS	0.0346		0.04740		73.0	50	200				

Sample ID: LCS-89674	SampType: LCS	TestCode: SVOC-EPA16		Units: ng/L	Prep Date: 11/6/2025			RunNo: 220639			
Client ID: BatchQC	Batch ID: 89674	TestNo: E1633		E1633	Analysis Date: 11/11/2025			SeqNo: 5883614			
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual
11Cl-PF3OUdS	383	8.00	377	0	102	55	160				
3:3 FTCA	450	10.0	500	0	89.9	65	130				
4:2 FTS	331	8.00	375	0	88.4	70	145				
5:3 FTCA	2320	50.0	2500	0	92.8	70	135				
6:2 FTS	322	8.00	380	0	84.8	65	155				
7:3 FTCA	2470	50.0	2500	0	98.7	50	145				
8:2 FTS	354	8.00	384	0	92.2	60	150				
9Cl-PF3ONS	422	8.00	373	0	113	70	155				
ADONA	427	8.00	378	0	113	65	145				
HFPO-DA	383	8.00	400	0	95.7	70	140				
NEtFOSA	100	2.00	100	0	100	65	145				
NEtFOSAA	97.9	2.00	100	0	97.9	70	145				
NEtFOSE	941	20.0	1000	0	94.1	70	135				
NFDHA	196	4.00	200	0	97.8	50	150				
NMeFOSA	91.4	2.00	100	0	91.4	60	150				
NMeFOSAA	86.1	2.00	100	0	86.1	50	140				

Qualifiers: H Holding times for preparation or analysis exceeded
ND Not Detected
S Spike Recovery outside accepted recovery limits

J Analyte detected below quantitation limits
PL Permit Limit
U Samples with CalcVal < MDL

M Manual Integration used to determine area response
RL Reporting Detection Limit
W Sample container temperature is out of limit as spec

PRELIMINARY

Original



Alliance Technical Group - Akron
3310 Win St.
Cuyahoga Falls, Ohio 44223
TEL: (330) 253-8211 FAX: (330) 253-4489
Website: <http://www.settek.com>

QC SUMMARY REPORT

WO#: 25110242

19-Nov-25

Client: ATG - NEWARK LAB

Project: Q3535

BatchID: 89674

Sample ID: LCS-89674		SampType: LCS	TestCode: SVOC-EPA16 Units: ng/L			Prep Date: 11/6/2025			RunNo: 220639		
Client ID: BatchQC	Batch ID: 89674	TestNo: E1633		E1633		Analysis Date: 11/11/2025			SeqNo: 5883614		
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual
NMeFOSE	1030	20.0	1000	0	103	70	145				
PFBA	352	8.00	400	0	88.0	70	140				
PFBS	80.8	2.00	88.8	0	90.9	60	145				
PFDA	87.7	2.00	100	0	87.7	70	140				
PFDaA	91.1	2.00	100	0	91.1	70	140				
PFDoS	83.6	2.00	96.8	0	86.4	50	145				
PFDS	95.5	2.00	96.4	0	99.1	60	145				
PFEESA	174	4.00	178	0	97.4	70	140				
PFHpA	102	2.00	100	0	102	70	150				
PFHpS	101	2.00	95.2	0	106	70	150				
PFHxA	91.9	2.00	100	0	91.9	70	145				
PFHxS	92.2	2.00	91.2	0	101	65	145				
PFMBA	182	4.00	200	0	91.1	60	150				
PFMPA	161	4.00	200	0	80.3	55	140				
PFNA	92.1	2.00	100	0	92.1	70	150				
PFNS	101	2.00	96.0	0	105	56	145				
PFOA	105	2.00	100	0	105	70	150				
PFOS	86.3	2.00	92.8	0	93.0	55	150				
PFOSA	100	2.00	100	0	100	70	145				
PFPeA	203	4.00	200	0	101	65	135				
PFPeS	88.4	2.00	94.0	0	94.1	65	140				
PFTeDA	94.5	2.00	100	0	94.5	60	140				
PFTrDA	107	2.00	100	0	107	65	140				

Qualifiers:

H Holding times for preparation or analysis exceeded
ND Not Detected
S Spike Recovery outside accepted recovery limits

J Analyte detected below quantitation limits
PL Permit Limit
U Samples with CalcVal < MDL

M Manual Integration used to determine area response
RL Reporting Detection Limit
W Sample container temperature is out of limit as spec

PRELIMINARY

Original



Alliance Technical Group - Akron
3310 Win St.
Cuyahoga Falls, Ohio 44223
TEL: (330) 253-8211 FAX: (330) 253-4489
Website: <http://www.settek.com>

QC SUMMARY REPORT

WO#: 25110242

19-Nov-25

Client: ATG - NEWARK LAB

Project: Q3535

BatchID: 89674

Sample ID: LCS-89674	SampType: LCS	TestCode: SVOC-EPA16	Units: ng/L	Prep Date: 11/6/2025	RunNo: 220639						
Client ID: BatchQC	Batch ID: 89674	TestNo: E1633	E1633	Analysis Date: 11/11/2025	SeqNo: 5883614						
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual
PFUnA	92.5	2.00	100	0	92.5	70	145				
Surr: 13C2-PFDA	0.0205		0.02500		82.0	50	200				
Surr: 13C2-PFHxA	0.0410		0.05000		82.0	50	200				
Surr: 13C3-PFBA	0.0651		0.1000		65.1	50	200				
Surr: 13C4-PFOA	0.0417		0.05000		83.3	50	200				
Surr: 13C4-PFOS	0.0422		0.04790		88.2	50	200				
Surr: 13C5-PFNA	0.0219		0.02500		87.5	50	200				
Surr: 18O2-PFHxS	0.0432		0.04740		91.2	50	200				

Qualifiers: H Holding times for preparation or analysis exceeded
ND Not Detected
S Spike Recovery outside accepted recovery limits

J Analyte detected below quantitation limits
PL Permit Limit
U Samples with CalcVal < MDL

M Manual Integration used to determine area response
RL Reporting Detection Limit
W Sample container temperature is out of limit as specified

PRELIMINARY

Original



Alliance Technical Group - Akron
3310 Win St.
Cuyahoga Falls, Ohio 44223
TEL: (330) 253-8211 FAX: (330) 253-4489
Website: <http://www.settek.com>

QC SUMMARY REPORT

WO#: 25110242

19-Nov-25

Client: ATG - NEWARK LAB

Project: Q3535

BatchID: R220410

Sample ID: MB-R220410	SampType: MBLK	TestCode: TSS_NPW(25 Units: mg/L				Prep Date:				RunNo: 220410		
Client ID: BatchQC	Batch ID: R220410	TestNo: A2540D				Analysis Date: 11/6/2025				SeqNo: 5877980		
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual	
TSS	ND	2.50									U	

Sample ID: LCS-R220410	SampType: LCS	TestCode: TSS_NPW(25 Units: mg/L				Prep Date:				RunNo: 220410		
Client ID: BatchQC	Batch ID: R220410	TestNo: A2540D				Analysis Date: 11/6/2025				SeqNo: 5877981		
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual	
TSS	93.5	25.0	100.0	0	93.5	85	115					

Sample ID: 25110284-001ADUP	SampType: DUP	TestCode: TSS_NPW(25	Units: mg/L	Prep Date:	RunNo: 220410						
Client ID: BatchQC	Batch ID: R220410	TestNo: A2540D		Analysis Date: 11/6/2025	SeqNo: 5877983						
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual
TSS	1560	33.3						1513	2.78	5	

Sample ID: 25110284-002ADUP	SampType: DUP	TestCode: TSS_NPW(25 Units: mg/L				Prep Date:				RunNo: 220410		
Client ID: BatchQC	Batch ID: R220410	TestNo: A2540D				Analysis Date: 11/6/2025				SeqNo: 5877985		
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual	
TSS	376	16.7						368.3	2.15	5		

Qualifiers: H Holding times for preparation or analysis exceeded
ND Not Detected
S Spike Recovery outside accepted recovery limits

J Analyte detected below quantitation limits
PL Permit Limit
U Samples with CalcVal < MDL

M Manual Integration used to determine area response
RL Reporting Detection Limit
W Sample container temperature is out of limit as specified

PRELIMINARY

Original



Alliance Technical Group - Akron
3310 Win St.
Cuyahoga Falls, Ohio 44223
TEL: (330) 253-8211 FAX: (330) 253-4489
Website: <http://www.settek.com>

QC SUMMARY REPORT

WO#: 25110242

19-Nov-25

Client: ATG - NEWARK LAB

Project: Q3535

BatchID: R220410

Sample ID: 25110284-002ADUP		SampType: DUP		TestCode: TSS_NPW(25		Units: mg/L		Prep Date:		RunNo: 220410	
Client ID: BatchQC		Batch ID: R220410		TestNo: A2540D				Analysis Date: 11/6/2025		SeqNo: 5877985	
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual

Qualifiers: H Holding times for preparation or analysis exceeded
ND Not Detected
S Spike Recovery outside accepted recovery limits

J Analyte detected below quantitation limits
PL Permit Limit
U Samples with CalcVal < MDL

M Manual Integration used to determine area response
RL Reporting Detection Limit
W Sample container temperature is out of limit as specified

PRELIMINARY

Original



Alliance Technical Group - Akron
3310 Win St.
Cuyahoga Falls, Ohio 44223
TEL: (330) 253-8211 FAX: (330) 253-4489
Website: <http://www.settek.com>

QC SUMMARY REPORT

WO#: 25110242

19-Nov-25

Client: ATG - NEWARK LAB

Project: Q3535

BatchID: R220618

Sample ID: MB-R220618	SampType: MBLK	TestCode: TSS_NPW(25 Units: mg/L				Prep Date:				RunNo: 220618		
Client ID: BatchQC	Batch ID: R220618	TestNo: A2540D				Analysis Date: 11/10/2025				SeqNo: 5883092		
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual	
TSS	ND	2.50									U	

Sample ID: LCS-R220618	SampType: LCS	TestCode: TSS_NPW(25 Units: mg/L				Prep Date:			RunNo: 220618		
Client ID: BatchQC	Batch ID: R220618	TestNo: A2540D				Analysis Date: 11/10/2025			SeqNo: 5883093		
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual
TSS	94.5	25.0	100.0	0	94.5	85	115				

Sample ID: 25110453-003ADUP		SampType: DUP		TestCode: TSS_NPW(25			Units: mg/L		Prep Date:			RunNo: 220618		
Client ID: BatchQC		Batch ID: R220618		TestNo: A2540D			Analysis Date: 11/10/2025			SeqNo: 5883097				
Analyte		Result		PQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual	
TSS		8800		500						8690	1.26	5	W	

Qualifiers: H Holding times for preparation or analysis exceeded
ND Not Detected
S Spike Recovery outside accepted recovery limits

J Analyte detected below quantitation limits
PL Permit Limit
U Samples with CalcVal < MDL

M Manual Integration used to determine area response
RL Reporting Detection Limit
W Sample container temperature is out of limit as specified

PRELIMINARY

Original

CLIENT INFORMATION

REPORT TO BE SENT TO:
COMPANY: Remington Vernick Engineers
ADDRESS: 2059 Springdale Road
CITY: Cherry Hill STATE: NJ ZIP: 08003
ATTENTION: Kyle Carson @ rve.com
PHONE: 609-682-3049 FAX:

CLIENT PROJECT INFORMATION

PROJECT NAME: Edison Landfill
PROJECT NO.: 12051007 LOCATION: Edison NJ
PROJECT MANAGER: Leslie Hartzell
e-mail: Leslie.Hartzell@rve.com
PHONE: 610-663-4165 FAX:

CLIENT BILLING INFORMATION

BILL TO: AP Invoicing One Corp PO#: 12051007
ADDRESS: 2059 Springdale Rd
CITY: Cherry Hill STATE: NJ ZIP: 08003
ATTENTION: Leslie Hartzell PHONE: 610-663-4165

ANALYSIS

DATA TURNAROUND INFORMATION

FAX (RUSH) _____ DAYS*
HARDCOPY (DATA PACKAGE): _____ DAYS*
EDD: _____ DAYS*
*TO BE APPROVED BY CHEMTECH
STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS

DATA DELIVERABLE INFORMATION

☐ Level 1 (Results Only) ☐ Level 4 (QC + Full Raw Data)
☐ Level 2 (Results + QC) ☒ NJ Reduced ☐ US EPA CLP
☐ Level 3 (Results + QC) ☐ NYS ASP A ☐ NYS ASP B
+ Raw Data ☐ Other _____
☒ EDD FORMAT Handsite

1 2 3 4 5 6 7 8 9

ALLIANCE SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE		SAMPLE COLLECTION		# OF BOTTLES	PRESERVATIVES									COMMENTS	
			COMP	GRAB	DATE	TIME		1	2	3	4	5	6	7	8	9	← Specify Preservatives A-HCl D-NaOH B-HNO3 E-ICE C-H2SO4 F-OTHER	
1.	MW-6	GW	Y	Y	11/3/25	10:45	3	3										
2.	MW-1	GW	Y	Y	11/3/25	11:50	3	3										
3.	MW-11	GW	Y	Y	11/3/25	12:00	3	3										
4.	MW-5	GW	Y	Y	11/4/25	10:20	3	3										
5.	MW-EB-1	GW	Y	Y	11/3/25	15:00	3	3										
6.	MW-EB-2	GW	Y	Y	11/4/25	10:05	3	3										
7.	MW-EB-3	GW	Y	Y	11/3/25	14:30	3	3										
8.	FB-1	GW	Y	Y	11/4/25	12:00	3	3										field blank
9.	FB-1	GW	Y	Y	11/4/25	12:00	3	3										equipment blank
10.																		

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER: 1. <u>ARRANDER</u>	DATE/TIME: <u>11/4/25 01:30</u>	RECEIVED BY: <u>[Signature]</u> <u>1218</u> <u>11-4-25</u>	Conditions of bottles or coolers at receipt: ☐ COMPLIANT ☐ NON COMPLIANT ☐ COOLER TEMP <u>3.1</u> °C
RELINQUISHED BY SAMPLER: 2. <u>[Signature]</u>	DATE/TIME:	RECEIVED BY:	Comments:
RELINQUISHED BY SAMPLER: 3. <u>[Signature]</u>	DATE/TIME: <u>11-4-25</u>	RECEIVED BY: 3.	Page <u>1</u> of <u>1</u> CLIENT: ☐ Hand Delivered ☐ Other

Shipment Complete
☐ YES ☐ NO



284 Sheffield Street, Mountainside, NJ 07092
(908) 789-8900 Fax (908) 789-8922
WWW.CHEMTECH.NET

25110242

CHAIN OF CUSTODY RECORD

Sub Lab INFORMATION	CLIENT PROJECT INFORMATION	CLIENT BILLING INFORMATION
COMPANY : Alliance Technical Group - Akron	ORDER ID : Q3535	BILL TO: CHEMTECH PO# : Q3535
ADDRESS : 3310 Win Street	PROJECT ID:Edison Landfill	ADDRESS : 284, Sheffield Street
CITY:Cuyahoga Fal State :OH ZIP :44223	PROJECT MANAGER YAZMEEN	CITY: Mountainside State : NJ ZIP : 07092
E-mail :jennifer.woolf@alliancetg.com	E-mail : YAZMEEN.GOMEZ@AllianceTG.com	ATTENTION :YAZMEE
PHONE :330-253-8211	PHONE : (908) 789 8900 FAX: (908) 789 8922	PHONE : (908) 789 8900 FAX : (908) 789 8922

EDD : Excel NJ Report : NJ Reduced Comment : NJ. GRAB.

ID	CLIENT SAMPLE IDENTIFICATION	SAMPLE MATRIX	ANALYSIS	Preservative	Method	SAMPLE COLLECTION		# OF BOTTLES	TAT DAYS
						DATE	TIME		
01	MW-6	Water	PFAS	Cool 4 deg C	1633	11/03/2025	10:45:00	3	10
02	MW-1	Water	PFAS	Cool 4 deg C	1633	11/03/2025	11:50:00	3	10
03	MW-11	Water	PFAS	Cool 4 deg C	1633	11/03/2025	12:00:00	3	10
04	MW-5	Water	PFAS	Cool 4 deg C	1633	11/04/2025	10:20:00	3	10
05	MW-EB-1	Water	PFAS	Cool 4 deg C	1633	11/03/2025	15:00:00	3	10
06	MW-EB-2	Water	PFAS	Cool 4 deg C	1633	11/04/2025	10:05:00	3	10
07	MW-EB-3	Water	PFAS	Cool 4 deg C	1633	11/03/2025	14:30:00	3	10
08	FB-1	Water	PFAS	Cool 4 deg C	1633	11/04/2025	12:00:00	3	10
09	EB-1	Water	PFAS	Cool 4 deg C	1633	11/04/2025	12:00:00	3	10

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGES POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER:	DATETIME: 11/4/25 15:30	RECEIVED BY:	Conditions of bottles or Coolers at receipt:	☐ Compliant ☐ Non Compliant	Cooler Temp _____ Ice or Cooler? _____
1. [Signature]		1. [Signature]			
RELINQUISHED BY:	DATETIME: 11/5/25	RECEIVED BY:	2. [Signature] 10:30am 2.8+0.2 = 3.0 FedEx		
2. [Signature]		2. [Signature]			
RELINQUISHED BY:	DATETIME:	RECEIVED BY:		☐ OVERNIGHT ☐ OVERNIGHT	Shipment Complete: ☐ YES ☐ NO
3. [Signature]		3. [Signature]			

Sample Log-In Check List

Client Name: **CHE-NJ-07972**


Work Order Number: **25110242**

RcptNo: **1**

Logged by: **Nick Rigby** 11/5/2025 10:30:00 AM

Completed By: **Maggie M. Grmek** 11/5/2025 12:58:36 PM

Reviewed By: **Jennifer Woolf** 11/5/2025 1:16:38 PM



Chain of Custody

1. Is Chain of Custody complete? Yes ☒ No ☐ Not Present ☐

2. How was the sample delivered? FedEx
Tracking No.: 885632192110

Log In

3. Coolers are present? Yes ☒ No ☐ NA ☐

4. Shipping container/cooler in good condition? Yes ☒ No ☐
Custody seals intact on shipping container/cooler? Yes ☐ No ☐ Not Present ☒ NA ☐

No. Seal Date: Signed By:
5. Was an attempt made to cool the samples? Yes ☒ No ☐ NA ☐

6. Were all samples received at a temperature of >0° C to 6.0° C Yes ☒ No ☐ NA ☐

7. Sample(s) in proper container(s)? Yes ☒ No ☐

8. Sufficient sample volume for indicated test(s)? Yes ☒ No ☐

9. Are samples (except VOA and ONG) properly preserved? Yes ☒ No ☐

10. Was preservative added to bottles? Yes ☐ No ☒ NA ☐

11. Is the headspace in the VOA vials less than 1/4 inch or 6 mm? Yes ☐ No ☐ No VOA Vials ☒

12. Were any sample containers received broken? Yes ☐ No ☒

13. Does paperwork match bottle labels?
(Note discrepancies on chain of custody) Yes ☒ No ☐

14. Are matrices correctly identified on Chain of Custody? Yes ☒ No ☐

15. Is it clear what analyses were requested? Yes ☒ No ☐

16. Were all holding times able to be met?
(If no, notify customer for authorization.) Yes ☒ No ☐

Special Handling (if applicable)

17. Was client notified of all discrepancies with this order? Yes ☐ No ☐ NA ☒

Person Notified: Date:
By Whom: Via: ☐ eMail ☐ Phone ☐ Fax ☐ In Person
Regarding:
Client Instructions:

18. Additional remarks:

Cooler Information

Cooler No	Temp °C	Condition	Seal Intact	Seal No	Seal Date	Signed By
1	3	Good	Not Present			