



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: Ambric Technology Corporation

SDG No.: Q3538

Project: NEWPCP Sludge Screen 13923-GE-25

RunNo.: LB137781

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: Corrosivity	ICV	pH	7.02	7	100	90-110	11/05/2025
Sample ID: Corrosivity	CCV1	pH	2.02	2.00	101	90-110	11/05/2025
Sample ID: Corrosivity	CCV2	pH	12.02	12.00	100	90-110	11/05/2025