

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: Alliance

Contract: PSEG03

Lab Code: ACE

SDG No.: Q3541

Instrument ID: BNA_P

Calibration Date(s): 10/29/2025 10/29/2025

Calibration Time(s): 09:26 14:13

LAB FILE ID:		RRF2.5 = BP026028.D			RRF005 = BP026029.D		RRF010 = BP026030.D	
		RRF020 = BP026031.D			RRF040 = BP026032.D		RRF050 = BP026033.D	
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
2-Fluorophenol		1.055	1.255	1.234	1.249	1.123	1.212	7.4
Benzaldehyde			0.855	0.938	0.809	0.670	0.802	11.5
Phenol-d6		1.366	1.580	1.574	1.606	1.425	1.543	6.8
Phenol		1.542	1.771	1.750	1.773	1.568	1.713	6.4
bis(2-Chloroethyl)ether		1.271	1.434	1.372	1.373	1.228	1.352	5.6
2-Chlorophenol		1.116	1.354	1.356	1.403	1.262	1.342	9.0
2-Methylphenol		0.958	1.129	1.145	1.175	1.057	1.121	7.8
2,2-oxybis(1-Chloropropane)		1.978	2.210	2.102	2.099	1.846	2.043	5.8
Acetophenone		0.481	0.548	0.519	0.518	0.455	0.506	6.1
3+4-Methylphenols			1.521	1.557	1.613	1.430	1.560	5.0
n-Nitroso-di-n-propylamine	0.795	0.894	1.013	1.016	1.027	0.904	0.958	8.8
Nitrobenzene-d5		0.256	0.317	0.327	0.338	0.306	0.321	10.3
Hexachloroethane		0.481	0.562	0.535	0.552	0.490	0.533	6.5
Nitrobenzene		0.282	0.345	0.346	0.357	0.317	0.339	8.8
Isophorone		0.606	0.721	0.705	0.720	0.632	0.687	7.0
2-Nitrophenol		0.069	0.092	0.107	0.128	0.124	0.104	23.2
2,4-Dimethylphenol		0.243	0.306	0.281	0.302	0.269	0.289	8.9
bis(2-Chloroethoxy)methane		0.406	0.466	0.436	0.447	0.396	0.433	5.8
2,4-Dichlorophenol		0.241	0.302	0.308	0.327	0.292	0.305	10.8
Naphthalene		1.039	1.181	1.121	1.115	0.978	1.091	6.3
4-Chloroaniline		0.374	0.440	0.427	0.437	0.384	0.419	6.8
Hexachlorobutadiene		0.198	0.231	0.209	0.216	0.190	0.211	6.7
Caprolactam			0.099	0.103	0.113	0.099	0.107	7.0
4-Chloro-3-methylphenol		0.257	0.314	0.325	0.345	0.303	0.320	10.3
2-Methylnaphthalene		0.717	0.823	0.771	0.795	0.694	0.763	6.0
Hexachlorocyclopentadiene			0.200	0.198	0.237	0.218	0.227	11.5
2,4,6-Trichlorophenol		0.256	0.342	0.372	0.402	0.360	0.367	15.7
2-Fluorobiphenyl		1.350	1.549	1.454	1.421	1.231	1.392	7.5
2,4,5-Trichlorophenol		0.309	0.414	0.425	0.456	0.405	0.419	12.9
1,1-Biphenyl		1.454	1.673	1.577	1.560	1.378	1.531	6.4
2-Chloronaphthalene		1.130	1.304	1.248	1.229	1.090	1.205	6.2
2-Nitroaniline		0.166	0.226	0.269	0.304	0.277	0.272	21.9
Dimethylphthalate		1.282	1.574	1.491	1.522	1.323	1.455	7.8
Acenaphthylene		1.547	1.845	1.832	1.832	1.618	1.756	7.2

All other compounds must meet a minimum RRF of 0.010.

Form VI SV-1

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: Alliance

Contract: PSEG03

Lab Code: ACE

SDG No.: Q3541

Instrument ID: BNA_P

Calibration Date(s): 10/29/2025 10/29/2025

Calibration Time(s): 09:26 14:13

LAB FILE ID:		RRF2.5 = BP026028.D			RRF005 = BP026029.D		RRF010 = BP026030.D	
		RRF020 = BP026031.D			RRF040 = BP026032.D		RRF050 = BP026033.D	
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
2,6-Dinitrotoluene		0.142	0.196	0.234	0.264	0.245	0.237	22.6
3-Nitroaniline		0.177	0.254	0.296	0.324	0.295	0.291	20.7
Acenaphthene		1.101	1.285	1.247	1.249	1.089	1.206	6.8
2,4-Dinitrophenol			0.069	0.082	0.100	0.096	0.100	22.4
4-Nitrophenol			0.277	0.309	0.285	0.258	0.290	7.0
Dibenzofuran		1.725	1.997	1.886	1.869	1.614	1.822	6.9
2,4-Dinitrotoluene		0.195	0.288	0.340	0.393	0.360	0.347	24.0
Diethylphthalate		1.240	1.570	1.484	1.549	1.350	1.463	8.7
4-Chlorophenyl-phenylether		0.699	0.803	0.736	0.731	0.623	0.713	8.3
Fluorene		1.347	1.621	1.501	1.450	1.260	1.427	8.8
4-Nitroaniline		0.207	0.294	0.328	0.342	0.298	0.311	16.9
4,6-Dinitro-2-methylphenol			0.053	0.065	0.076	0.075	0.077	22.1
n-Nitrosodiphenylamine		0.574	0.678	0.641	0.647	0.567	0.625	6.8
2,4,6-Tribromophenol		0.155	0.211	0.218	0.236	0.209	0.216	14.2
4-Bromophenyl-phenylether		0.206	0.234	0.222	0.228	0.204	0.222	5.8
Hexachlorobenzene		0.239	0.273	0.256	0.258	0.226	0.253	6.5
Atrazine		0.177	0.213	0.221	0.219	0.193	0.211	9.6
Pentachlorophenol			0.126	0.143	0.161	0.148	0.154	12.4
Phenanthrene		1.133	1.288	1.209	1.189	1.031	1.164	7.1
Anthracene		1.064	1.262	1.195	1.177	1.044	1.154	7.1
Carbazole		0.985	1.172	1.164	1.129	0.969	1.092	7.7
Di-n-butylphthalate		0.937	1.219	1.241	1.333	1.196	1.222	11.6
Fluoranthene		1.259	1.469	1.426	1.411	1.224	1.358	7.1
Pyrene		1.201	1.488	1.395	1.427	1.217	1.353	8.2
Terphenyl-d14		0.953	1.125	1.013	1.014	0.860	0.984	8.8
Butylbenzylphthalate		0.235	0.337	0.407	0.496	0.468	0.432	26.4
3,3-Dichlorobenzidine			0.390	0.424	0.462	0.416	0.443	8.9
Benzo (a) anthracene		1.279	1.497	1.407	1.420	1.224	1.374	7.0
Chrysene		1.210	1.424	1.348	1.345	1.169	1.302	6.9
Bis (2-ethylhexyl) phthalate		0.453	0.613	0.690	0.829	0.768	0.718	20.3
Di-n-octyl phthalate			0.822	1.025	1.287	1.238	1.191	18.9
Benzo (b) fluoranthene		1.074	1.277	1.278	1.306	1.149	1.242	7.7
Benzo (k) fluoranthene		1.119	1.365	1.306	1.331	1.141	1.268	7.9
Benzo (a) pyrene		0.941	1.169	1.171	1.210	1.054	1.141	9.4
Indeno (1,2,3-cd) pyrene		1.190	1.486	1.539	1.564	1.371	1.476	10.1

All other compounds must meet a minimum RRF of 0.010.

Form VI SV-1

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: Alliance

Contract: PSEG03

Lab Code: ACE

SDG No.: Q3541

Instrument ID: BNA_P

Calibration Date(s): 10/29/2025 10/29/2025

Calibration Time(s): 09:26 14:13

LAB FILE ID:		RRF2.5 = BP026028.D			RRF005 = BP026029.D		RRF010 = BP026030.D	
		RRF020 = BP026031.D			RRF040 = BP026032.D		RRF050 = BP026033.D	
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
Dibenzo(a,h)anthracene		0.978	1.219	1.249	1.274	1.117	1.202	9.7
Benzo(g,h,i)perylene		0.966	1.167	1.219	1.212	1.068	1.156	8.9
1,2,4,5-Tetrachlorobenzene		0.574	0.670	0.637	0.632	0.565	0.621	6.3
1,4-Dioxane		0.472	0.567	0.519	0.515	0.453	0.511	7.6
2,3,4,6-Tetrachlorophenol		0.223	0.303	0.332	0.364	0.328	0.330	16.7