

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3541

Client ID: SSTDCCC040

Date Analyzed: 11/06/2025

Lab File ID: BF144180.D

Time Analyzed: 11:13

Instrument ID: BNA_F

GC Column: DB-UI ID: 0.18 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	301670	11.404	184652	14.057	216041	15.539
UPPER LIMIT	603340	11.904	369304	14.557	432082	16.039
LOWER LIMIT	150835	10.904	92326	13.557	108021	15.039
EPA SAMPLE NO.						
01 N48969	224075	11.40	167742	14.05	139517	15.53

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.