

### CALIBRATION VERIFICATION SUMMARY

**Lab Name:** Alliance **Contract:** CLOU03  
**Lab Code:** ACE **SDG NO.:** Q3545  
**GC Column:** ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

**Client Sample No.:** CCAL01 **Date Analyzed:** 11/07/2025  
**Lab Sample No.:** AR1660CCC500 **Data File :** PP076373.D **Time Analyzed:** 09:42

| COMPOUND             | RT     | RT WINDOW |        | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|--------|-----------|--------|--------------------|-------------------|------|
|                      |        | FROM      | TO     |                    |                   |      |
| Aroclor-1016-1       | 5.787  | 5.695     | 5.895  | 576.250            | 500.000           | 15.3 |
| Aroclor-1016-2       | 5.809  | 5.717     | 5.917  | 503.130            | 500.000           | 0.6  |
| Aroclor-1016-3       | 5.872  | 5.780     | 5.980  | 506.540            | 500.000           | 1.3  |
| Aroclor-1016-4       | 5.969  | 5.877     | 6.077  | 520.000            | 500.000           | 4.0  |
| Aroclor-1016-5       | 6.261  | 6.169     | 6.369  | 537.980            | 500.000           | 7.6  |
| Aroclor-1260-1       | 7.380  | 7.287     | 7.487  | 570.690            | 500.000           | 14.1 |
| Aroclor-1260-2       | 7.633  | 7.540     | 7.740  | 544.650            | 500.000           | 8.9  |
| Aroclor-1260-3       | 7.991  | 7.899     | 8.099  | 559.970            | 500.000           | 12.0 |
| Aroclor-1260-4       | 8.219  | 8.127     | 8.327  | 560.670            | 500.000           | 12.1 |
| Aroclor-1260-5       | 8.544  | 8.452     | 8.652  | 530.860            | 500.000           | 6.2  |
| Decachlorobiphenyl   | 10.407 | 10.319    | 10.519 | 49.920             | 50.000            | -0.2 |
| Tetrachloro-m-xylene | 4.637  | 4.542     | 4.742  | 53.560             | 50.000            | 7.1  |

### CALIBRATION VERIFICATION SUMMARY

**Lab Name:** Alliance **Contract:** CLOU03  
**Lab Code:** ACE **SDG NO.:** Q3545  
**GC Column:** ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

**Client Sample No.:** CCAL01 **Date Analyzed:** 11/07/2025  
**Lab Sample No.:** AR1660CCC500 **Data File :** PP076373.D **Time Analyzed:** 09:42

| COMPOUND             | RT    | RT WINDOW |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|-------|-----------|-------|--------------------|-------------------|------|
|                      |       | FROM      | TO    |                    |                   |      |
| Aroclor-1016-1       | 4.870 | 4.778     | 4.978 | 524.510            | 500.000           | 4.9  |
| Aroclor-1016-2       | 4.930 | 4.836     | 5.036 | 483.130            | 500.000           | -3.4 |
| Aroclor-1016-3       | 5.048 | 4.954     | 5.154 | 482.550            | 500.000           | -3.5 |
| Aroclor-1016-4       | 5.090 | 4.996     | 5.196 | 545.090            | 500.000           | 9.0  |
| Aroclor-1016-5       | 5.303 | 5.210     | 5.410 | 504.070            | 500.000           | 0.8  |
| Aroclor-1260-1       | 6.330 | 6.238     | 6.438 | 577.760            | 500.000           | 15.6 |
| Aroclor-1260-2       | 6.517 | 6.425     | 6.625 | 566.710            | 500.000           | 13.3 |
| Aroclor-1260-3       | 6.670 | 6.578     | 6.778 | 575.490            | 500.000           | 15.1 |
| Aroclor-1260-4       | 7.138 | 7.047     | 7.247 | 584.730            | 500.000           | 16.9 |
| Aroclor-1260-5       | 7.380 | 7.287     | 7.487 | 568.140            | 500.000           | 13.6 |
| Decachlorobiphenyl   | 8.772 | 8.682     | 8.882 | 47.390             | 50.000            | -5.2 |
| Tetrachloro-m-xylene | 3.777 | 3.681     | 3.881 | 56.240             | 50.000            | 12.5 |