



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q3545

Order ID: Q3545

Client: CHA Companies, Inc.

Project ID: NYSOMH Emergency Service

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID :				Total Concentration:	0.000			



SAMPLE DATA

A

B

C

D

Report of Analysis

Client: CHA Companies, Inc.
Project: NYSOMH Emergency Service
Client Sample ID: PCB-110325KC-01 Hard Doorframe Caulk
Lab Sample ID: Q3545-01
Analytical Method: 8082A
Sample Wt/Vol: 1.31 g Final Vol: 10000 uL
Prep Method: SW3541B Prep Date: 11/07/25

Date Collected: 11/03/25
Date Received: 11/05/25
SDG No.: Q3545
Matrix: CAULK
% Solid: 100
Test: PCB

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	905	U	10	905	3900	ug/kg	11/07/25 12:33	PB170432
11104-28-2	Aroclor-1221	923	U	10	923	3900	ug/kg	11/07/25 12:33	PB170432
11141-16-5	Aroclor-1232	852	U	10	852	3900	ug/kg	11/07/25 12:33	PB170432
53469-21-9	Aroclor-1242	918	U	10	918	3900	ug/kg	11/07/25 12:33	PB170432
12672-29-6	Aroclor-1248	1400	U	10	1400	3900	ug/kg	11/07/25 12:33	PB170432
11097-69-1	Aroclor-1254	735	U	10	735	3900	ug/kg	11/07/25 12:33	PB170432
37324-23-5	Aroclor-1262	1100	U	10	1100	3900	ug/kg	11/07/25 12:33	PB170432
11100-14-4	Aroclor-1268	824	U	10	824	3900	ug/kg	11/07/25 12:33	PB170432
11096-82-5	Aroclor-1260	740	U	10	740	3900	ug/kg	11/07/25 12:33	PB170432
SURROGATES									
877-09-8	Tetrachloro-m-xylene	18.7			32 - 144	94%	SPK: 20		
2051-24-3	Decachlorobiphenyl	51.0	*		32 - 175	255%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

LAB CHRONICLE

OrderID:	Q3545	OrderDate:	11/5/2025 11:53:00 AM
Client:	CHA Companies, Inc.	Project:	NYSOMH Emergency Service
Contact:	Carrie Robinson	Location:	A11

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q3545-01	PCB-110325KC-01 Hard Doorframe Caulk	CAULK	PCB	8082A	11/03/25	11/07/25	11/07/25	11/05/25