

Surrogate Summary

SDG No.: **Q3545**

Client: **CHA Companies, Inc.**

Analytical Method: **8082A**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PP075887.D	PIBLK-PP075887.D	Tetrachloro-m-xyl	1	20	19.9	100		60	140
		Decachlorobiphen	1	20	20.4	102		60	140
		Tetrachloro-m-xyl	2	20	18.8	94		60	140
		Decachlorobiphen	2	20	19.3	96		60	140
I.BLK-PP076377.D	PIBLK-PP076377.D	Tetrachloro-m-xyl	1	20	24.1	121		60	140
		Decachlorobiphen	1	20	23.4	117		60	140
		Tetrachloro-m-xyl	2	20	21.1	105		60	140
		Decachlorobiphen	2	20	20.4	102		60	140
PB170432BL	PB170432BL	Tetrachloro-m-xyl	1	20	26.5	133		32	144
		Decachlorobiphen	1	20	26.2	131		32	175
		Tetrachloro-m-xyl	2	20	25.3	126		32	144
		Decachlorobiphen	2	20	23.4	117		32	175
PB170432BS	PB170432BS	Tetrachloro-m-xyl	1	20	28.3	142		32	144
		Decachlorobiphen	1	20	26.4	132		32	175
		Tetrachloro-m-xyl	2	20	26.4	132		32	144
		Decachlorobiphen	2	20	23.4	117		32	175
Q3545-01	PCB-110325KC-01 Hard Doorframe	Tetrachloro-m-xyl	1	20	18.7	94		32	144
		Decachlorobiphen	1	20	51.0	255	*	32	175
		Tetrachloro-m-xyl	2	20	11.3	57		32	144
		Decachlorobiphen	2	20	31.4	157		32	175
I.BLK-PP076391.D	PIBLK-PP076391.D	Tetrachloro-m-xyl	1	20	22.9	115		60	140
		Decachlorobiphen	1	20	23.8	119		60	140
		Tetrachloro-m-xyl	2	20	21.5	108		60	140
		Decachlorobiphen	2	20	21.0	105		60	140