

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms

SDG No.: Q3547

Contract: ARAM01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV103	Mercury	3.99	4.0	100	95 - 105	CV	11/07/2025	14:28	LB137819

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms

SDG No.: Q3547

Contract: ARAM01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV40	Mercury	4.99	5.0	100	90 - 110	CV	11/07/2025	14:33	LB137819
CCV41	Mercury	4.94	5.0	99	90 - 110	CV	11/07/2025	15:06	LB137819
CCV42	Mercury	4.90	5.0	98	90 - 110	CV	11/07/2025	15:15	LB137819

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms

SDG No.: Q3547

Contract: ARAM01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Cadmium	1960	2000	98	95 - 105	P	11/10/2025	11:45	LB137841
	Copper	1040	1000	104	95 - 105	P	11/10/2025	11:45	LB137841
	Lead	4190	4000	105	95 - 105	P	11/10/2025	11:45	LB137841
	Nickel	2010	2000	100	95 - 105	P	11/10/2025	11:45	LB137841
	Zinc	2020	2000	101	95 - 105	P	11/10/2025	11:45	LB137841

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms

SDG No.: Q3547

Contract: ARAM01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Cadmium	5.56	6.0	93	80 - 120	P	11/10/2025	11:57	LB137841
	Copper	20.3	20.0	101	80 - 120	P	11/10/2025	11:57	LB137841
	Lead	10.8	12.0	90	80 - 120	P	11/10/2025	11:57	LB137841
	Nickel	38.3	40.0	96	80 - 120	P	11/10/2025	11:57	LB137841
	Zinc	40.9	40.0	102	80 - 120	P	11/10/2025	11:57	LB137841

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms

SDG No.: Q3547

Contract: ARAM01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Cadmium	2460	2500	98	90 - 110	P	11/10/2025	12:38	LB137841
	Copper	1240	1250	99	90 - 110	P	11/10/2025	12:38	LB137841
	Lead	4910	5000	98	90 - 110	P	11/10/2025	12:38	LB137841
	Nickel	2460	2500	98	90 - 110	P	11/10/2025	12:38	LB137841
	Zinc	2490	2500	99	90 - 110	P	11/10/2025	12:38	LB137841
CCV02	Cadmium	2480	2500	99	90 - 110	P	11/10/2025	13:28	LB137841
	Copper	1250	1250	100	90 - 110	P	11/10/2025	13:28	LB137841
	Lead	4950	5000	99	90 - 110	P	11/10/2025	13:28	LB137841
	Nickel	2470	2500	99	90 - 110	P	11/10/2025	13:28	LB137841
	Zinc	2500	2500	100	90 - 110	P	11/10/2025	13:28	LB137841
CCV03	Cadmium	2440	2500	98	90 - 110	P	11/10/2025	14:46	LB137841
	Copper	1240	1250	99	90 - 110	P	11/10/2025	14:46	LB137841
	Lead	4860	5000	97	90 - 110	P	11/10/2025	14:46	LB137841
	Nickel	2440	2500	98	90 - 110	P	11/10/2025	14:46	LB137841
	Zinc	2490	2500	100	90 - 110	P	11/10/2025	14:46	LB137841
CCV04	Cadmium	2470	2500	99	90 - 110	P	11/10/2025	15:44	LB137841
	Copper	1260	1250	101	90 - 110	P	11/10/2025	15:44	LB137841
	Lead	4920	5000	98	90 - 110	P	11/10/2025	15:44	LB137841
	Nickel	2470	2500	99	90 - 110	P	11/10/2025	15:44	LB137841
	Zinc	2530	2500	101	90 - 110	P	11/10/2025	15:44	LB137841
CCV05	Cadmium	2430	2500	97	90 - 110	P	11/10/2025	16:35	LB137841
	Copper	1250	1250	100	90 - 110	P	11/10/2025	16:35	LB137841
	Lead	4810	5000	96	90 - 110	P	11/10/2025	16:35	LB137841
	Nickel	2420	2500	97	90 - 110	P	11/10/2025	16:35	LB137841
	Zinc	2500	2500	100	90 - 110	P	11/10/2025	16:35	LB137841
CCV06	Cadmium	2470	2500	99	90 - 110	P	11/10/2025	17:27	LB137841
	Copper	1270	1250	102	90 - 110	P	11/10/2025	17:27	LB137841
	Lead	4880	5000	98	90 - 110	P	11/10/2025	17:27	LB137841
	Nickel	2460	2500	99	90 - 110	P	11/10/2025	17:27	LB137841
	Zinc	2590	2500	103	90 - 110	P	11/10/2025	17:27	LB137841
CCV07	Cadmium	2590	2500	104	90 - 110	P	11/10/2025	19:14	LB137841
	Copper	1320	1250	105	90 - 110	P	11/10/2025	19:14	LB137841
	Lead	5130	5000	103	90 - 110	P	11/10/2025	19:14	LB137841
	Nickel	2570	2500	103	90 - 110	P	11/10/2025	19:14	LB137841

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms

SDG No.: Q3547

Contract: ARAM01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV07	Zinc	2680	2500	107	90 - 110	P	11/10/2025	19:14	LB137841
CCV08	Cadmium	2600	2500	104	90 - 110	P	11/10/2025	20:11	LB137841
	Copper	1320	1250	106	90 - 110	P	11/10/2025	20:11	LB137841
	Lead	5140	5000	103	90 - 110	P	11/10/2025	20:11	LB137841
	Nickel	2580	2500	103	90 - 110	P	11/10/2025	20:11	LB137841
	Zinc	2680	2500	107	90 - 110	P	11/10/2025	20:11	LB137841
CCV09	Cadmium	2490	2500	100	90 - 110	P	11/10/2025	21:31	LB137841
	Copper	1280	1250	103	90 - 110	P	11/10/2025	21:31	LB137841
	Lead	4890	5000	98	90 - 110	P	11/10/2025	21:31	LB137841
	Nickel	2480	2500	99	90 - 110	P	11/10/2025	21:31	LB137841
	Zinc	2630	2500	105	90 - 110	P	11/10/2025	21:31	LB137841
CCV10	Cadmium	2450	2500	98	90 - 110	P	11/10/2025	22:29	LB137841
	Copper	1260	1250	101	90 - 110	P	11/10/2025	22:29	LB137841
	Lead	4820	5000	96	90 - 110	P	11/10/2025	22:29	LB137841
	Nickel	2440	2500	98	90 - 110	P	11/10/2025	22:29	LB137841
	Zinc	2580	2500	103	90 - 110	P	11/10/2025	22:29	LB137841
CCV11	Cadmium	2480	2500	99	90 - 110	P	11/10/2025	23:30	LB137841
	Copper	1270	1250	102	90 - 110	P	11/10/2025	23:30	LB137841
	Lead	4890	5000	98	90 - 110	P	11/10/2025	23:30	LB137841
	Nickel	2470	2500	99	90 - 110	P	11/10/2025	23:30	LB137841
	Zinc	2590	2500	103	90 - 110	P	11/10/2025	23:30	LB137841