

metals
- 5a -
MATRIX SPIKE SUMMARY

client: <u>Aramark Uniforms</u>	level: <u>low</u>	sdg no.: <u>Q3547</u>
contract: <u>ARAM01</u>		lab code: <u>ACE</u>
matrix: <u>Water</u>	sample id: <u>Q3547-02</u>	client id: <u>CompMS</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q3547-02MS</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Cadmium	ug/L	75 - 125	99.1		1.62	J	100	97		P
Copper	ug/L	75 - 125	264		121		150	95		P
Lead	ug/L	75 - 125	487		20.0		500	93		P
Nickel	ug/L	75 - 125	269		20.5		250	99		P
Zinc	ug/L	75 - 125	628		524		100	104		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: <u>Aramark Uniforms</u>	level: <u>low</u>	sdg no.: <u>Q3547</u>
contract: <u>ARAM01</u>		lab code: <u>ACE</u>
matrix: <u>Water</u>	sample id: <u>Q3547-02</u>	client id: <u>CompMSD</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q3547-02MSD</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Cadmium	ug/L	75 - 125	104		1.62	J	100	102		P
Copper	ug/L	75 - 125	272		121		150	101		P
Lead	ug/L	75 - 125	513		20.0		500	99		P
Nickel	ug/L	75 - 125	281		20.5		250	104		P
Zinc	ug/L	75 - 125	632		524		100	108		P

metals
- 5a -
MATRIX SPIKE SUMMARY

client: Aramark Uniforms **level:** low **sdg no.:** Q3547
contract: ARAM01 **lab code:** ACE
matrix: Water **sample id:** Q3551-01 **client id:** MANHOLEMS
Percent Solids for Sample: NA **Spiked ID:** Q3551-01MS **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Mercury	ug/L	75 - 125	3.56		0.20	U	4.0	89		CV

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: <u>Aramark Uniforms</u>	level: <u>low</u>	sdg no.: <u>Q3547</u>
contract: <u>ARAM01</u>		lab code: <u>ACE</u>
matrix: <u>Water</u>	sample id: <u>Q3551-01</u>	client id: <u>MANHOLEMSD</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q3551-01MSD</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Mercury	ug/L	75 - 125	4.15		0.20	U	4.0	104		CV