

Laboratory Control Sample Summary

Client: Major Products

SDG No.: Q3550

Project: Floor Drain Water 2025

Run No.: LB137795

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137795BS							
BOD5	mg/L	198	191		97	1	84.6-115.4	11/05/2025

Laboratory Control Sample Summary

Client: Major Products

SDG No.: Q3550

Project: Floor Drain Water 2025

Run No.: LB137812

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137812BS							
TSS	mg/L	550	531		96	1	90-110	11/07/2025