

Initial and Continuing Calibration Blank Summary

Client: Spectra East Inc.

SDG No.: Q3551

Project: Outfall 001 - Orangetown Dis Permit 2025

RunNo.: LB137801

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.030	0.1	11/06/2025
Sample ID: CCB1 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.030	0.1	11/06/2025
Sample ID: CCB2 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.030	0.1	11/06/2025
Sample ID: CCB3 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.030	0.1	11/06/2025

Initial and Continuing Calibration Blank Summary

Client: Spectra East Inc.

SDG No.: Q3551

Project: Outfall 001 - Orangetown Dis Permit 2025

RunNo.: LB137803

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1							
Bromide	mg/L	< 1.0000	1.0000	U	0.37	2	11/03/2025
Chloride	mg/L	< 0.3000	0.3000	U	0.19	0.6	11/03/2025
Fluoride	mg/L	< 0.2000	0.2000	U	0.11	0.4	11/03/2025
Nitrite	mg/L	< 0.3000	0.3000	U	0.074	0.6	11/03/2025
Nitrate	mg/L	< 0.2500	0.2500	U	0.095	0.5	11/03/2025
Sulfate	mg/L	< 1.5000	1.5000	U	0.46	3	11/03/2025
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.34	1	11/03/2025
Sample ID: CCB1							
Bromide	mg/L	< 1.0000	1.0000	U	0.37	2	11/06/2025
Chloride	mg/L	< 0.3000	0.3000	U	0.19	0.6	11/06/2025
Fluoride	mg/L	< 0.2000	0.2000	U	0.11	0.4	11/06/2025
Nitrite	mg/L	< 0.3000	0.3000	U	0.074	0.6	11/06/2025
Nitrate	mg/L	< 0.2500	0.2500	U	0.095	0.5	11/06/2025
Sulfate	mg/L	< 1.5000	1.5000	U	0.46	3	11/06/2025
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.34	1	11/06/2025
Sample ID: CCB2							
Bromide	mg/L	< 1.0000	1.0000	U	0.37	2	11/06/2025
Chloride	mg/L	< 0.3000	0.3000	U	0.19	0.6	11/06/2025
Fluoride	mg/L	< 0.2000	0.2000	U	0.11	0.4	11/06/2025
Nitrite	mg/L	< 0.3000	0.3000	U	0.074	0.6	11/06/2025
Nitrate	mg/L	< 0.2500	0.2500	U	0.095	0.5	11/06/2025
Sulfate	mg/L	< 1.5000	1.5000	U	0.46	3	11/06/2025
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.34	1	11/06/2025
Sample ID: CCB3							
Bromide	mg/L	< 1.0000	1.0000	U	0.37	2	11/06/2025
Chloride	mg/L	< 0.3000	0.3000	U	0.19	0.6	11/06/2025
Fluoride	mg/L	< 0.2000	0.2000	U	0.11	0.4	11/06/2025
Nitrite	mg/L	< 0.3000	0.3000	U	0.074	0.6	11/06/2025
Nitrate	mg/L	< 0.2500	0.2500	U	0.095	0.5	11/06/2025
Sulfate	mg/L	< 1.5000	1.5000	U	0.46	3	11/06/2025
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.34	1	11/06/2025
Sample ID: CCB4							
Bromide	mg/L	< 1.0000	1.0000	U	0.37	2	11/07/2025
Chloride	mg/L	< 0.3000	0.3000	U	0.19	0.6	11/07/2025
Fluoride	mg/L	< 0.2000	0.2000	U	0.11	0.4	11/07/2025
Nitrite	mg/L	< 0.3000	0.3000	U	0.074	0.6	11/07/2025
Nitrate	mg/L	< 0.2500	0.2500	U	0.095	0.5	11/07/2025
Sulfate	mg/L	< 1.5000	1.5000	U	0.46	3	11/07/2025
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.34	1	11/07/2025
Sample ID: CCB5							
Bromide	mg/L	< 1.0000	1.0000	U	0.37	2	11/07/2025
Chloride	mg/L	< 0.3000	0.3000	U	0.19	0.6	11/07/2025
Fluoride	mg/L	< 0.2000	0.2000	U	0.11	0.4	11/07/2025
Nitrite	mg/L	< 0.3000	0.3000	U	0.074	0.6	11/07/2025
Nitrate	mg/L	< 0.2500	0.2500	U	0.095	0.5	11/07/2025
Sulfate	mg/L	< 1.5000	1.5000	U	0.46	3	11/07/2025

Initial and Continuing Calibration Blank Summary

Client: Spectra East Inc.

SDG No.: Q3551

Project: Outfall 001 - Orangetown Dis Permit 2025

RunNo.: LB137803

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.34	1	11/07/2025
Sample ID: CCB6							
Bromide	mg/L	< 1.0000	1.0000	U	0.37	2	11/07/2025
Chloride	mg/L	< 0.3000	0.3000	U	0.19	0.6	11/07/2025
Fluoride	mg/L	< 0.2000	0.2000	U	0.11	0.4	11/07/2025
Nitrite	mg/L	< 0.3000	0.3000	U	0.074	0.6	11/07/2025
Nitrate	mg/L	< 0.2500	0.2500	U	0.095	0.5	11/07/2025
Sulfate	mg/L	< 1.5000	1.5000	U	0.46	3	11/07/2025
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.34	1	11/07/2025

Initial and Continuing Calibration Blank Summary

Client: Spectra East Inc.

SDG No.: Q3551

Project: Outfall 001 - Orangetown Dis Permit 2025

RunNo.: LB137822

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Cyanide	mg/L	< 0.0025	0.0025	U	0.0012	0.005	11/07/2025
Sample ID: CCB1 Cyanide	mg/L	< 0.0025	0.0025	U	0.0012	0.005	11/07/2025
Sample ID: CCB2 Cyanide	mg/L	< 0.0025	0.0025	U	0.0012	0.005	11/07/2025

Initial and Continuing Calibration Blank Summary

Client: Spectra East Inc.

SDG No.: Q3551

Project: Outfall 001 - Orangetown Dis Permit 2025

RunNo.: LB137904

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB COD	mg/L	< 5.0000	5.0000	U	1.50	10	09/08/2025
Sample ID: CCB1 COD	mg/L	< 5.0000	5.0000	U	1.50	10	11/14/2025
Sample ID: CCB2 COD	mg/L	< 5.0000	5.0000	U	1.50	10	11/14/2025
Sample ID: CCB3 COD	mg/L	< 5.0000	5.0000	U	1.50	10	11/14/2025
Sample ID: CCB4 COD	mg/L	< 5.0000	5.0000	U	1.50	10	11/14/2025

Initial and Continuing Calibration Blank Summary

Client: Spectra East Inc.

SDG No.: Q3551

Project: Outfall 001 - Orangetown Dis Permit 2025

RunNo.: LB137926

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Phenolics	mg/L	< 0.0250	0.0250	U	0.015	0.05	11/17/2025
Sample ID: CCB1 Phenolics	mg/L	< 0.0250	0.0250	U	0.015	0.05	11/17/2025
Sample ID: CCB2 Phenolics	mg/L	< 0.0250	0.0250	U	0.015	0.05	11/17/2025