



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Spectra East Inc.

SDG No.: Q3551

Contract: SPEC01

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.043	0.05	86	60 - 140	CV	11/07/2025	14:43	LB137819
CRI01	Arsenic	19.4	20.0	97	65 - 135	P	11/10/2025	12:16	LB137841
	Beryllium	6.53	6.0	109	65 - 135	P	11/10/2025	12:16	LB137841
	Cadmium	5.98	6.0	100	65 - 135	P	11/10/2025	12:16	LB137841
	Chromium	9.05	10.0	90	65 - 135	P	11/10/2025	12:16	LB137841
	Copper	21.9	20.0	110	65 - 135	P	11/10/2025	12:16	LB137841
	Lead	11.1	12.0	93	65 - 135	P	11/10/2025	12:16	LB137841
	Nickel	39.7	40.0	99	65 - 135	P	11/10/2025	12:16	LB137841
	Selenium	22.6	20.0	113	65 - 135	P	11/10/2025	12:16	LB137841
	Silver	10.1	10.0	101	65 - 135	P	11/10/2025	12:16	LB137841
	Zinc	43.2	40.0	108	65 - 135	P	11/10/2025	12:16	LB137841