

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Spectra East Inc.

SDG No.: Q3551

Instrument: CV1

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB170437BL		WATER		Batch Number:	PB170437		Prep Date:	11/07/2025	
	Mercury	0.027	<0.2	U	0.20	CV	11/07/2025	14:50	LB137819

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Spectra East Inc.

SDG No.: Q3551

Instrument: P4

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB170412BL	WATER			Batch Number:	PB170412		Prep Date:	11/06/2025	
	Arsenic	1.85	<5	U	10.0	P	11/10/2025	13:58	LB137841
	Beryllium	0.27	<1.5	U	3.00	P	11/10/2025	13:58	LB137841
	Cadmium	0.31	<1.5	U	3.00	P	11/10/2025	13:58	LB137841
	Chromium	0.55	<2.5	J	5.00	P	11/10/2025	13:58	LB137841
	Copper	1.89	<5	U	10.0	P	11/10/2025	13:58	LB137841
	Lead	1.21	<3	U	6.00	P	11/10/2025	13:58	LB137841
	Nickel	1.90	<10	U	20.0	P	11/10/2025	13:58	LB137841
	Selenium	2.52	<5	U	10.0	P	11/10/2025	13:58	LB137841
	Silver	0.81	<2.5	U	5.00	P	11/10/2025	13:58	LB137841
	Zinc	2.00	<10	U	20.0	P	11/10/2025	13:58	LB137841