

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Spectra East Inc.

SDG No.: Q3551

Contract: SPEC01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB170412BS							
Arsenic	ug/L	400	391		98	85 - 115	P
Beryllium	ug/L	100	104		104	85 - 115	P
Cadmium	ug/L	100	94.7		95	85 - 115	P
Chromium	ug/L	200	203		102	85 - 115	P
Copper	ug/L	150	154		103	85 - 115	P
Lead	ug/L	500	468		94	85 - 115	P
Nickel	ug/L	250	244		98	85 - 115	P
Selenium	ug/L	1000	998		100	85 - 115	P
Silver	ug/L	37.5	37.0		99	85 - 115	P
Zinc	ug/L	100	103		103	85 - 115	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Spectra East Inc.

SDG No.: Q3551

Contract: SPEC01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB170437BS Mercury	ug/L	4.0	4.14		104	85 - 115	CV