

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: SPEC01

Lab Code: ACE

SDG NO.: Q3551

Continuing Calib Date: 11/11/2025

Initial Calibration Date(s): 11/04/2025

11/04/2025

Continuing Calib Time: 22:14

Initial Calibration Time(s): 09:39

18:31

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|----------------------|------------|-----------|-----------|-------|------------|
|                      |            |           | FROM      | TO    |            |
| Aroclor-1242-1 (1)   | 5.49       | 5.49      | 5.39      | 5.59  | 0.00       |
| Aroclor-1242-2 (2)   | 5.51       | 5.51      | 5.41      | 5.61  | 0.00       |
| Aroclor-1242-3 (3)   | 5.57       | 5.57      | 5.47      | 5.67  | 0.00       |
| Aroclor-1242-4 (4)   | 5.67       | 5.67      | 5.57      | 5.77  | 0.01       |
| Aroclor-1242-5 (5)   | 6.39       | 6.39      | 6.29      | 6.49  | 0.00       |
| Tetrachloro-m-xylene | 4.35       | 4.35      | 4.25      | 4.45  | 0.00       |
| Decachlorobiphenyl   | 9.93       | 9.93      | 9.83      | 10.03 | 0.00       |

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: SPEC01

Lab Code: ACE

SDG NO.: Q3551

Continuing Calib Date: 11/11/2025

Initial Calibration Date(s): 11/04/2025

11/04/2025

Continuing Calib Time: 22:14

Initial Calibration Time(s): 09:39

18:31

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|------------|-----------|-----------|------|------------|
|                      |            |           | FROM      | TO   |            |
| Aroclor-1242-1 (1)   | 4.73       | 4.73      | 4.63      | 4.83 | 0.00       |
| Aroclor-1242-2 (2)   | 4.75       | 4.75      | 4.65      | 4.85 | 0.00       |
| Aroclor-1242-3 (3)   | 4.92       | 4.92      | 4.82      | 5.02 | 0.00       |
| Aroclor-1242-4 (4)   | 5.01       | 5.01      | 4.91      | 5.11 | 0.00       |
| Aroclor-1242-5 (5)   | 5.52       | 5.53      | 5.43      | 5.63 | 0.01       |
| Tetrachloro-m-xylene | 3.65       | 3.65      | 3.55      | 3.75 | 0.00       |
| Decachlorobiphenyl   | 8.64       | 8.64      | 8.54      | 8.74 | 0.00       |