

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** SPEC01
Lab Code: ACE **SDG NO.:** Q3551
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL01 **Date Analyzed:** 11/11/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO115010.D **Time Analyzed:** 15:49

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.486	5.387	5.587	516.170	500.000	3.2
Aroclor-1016-2	5.508	5.408	5.608	515.680	500.000	3.1
Aroclor-1016-3	5.569	5.470	5.670	531.180	500.000	6.2
Aroclor-1016-4	5.666	5.566	5.766	520.070	500.000	4.0
Aroclor-1016-5	5.957	5.858	6.058	505.450	500.000	1.1
Aroclor-1260-1	7.071	6.971	7.171	495.620	500.000	-0.9
Aroclor-1260-2	7.325	7.225	7.425	531.970	500.000	6.4
Aroclor-1260-3	7.683	7.582	7.782	531.940	500.000	6.4
Aroclor-1260-4	7.904	7.803	8.003	516.150	500.000	3.2
Aroclor-1260-5	8.212	8.110	8.310	512.360	500.000	2.5
Decachlorobiphenyl	9.933	9.830	10.030	51.230	50.000	2.5
Tetrachloro-m-xylene	4.347	4.248	4.448	53.580	50.000	7.2

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** SPEC01
Lab Code: ACE **SDG NO.:** Q3551
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL01 **Date Analyzed:** 11/11/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO115010.D **Time Analyzed:** 15:49

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.727	4.628	4.828	517.240	500.000	3.4
Aroclor-1016-2	4.746	4.647	4.847	515.690	500.000	3.1
Aroclor-1016-3	4.920	4.821	5.021	522.320	500.000	4.5
Aroclor-1016-4	4.963	4.864	5.064	520.080	500.000	4.0
Aroclor-1016-5	5.175	5.076	5.276	517.410	500.000	3.5
Aroclor-1260-1	6.205	6.106	6.306	495.010	500.000	-1.0
Aroclor-1260-2	6.393	6.294	6.494	487.870	500.000	-2.4
Aroclor-1260-3	6.545	6.446	6.646	496.040	500.000	-0.8
Aroclor-1260-4	7.016	6.917	7.117	503.530	500.000	0.7
Aroclor-1260-5	7.257	7.158	7.358	499.860	500.000	0.0
Decachlorobiphenyl	8.642	8.541	8.741	53.840	50.000	7.7
Tetrachloro-m-xylene	3.648	3.548	3.748	52.450	50.000	4.9