

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** SPEC01
Lab Code: ACE **SDG NO.:** Q3551
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL02 **Date Analyzed:** 11/11/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO115025.D **Time Analyzed:** 21:19

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.487	5.387	5.587	519.320	500.000	3.9
Aroclor-1016-2	5.508	5.408	5.608	514.000	500.000	2.8
Aroclor-1016-3	5.569	5.470	5.670	533.200	500.000	6.6
Aroclor-1016-4	5.666	5.566	5.766	520.550	500.000	4.1
Aroclor-1016-5	5.958	5.858	6.058	507.710	500.000	1.5
Aroclor-1260-1	7.071	6.971	7.171	502.610	500.000	0.5
Aroclor-1260-2	7.324	7.225	7.425	537.250	500.000	7.5
Aroclor-1260-3	7.682	7.582	7.782	531.510	500.000	6.3
Aroclor-1260-4	7.904	7.803	8.003	517.550	500.000	3.5
Aroclor-1260-5	8.211	8.110	8.310	516.110	500.000	3.2
Decachlorobiphenyl	9.931	9.830	10.030	51.650	50.000	3.3
Tetrachloro-m-xylene	4.347	4.248	4.448	52.990	50.000	6.0

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** SPEC01
Lab Code: ACE **SDG NO.:** Q3551
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL02 **Date Analyzed:** 11/11/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO115025.D **Time Analyzed:** 21:19

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.727	4.628	4.828	505.020	500.000	1.0
Aroclor-1016-2	4.746	4.647	4.847	508.360	500.000	1.7
Aroclor-1016-3	4.920	4.821	5.021	517.080	500.000	3.4
Aroclor-1016-4	4.963	4.864	5.064	509.100	500.000	1.8
Aroclor-1016-5	5.175	5.076	5.276	523.950	500.000	4.8
Aroclor-1260-1	6.205	6.106	6.306	485.510	500.000	-2.9
Aroclor-1260-2	6.393	6.294	6.494	485.930	500.000	-2.8
Aroclor-1260-3	6.545	6.446	6.646	492.220	500.000	-1.6
Aroclor-1260-4	7.016	6.917	7.117	496.900	500.000	-0.6
Aroclor-1260-5	7.257	7.158	7.358	495.370	500.000	-0.9
Decachlorobiphenyl	8.641	8.541	8.741	53.840	50.000	7.7
Tetrachloro-m-xylene	3.648	3.548	3.748	51.310	50.000	2.6