

Surrogate Summary

SDG No.: Q3551

Client: Spectra East Inc.

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PO114870.D	PIBLK-PO114870.D	Tetrachloro-m-xyl	1	20	15.5	77		60	140
		Decachlorobiphen	1	20	16.8	84		60	140
		Tetrachloro-m-xyl	2	20	16.2	81		60	140
		Decachlorobiphen	2	20	17.3	86		60	140
I.BLK-PO115014.D	PIBLK-PO115014.D	Tetrachloro-m-xyl	1	20	19.8	99		60	140
		Decachlorobiphen	1	20	19.8	99		60	140
		Tetrachloro-m-xyl	2	20	19.9	100		60	140
		Decachlorobiphen	2	20	21.7	108		60	140
PB170484BL	PB170484BL	Tetrachloro-m-xyl	1	20	23.3	117		30	173
		Decachlorobiphen	1	20	23.3	117		10	173
		Tetrachloro-m-xyl	2	20	23.5	118		30	173
		Decachlorobiphen	2	20	25.5	128		10	173
PB170484BS	PB170484BS	Tetrachloro-m-xyl	1	20	24.5	123		30	173
		Decachlorobiphen	1	20	24.5	122		10	173
		Tetrachloro-m-xyl	2	20	22.5	113		30	173
		Decachlorobiphen	2	20	26.8	134		10	173
PB170484BSD	PB170484BSD	Tetrachloro-m-xyl	1	20	24.3	121		30	173
		Decachlorobiphen	1	20	24.4	122		10	173
		Tetrachloro-m-xyl	2	20	22.8	114		30	173
		Decachlorobiphen	2	20	26.1	130		10	173
Q3551-01	MANHOLE	Tetrachloro-m-xyl	1	20	16.2	81		30	173
		Decachlorobiphen	1	20	8.56	43		10	173
		Tetrachloro-m-xyl	2	20	16.1	81		30	173
		Decachlorobiphen	2	20	10.5	52		10	173
I.BLK-PO115029.D	PIBLK-PO115029.D	Tetrachloro-m-xyl	1	20	19.5	97		60	140
		Decachlorobiphen	1	20	19.6	98		60	140
		Tetrachloro-m-xyl	2	20	19.5	97		60	140
		Decachlorobiphen	2	20	20.7	103		60	140