



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3551</u>	Analytical Method:	<u>8082A</u>
Client:	<u>Spectra East Inc.</u>	Datafile :	<u>PO115016.D</u>

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD	Low	Limits	RPD
								Qual		High	
PB170484BS (Column 1)	AR1016	5	4.10	ug/L	82				77	107	
	AR1260	5	4.00	ug/L	80				66	113	
PB170484BS (Column 2)	AR1016	5	4.00	ug/L	80				77	107	
	AR1260	5	3.90	ug/L	78				66	113	