



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q3551

Analytical Method: 8082A

Client: Spectra East Inc.

Datafile : PO115017.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB170484BSD (Column 1)	AR1016	5	4.20	ug/L	84	2			77	107	20
	AR1260	5	4.00	ug/L	80	0			66	113	20
PB170484BSD (Column 2)	AR1016	5	4.10	ug/L	82	2			77	107	20
	AR1260	5	3.90	ug/L	78	0			66	113	20