

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** SPEC01
Lab Code: ACE **SDG NO.:** Q3551
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/17/2025 10/17/2025

Client Sample No.: CCAL03 **Date Analyzed:** 11/10/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD091036.D **Time Analyzed:** 11:56

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.704	6.600	6.800	49.290	50.000	-1.4
4,4'-DDE	6.194	6.090	6.290	46.340	50.000	-7.3
4,4'-DDT	7.020	6.915	7.115	49.250	50.000	-1.5
Aldrin	5.269	5.164	5.364	47.220	50.000	-5.6
alpha-BHC	3.999	3.894	4.094	48.530	50.000	-2.9
alpha-Chlordane	6.025	5.921	6.121	45.310	50.000	-9.4
beta-BHC	4.517	4.412	4.612	46.600	50.000	-6.8
Decachlorobiphenyl	9.072	8.964	9.164	44.030	50.000	-11.9
delta-BHC	4.765	4.660	4.860	47.410	50.000	-5.2
Dieldrin	6.345	6.240	6.440	47.330	50.000	-5.3
Endosulfan I	6.073	5.968	6.168	46.860	50.000	-6.3
Endosulfan II	6.786	6.681	6.881	46.040	50.000	-7.9
Endosulfan sulfate	7.149	7.045	7.245	46.490	50.000	-7.0
Endrin	6.573	6.468	6.668	45.320	50.000	-9.4
Endrin aldehyde	6.915	6.810	7.010	47.110	50.000	-5.8
Endrin ketone	7.631	7.525	7.725	48.410	50.000	-3.2
gamma-BHC (Lindane)	4.330	4.225	4.425	47.760	50.000	-4.5
gamma-Chlordane	5.944	5.840	6.040	46.830	50.000	-6.3
Heptachlor	4.927	4.823	5.023	57.550	50.000	15.1
Heptachlor epoxide	5.689	5.584	5.784	46.690	50.000	-6.6
Methoxychlor	7.493	7.388	7.588	48.330	50.000	-3.3
Tetrachloro-m-xylene	3.549	3.445	3.645	46.040	50.000	-7.9

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** SPEC01
Lab Code: ACE **SDG NO.:** Q3551
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/17/2025 10/17/2025

Client Sample No.: CCAL03 **Date Analyzed:** 11/10/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD091036.D **Time Analyzed:** 11:56

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.921	5.821	6.021	52.410	50.000	4.8
4,4'-DDE	5.365	5.266	5.466	50.660	50.000	1.3
4,4'-DDT	6.174	6.074	6.274	53.560	50.000	7.1
Aldrin	4.360	4.261	4.461	50.810	50.000	1.6
alpha-BHC	3.386	3.287	3.487	51.080	50.000	2.2
alpha-Chlordane	5.181	5.081	5.281	50.500	50.000	1.0
beta-BHC	4.019	3.919	4.119	50.120	50.000	0.2
Decachlorobiphenyl	8.062	7.962	8.162	53.960	50.000	7.9
delta-BHC	4.255	4.155	4.355	50.690	50.000	1.4
Dieldrin	5.504	5.404	5.604	51.360	50.000	2.7
Endosulfan I	5.238	5.138	5.338	50.630	50.000	1.3
Endosulfan II	6.072	5.972	6.172	51.880	50.000	3.8
Endosulfan sulfate	6.474	6.374	6.574	51.930	50.000	3.9
Endrin	5.781	5.681	5.881	49.740	50.000	-0.5
Endrin aldehyde	6.250	6.150	6.350	52.650	50.000	5.3
Endrin ketone	6.984	6.883	7.083	55.010	50.000	10.0
gamma-BHC (Lindane)	3.723	3.623	3.823	50.800	50.000	1.6
gamma-Chlordane	5.116	5.017	5.217	50.680	50.000	1.4
Heptachlor	4.074	3.975	4.175	52.380	50.000	4.8
Heptachlor epoxide	4.864	4.764	4.964	49.850	50.000	-0.3
Methoxychlor	6.744	6.645	6.845	52.100	50.000	4.2
Tetrachloro-m-xylene	2.874	2.775	2.975	50.470	50.000	0.9