

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q3551

Analytical Method: 608.3 Pest

Client: Spectra East Inc.

Datafile : PD091029.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB170447BSD (Column 1)	alpha-BHC	0.0125	0.0089	ug/L	71	1			37	140	20
	gamma-BHC (Lindane)	0.0125	0.0091	ug/L	73	3			32	140	20
	Heptachlor	0.0125	0.012	ug/L	95	1			34	140	20
	Aldrin	0.0125	0.0096	ug/L	77	3			42	140	20
	beta-BHC	0.0125	0.011	ug/L	86	1			17	147	20
	delta-BHC	0.0125	0.0084	ug/L	67	2			19	140	20
	Heptachlor epoxide	0.0125	0.010	ug/L	80	3			37	142	20
	Endosulfan I	0.0125	0.010	ug/L	81	3			45	153	20
	gamma-Chlordane	0.0125	0.0098	ug/L	78	3			45	140	20
	alpha-Chlordane	0.0125	0.0099	ug/L	79	0			45	140	20
	4,4'-DDE	0.0125	0.0094	ug/L	75	3			30	145	20
	Dieldrin	0.0125	0.0098	ug/L	78	3			36	146	20
	Endrin	0.0125	0.0094	ug/L	75	3			30	147	20
	Endosulfan II	0.0125	0.0099	ug/L	79	5			1	202	20
	4,4'-DDD	0.0125	0.010	ug/L	80	3			31	141	20
	4,4'-DDT	0.0125	0.010	ug/L	80	3			25	160	20
	Endrin aldehyde	0.0125	0.0099	ug/L	79	1			38	141	20
	Endosulfan sulfate	0.0125	0.010	ug/L	80	3			26	144	20
	Methoxychlor	0.0125	0.011	ug/L	89	3			30	151	20
	Endrin ketone	0.0125	0.011	ug/L	84	4			44	149	20
PB170447BSD (Column 2)	alpha-BHC	0.0125	0.011	ug/L	89	2			37	140	20
	gamma-BHC (Lindane)	0.0125	0.011	ug/L	90	3			32	140	20
	Heptachlor	0.0125	0.012	ug/L	96	2			34	140	20
	Aldrin	0.0125	0.011	ug/L	91	2			42	140	20
	beta-BHC	0.0125	0.012	ug/L	94	2			17	147	20
	delta-BHC	0.0125	0.010	ug/L	82	1			19	140	20
	Heptachlor epoxide	0.0125	0.011	ug/L	91	1			37	142	20
	Endosulfan I	0.0125	0.012	ug/L	94	4			45	153	20
	gamma-Chlordane	0.0125	0.011	ug/L	90	2			45	140	20
	alpha-Chlordane	0.0125	0.012	ug/L	92	3			45	140	20
	4,4'-DDE	0.0125	0.012	ug/L	92	2			30	145	20
	Dieldrin	0.0125	0.012	ug/L	94	3			36	146	20
	Endrin	0.0125	0.011	ug/L	90	3			30	147	20
	Endosulfan II	0.0125	0.012	ug/L	96	3			1	202	20
	4,4'-DDD	0.0125	0.012	ug/L	95	3			31	141	20
	4,4'-DDT	0.0125	0.012	ug/L	93	6			25	160	20
	Endrin aldehyde	0.0125	0.011	ug/L	91	3			38	141	20
	Endosulfan sulfate	0.0125	0.012	ug/L	93	3			26	144	20
	Methoxychlor	0.0125	0.012	ug/L	96	3			30	151	20
	Endrin ketone	0.0125	0.013	ug/L	102	4			44	149	20