

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: SPEC01
 Lab Code: ACE SDG No.: Q3551
 Instrument ID: BNA_G Calibration Date/Time: 11/07/2025 11:06
 Lab File ID: BG064713.D Init. Calib. Date(s): 10/24/2025 10/24/2025
 EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 09:00 15:32
 GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.192	1.167		-2.1	
Phenol-d6	1.635	1.530		-6.4	
Nitrobenzene-d5	0.333	0.407		22.2	
2-Fluorobiphenyl	1.319	1.406		6.6	
2,4,6-Tribromophenol	0.289	0.381		31.8	
Terphenyl-d14	1.060	1.075		1.4	
Bis(2-ethylhexyl)phthalate	0.590	0.615		4.2	

All other compounds must meet a minimum RRF of 0.010.