

## Report of Analysis

|                    |                                          |                 |               |
|--------------------|------------------------------------------|-----------------|---------------|
| Client:            | Spectra East Inc.                        | Date Collected: | 11/05/25      |
| Project:           | Outfall 001 - Orangetown Dis Permit 2025 | Date Received:  | 11/05/25      |
| Client Sample ID:  | MANHOLE                                  | SDG No.:        | Q3551         |
| Lab Sample ID:     | Q3551-01                                 | Matrix:         | Water         |
| Analytical Method: | 625.1                                    | % Solid:        | 0             |
| Sample Wt/Vol:     | 960 mL                                   | Test:           | SVOCMS Group1 |
| Prep Method :      | 3510C                                    |                 |               |
|                    | Level: LOW                               |                 |               |
|                    | Final Vol: 1000 uL                       |                 |               |
|                    | Prep Date: 11/07/25                      |                 |               |

| CAS Number                | Parameter                  | Conc.             | Qua. | DF | MDL      | LOQ / CRQL | Units    | Date Ana.      | Prep BatchID |
|---------------------------|----------------------------|-------------------|------|----|----------|------------|----------|----------------|--------------|
| <b>TARGETS</b>            |                            |                   |      |    |          |            |          |                |              |
| 117-81-7                  | Bis(2-ethylhexyl)phthalate | 1.70              | U    | 1  | 1.70     | 5.20       | ug/L     | 11/07/25 18:15 | PB170446     |
| <b>SURROGATES</b>         |                            |                   |      |    |          |            |          |                |              |
| 4165-60-0                 | Nitrobenzene-d5            | 86.5              |      |    | 60 - 140 | 87%        | SPK: 100 |                |              |
| 321-60-8                  | 2-Fluorobiphenyl           | 75.5              |      |    | 60 - 140 | 75%        | SPK: 100 |                |              |
| 1718-51-0                 | Terphenyl-d14              | 68.7              |      |    | 60 - 140 | 69%        | SPK: 100 |                |              |
| <b>INTERNAL STANDARDS</b> |                            |                   |      |    |          |            |          |                |              |
|                           |                            | <b>Area Count</b> |      |    |          |            |          |                |              |
| 3855-82-1                 | 1,4-Dichlorobenzene-d4     | 32100             |      |    |          |            |          |                |              |
| 1146-65-2                 | Naphthalene-d8             | 130000            |      |    |          |            |          |                |              |
| 15067-26-2                | Acenaphthene-d10           | 106000            |      |    |          |            |          |                |              |
| 1517-22-2                 | Phenanthrene-d10           | 265000            |      |    |          |            |          |                |              |
| 1719-03-5                 | Chrysene-d12               | 335000            |      |    |          |            |          |                |              |
| 1520-96-3                 | Perylene-d12               | 384000            |      |    |          |            |          |                |              |

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products