

Instrument ID: BNA_G

Daily Analysis Runlog For Sequence/QC Batch ID # BG102425

Review By	rahul	Review On	10/27/2025 10:16:00 AM
Supervise By	mohammad	Supervise On	11/5/2025 12:11:16 AM
SubDirectory	BG102425	HP Acquire Method	BNA_G
		HP Processing Method	BG102425
STD. NAME	STD REF.#		
Tune/Reschk	SP6875		
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864		
CCC	SP6860		
Internal Standard/PEM	S13179,10ul/1000ul sample		
ICV/I.BLK	SP6872		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BG064568.D	24 Oct 2025 8:20		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BG064569.D	24 Oct 2025 9:00		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BG064570.D	24 Oct 2025 9:41		RC/JU	Ok,M
4	SSTDICC010	SSTDICC010	BG064571.D	24 Oct 2025 11:02		RC/JU	Ok,M
5	SSTDICC020	SSTDICC020	BG064572.D	24 Oct 2025 12:23	Compound #32 kept on QR	RC/JU	Ok,M
6	SSTDICCC040	SSTDICCC040	BG064573.D	24 Oct 2025 13:03	Compound #26,48,51,54,57,65 are kept on LR	RC/JU	Ok,M
7	SSTDICC060	SSTDICC060	BG064574.D	24 Oct 2025 13:44		RC/JU	Ok,M
8	SSTDICC080	SSTDICC080	BG064575.D	24 Oct 2025 14:24		RC/JU	Ok,M
9	SSTDICC050	SSTDICC050	BG064576.D	24 Oct 2025 15:32		RC/JU	Ok,M
10	SSTDICV040	ICVBG102425	BG064577.D	24 Oct 2025 16:47	compound #77 failed in the ICV.	RC/JU	Ok
11	PB170222BL	PB170222BL	BG064578.D	24 Oct 2025 17:28	Analyzed for contamination check	RC/JU	Not Ok
12	Q3405-02	PAD-10-20-25-B	BG064579.D	24 Oct 2025 18:08		RC/JU	Ok
13	Q3405-03	PAD-10-20-25-C	BG064580.D	24 Oct 2025 18:49		RC/JU	Ok
14	Q3405-04	SW-10-20-25	BG064581.D	24 Oct 2025 19:30		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_G

Daily Analysis Runlog For Sequence/QC Batch ID # BG110725

Review By	rahul	Review On	11/7/2025 2:26:44 PM
Supervise By	Jagrut	Supervise On	11/10/2025 12:08:51 PM
SubDirectory	BG110725	HP Acquire Method	BNA_G
		HP Processing Method	BG102425
STD. NAME	STD REF.#		
Tune/Reschk	SP6875		
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864		
CCC	SP6860		
Internal Standard/PEM	S13181,10ul/1000ul sample		
ICV/I.BLK	SP6872		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BG064712.D	7 Nov 2025 10:25		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BG064713.D	7 Nov 2025 11:06		RC/JU	Ok,M
3	PB170428BL	PB170428BL	BG064714.D	7 Nov 2025 12:30		RC/JU	Ok
4	PB170428BS	PB170428BS	BG064715.D	7 Nov 2025 13:12		RC/JU	Ok,M
5	PB170394TB	PB170394TB	BG064716.D	7 Nov 2025 14:04		RC/JU	Ok
6	Q3537-02	MW-17-SOIL	BG064717.D	7 Nov 2025 14:53		RC/JU	Ok
7	PB170446BS	PB170446BS	BG064718.D	7 Nov 2025 16:13		RC/JU	Ok,M
8	PB170446BSD	PB170446BSD	BG064719.D	7 Nov 2025 16:54		RC/JU	Ok,M
9	PB170446BL	PB170446BL	BG064720.D	7 Nov 2025 17:35		RC/JU	Ok
10	Q3551-01	MANHOLE	BG064721.D	7 Nov 2025 18:15		RC/JU	Ok
11	Q3577-01	WC-1A	BG064722.D	7 Nov 2025 18:56		RC/JU	Ok
12	Q3580-01	346	BG064723.D	7 Nov 2025 19:37		RC/JU	Ok
13	Q3562-01	EO-03-110625	BG064724.D	7 Nov 2025 20:18		RC/JU	Ok,M
14	Q3564-01	OR-03-11062025	BG064725.D	7 Nov 2025 20:59	Need Further 2X Dilution	RC/JU	Dilution

M : Manual Integration