



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q3551

Analytical Method: 625.1

Client: Spectra East Inc.

DataFile: BG064718.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB170446BS	bis(2-Ethylhexyl)phthalate	50	47.5	ug/L	95				43	137	