

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: Alliance Contract: SPEC01
 Lab Code: ACE SDG No.: Q3551
 Instrument ID: MSVOA_N Calibration Date(s): 11/13/2025 11/13/2025
 Heated Purge: (Y/N) N Calibration Time(s): 10:14 11:50
 GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID:		RRF005 = VN088274.D		RRF020 = VN088275.D		RRF050 = VN088276.D		
		RRF100 = VN088277.D		RRF150 = VN088278.D		RRF =		
COMPOUND	RRF005	RRF020	RRF050	RRF100	RRF150	RRF	RRF	% RSD
Methylene Chloride	2.165	2.098	1.983	1.998	2.009		2.050	3.8
trans-1,2-Dichloroethene	1.917	2.036	1.848	1.810	1.993		1.921	4.9
Chloroform	3.861	3.873	3.674	3.755	3.958		3.824	2.9
1,1,1-Trichloroethane	0.623	0.624	0.578	0.609	0.651		0.617	4.3
Trichloroethene	0.350	0.340	0.321	0.335	0.349		0.339	3.4
Toluene	1.716	1.777	1.619	1.714	1.782		1.722	3.8
Tetrachloroethene	0.305	0.329	0.295	0.297	0.321		0.309	4.8
Ethyl Benzene	1.763	1.927	1.813	1.933	2.052		1.897	6
m/p-Xylenes	0.660	0.723	0.687	0.718	0.760		0.710	5.4
o-Xylene	0.619	0.665	0.643	0.698	0.721		0.669	6.2
1,3-Dichlorobenzene	0.726	0.795	0.753	0.802	0.815		0.778	4.8
1,2-Dichloroethane-d4	2.725	2.740	2.769	2.745	2.723		2.740	0.7
Toluene-d8	1.418	1.455	1.418	1.385	1.426		1.420	1.8
4-Bromofluorobenzene	0.479	0.494	0.513	0.515	0.534		0.507	4.1

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.